

CERTIFICATE OF COMPLIANCE

We,

Roche Diabetes Care GmbH Sandhofer Strasse 116 68305 Mannheim Germany

hereby certify that

Accu-Chek Softclix Lancet (25) (03307492001)

Batch no: 10821153

Production Date: 25.Oct.2021 Expiry Date: 30.Sep.2025

This lot contains these sterile components:

Product number: S/A Product description: S/A Lot number: S/A

Method of sterilization: Gamma Irradiation

Sterilization expiry date: Sep.2025

were controlled in accordance with the valid Quality Control procedures and have met the specifications.

The sterility of all sterile components in the product is guaranteed.

Mannheim, 29 March 2022

Roche Diabetes Care GmbH

Quality Assurance

Frank Reichert

Report No. AG2-211295

To: ASAHI POLYSLIDER COMPANY, LIMITED

KOGA ISOTOPE LIMITED

Jinbo, Koka, Shiga, Japan Tel:+81(0)748-88-3125 Fax:+81(0)748-88-2296

GAMMA PROCESSING REPORT

This is to report the results of the gamma absorbed dose measurements on the processed products cartons. The material represented by this report was sterilized in accordance with ISO11137.

Customer Name

Roche Diabetes Care GmbH

Customer Address

Sandhofer Strasse 116, 68305 Mannheim/Germany

Products

03307492001 /6

Lot No.

10821153 264 Carton(s) [2,442,000 pcs]

Irradiation Order Number

APS10821153-0

Products Receipt Date at our Warehouse

2021-10-26

Sterilization Date

2021-10-27

Dose Measurement Date

2021-10-27

Requirement Dose Range

25-50kGy

Irradiator

#1

Representative	Minimum Absorbed	Maximum Absorbed
Tote box	Dose	Dose
Control Number	(kGy)	(kGy)
1	34.2	43.6
29	34.4	43.8
53	34.1	43.4
Blank		

Note: Sterilization, dosimetric procedures are given on the over-leaf.

Approved by: V. Konishi

Date:

2021-10-27

Sterilization Supervisor Takamune Konishi

RS01-4AG2007

	ASAHI POLYSLIDER COMPANY, LIM	ITED
Products	Accu-Chek Softclix Lancet	Operator
LotNo.	10821153	合格
Sterilization Dat 滅菌年月日	2021-10-27 Approved	21.10.28 QA太田
Expiration Date 有効年月	2025-09	