

21.08.2016
İzmir / Turkey

DECLARATION FOR THE ISSUANCE OF QUALITY CERTIFICATES

To Whom It May Concern,

According to IVD 98/79/EC directive,

FOR ANNEX II LIST A which includes HIV, Hepatitis B and Hepatitis C tests; the Notified Body must verify that the product meets the Common Technical Specification (CTS) and must release each batch of product before it is placed on the European market. The batch release often requires testing. These have EC Design Examination certificates by the notified body.

FOR ANNEX III which includes all other tests for Professional use; the manufacturer prepares a declaration of conformity in a similar way to the general devices.

For the above mentioned reason, we hereby declare that we provide CE Certificate for only the Hepatitis B, Hepatitis C and HIV tests for Professional use. For the group of other Professional tests; it is enough to present a self-Declaration of Conformity to the EU standards.

Cordially,

TURKLAB TIBBİ MALZEMELER SAN TİC A.Ş



EC CERTIFICATE No. 1434-IVDD-56/2016

EC Design-Examination

Directive 98/79/EC on in vitro diagnostic medical devices

PCBC certifies that the design documentation relating to in vitro diagnostic medical device, List A:

HBsAg Test

Brands: Info®, Toyo®, Rapidan Tester®, Labmen®

manufactured by:

TÜKLAB Tıbbi Mal. San. Tic. A.Ş.
İTOB 10031 Sokak No: 15 Tekeli Menderes
İzmir, Turkey

was examined by PCBC according to Annex IV p. 4 Directive 98/79/EC (with subsequent amendments) transposed into the Polish law and comply with the essential requirements of the Directive.

This certificate is valid from **2016-08-29** to **2019-08-28**

Date of certificate issue: **2016-08-29**

Date of first certificate issue: **2008-08-29**



Anna Wyroba
Anna Wyroba
Vice President of PCBC

CE 1434

PCBC Notified Body
23A, Kłobucka Str., PL-02-699 Warsaw

Application No. 45/2016
Contract No. MD-18/2016

Module H6



EC CERTIFICATE No. 1434-IVDD-57/2016

Full Quality Assurance System

Directive 98/79/EC on in vitro diagnostic medical devices

PCBC certifies quality assurance system in company:

TÜRKLAB Tibbi Mal. San. Tic. A.Ş.
ITOB 10031 Sokak No: 15 Tekeli Menderes
Izmir, Turkey

for the design, manufacture and final inspection of in vitro diagnostic medical devices,
List A:

HBsAg Test
Brands: Info®, Toyo®, Rapidan Tester®, Labmen®


complies with the requirements of Annex IV excl. 4, 6 Directive 98/79/EC
(with subsequent amendments) transposed into the Polish law. The audit of the quality
assurance system carried out by PCBC has provided evidence of the above.

This certificate is valid from **2016-08-29** to **2019-08-28**

Date of certificate issue: **2016-08-29**

Date of first certificate issue: **2008-08-29**




Anna Wyroba
Vice President of PCBC

PCBC Notified Body
23A, Kłobucka Str., PL-02-699 Warsaw

CE 1434

Application No. 45/2016
Contract No. MD-18/2016

Module H7



EC CERTIFICATE No. 1434-IVDD-52/2016

EC Design-Examination

Directive 98/79/EC on in vitro diagnostic medical devices

PCBC certifies that the design documentation relating to in vitro diagnostic medical
device, List A:

Anti-HCV Test
Brands: Info®, Toyo®, Rapidan Tester®, Labmen®

manufactured by:

TÜRKLAB Tibbi Mal. San. Tic. A.Ş.
ITOB 10031 Sokak No: 15 Tekeli Menderes
Izmir, Turkey

was examined by PCBC according to Annex IV p. 4 Directive 98/79/EC
(with subsequent amendments) transposed into the Polish law and comply with the
essential requirements of the Directive.

This certificate is valid from **2016-08-29** to **2019-08-28**

Date of certificate issue: **2016-08-29**

Date of first certificate issue: **2008-08-29**




Anna Wyroba
Vice President of PCBC

PCBC Notified Body
23A, Kłobucka Str., PL-02-699 Warsaw

CE 1434

Application No. 43/2016
Contract No. MD-16/2016

Module H6



EC CERTIFICATE No. 1434-IVDD-53/2016

Full Quality Assurance System

Directive 98/79/EC on in vitro diagnostic medical devices

PCBC certifies quality assurance system in company:

TÜRKLAB Tibbi Mal. San. Tic. A.Ş.
İTOB 10031 Sokak No: 15 Tekeli Menderes
Izmir, Turkey

for the design, manufacture and final inspection of in vitro diagnostic medical devices,
List A:

Anti-HCV Test
Brands: Info®, Toyo®, Rapidan Tester®, Labmen®

complies with the requirements of Annex IV excl. p. 4, 6 Directive 98/79/EC
(with subsequent amendments) transposed into the Polish law. The audit of the quality
assurance system carried out by PCBC has provided evidence of the above.

This certificate is valid from **2016-08-29** to **2019-08-28**

Date of certificate issue: **2016-08-29**

Date of first certificate issue: **2008-08-29**



Anna Wyroba
Vice President of PCBC

PCBC Notified Body
23A, Kłobucka Str., PL-02-699 Warsaw

CE 1434

Application No. 43/2016
Contract No. MD-16/2016

Module H7



EC CERTIFICATE No. 1434-IVDD-54/2016

EC Design-Examination

Directive 98/79/EC on in vitro diagnostic medical devices

PCBC certifies that the design documentation relating to in vitro diagnostic medical
device, List A:

Anti-HBs Test
Brands: Info®, Toyo®, Rapidan Tester®, Labmen®

manufactured by:

TÜRKLAB Tibbi Mal. San. Tic. A.Ş.
İTOB 10031 Sokak No: 15 Tekeli Menderes
Izmir, Turkey

was examined by PCBC according to Annex IV p. 4 Directive 98/79/EC
(with subsequent amendments) transposed into the Polish law and comply with the
essential requirements of the Directive.

This certificate is valid from **2016-08-29** to **2019-08-28**

Date of certificate issue: **2016-08-29**

Date of first certificate issue: **2008-08-29**



Anna Wyroba
Vice President of PCBC

PCBC Notified Body
23A, Kłobucka Str., PL-02-699 Warsaw

CE 1434

Application No. 44/2016
Contract No. MD-17/2016

Module H6



EC CERTIFICATE No. 1434-IVDD-55/2016

Full Quality Assurance System

Directive 98/79/EC on in vitro diagnostic medical devices

PCBC certifies quality assurance system in company:

TÜRKLAB Tibbi Mal. San. Tic. A.Ş.
İTOB 10031 Sokak No: 15 Tekeli Menderes
Izmir, Turkey

for the design, manufacture and final inspection of in vitro diagnostic medical devices,
List A:

Anti-HBs Test
Brands: Info®, Toyo®, Rapidan Tester®, Labmen®

complies with the requirements of Annex IV excl. p. 4, 6 Directive 98/79/EC
(with subsequent amendments) transposed into the Polish law. The audit of the quality
assurance system carried out by PCBC has provided evidence of the above.

This certificate is valid from **2016-08-29** to **2019-08-28**

Date of certificate issue: **2016-08-29**

Date of first certificate issue: **2008-08-29**



Anna Wyroba
Anna Wyroba
Vice President of PCBC

PCBC Notified Body
23A, Kłobucka Str., PL-02-699 Warsaw

CE 1434

Application No. 44/2016
Contract No. MD-17/2016

Module H7



EC CERTIFICATE No. 1434-IVDD-58/2016

EC Design-Examination

Directive 98/79/EC on in vitro diagnostic medical devices

PCBC certifies that the design documentation relating to in vitro diagnostic medical
device, List A:

Anti - HIV 1/2 Test
Brands: Info®, Toyo®, Rapidan Tester®, Labmen®

manufactured by:

TÜRKLAB Tibbi Mal. San. Tic. A.Ş.
İTOB 10031 Sokak No: 15 Tekeli Menderes
Izmir, Turkey

was examined by PCBC according to Annex IV p. 4 Directive 98/79/EC
(with subsequent amendments) transposed into the Polish law and comply with the
essential requirements of the Directive.

This certificate is valid from **2016-08-29** to **2019-08-28**

Date of certificate issue: **2016-08-29**

Date of first certificate issue: **2008-08-29**



Anna Wyroba
Anna Wyroba
Vice President of PCBC

PCBC Notified Body
23A, Kłobucka Str., PL-02-699 Warsaw

CE 1434

Application No. 46/2016
Contract No. MD-19/2016

Module H6



EC CERTIFICATE No. 1434-IVDD-59/2016

Full Quality Assurance System

Directive 98/79/EC on in vitro diagnostic medical devices

PCBC certifies quality assurance system in company:

TÜRKLAB Tibbi Mal. San. Tic. A.Ş.
İTOB 10031 Sokak No: 15 Tekeli Menderes
Izmir, Turkey

for the design, manufacture and final inspection of in vitro diagnostic medical devices,
List A:

Anti - HIV 1/2 Test
Brands: Info®, Toyo®, Rapidan Tester®, Labmen®

complies with the requirements of Annex IV excl. p. 4, 6 Directive 98/79/EC
(with subsequent amendments) transposed into the Polish law. The audit of the quality
assurance system carried out by PCBC has provided evidence of the above.

This certificate is valid from **2016-08-29** to **2019-08-28**

Date of certificate issue: **2016-08-29**

Date of first certificate issue: **2008-08-29**



Anna Wyroba
Anna Wyroba
Vice President of PCBC

PCBC Notified Body
23A, Kłobucka Str., PL-02-699 Warsaw

CE 1434

Application No. 46/2016
Contract No. MD-19/2016

Module I17



EC CERTIFICATE No. 1434-IVDD-51/2016

EC Design-Examination

Directive 98/79/EC on in vitro diagnostic medical devices

PCBC certifies that the design documentation relating to in vitro diagnostic medical
device for self-testing:

hCG Pregnancy Test
Brands: Rapidan Nova®, Rapidan Optima®, Info®, Toyo®, Labmen®
Tester®, Rapidan Compact®, Labmen®

manufactured by:

TÜRKLAB Tibbi Mal. San. Tic. A.Ş.
İTOB 10031 Sokak No: 15 Tekeli Menderes
Izmir, Turkey

was examined by PCBC according to Annex III p. 6 Directive 98/79/EC
(with subsequent amendments) transposed into the Polish law and comply with the
essential requirements of the Directive.

This certificate is valid from **2016-08-29** to **2019-08-28**

Date of certificate issue: **2016-08-29**

Date of first certificate issue: **2008-08-29**



Anna Wyroba
Anna Wyroba
Vice President of PCBC

PCBC Notified Body
23A, Kłobucka Str., PL-02-699 Warsaw

CE 1434

Application No. 42/2016
Contract No. MD-15/2016

Module A1