

ITEM 2
FREEZER

| No. | Minimum Technical Requirement | Conformity | | Bidder's Offer | Remarks |
|-----|--|----------------|----|----------------|------------|
| | | Yes | No | | |
| 1 | Model | NS 1400/40 | | | |
| 2 | Country of Origin | Czech republic | | | |
| 3 | Name of Manufacture | Frigera s.r.o. | | | |
| 4 | For storage and preservation of frozen blood products | yes | | | |
| 5 | Type: vertical, with two doors | Yes | | | |
| 6 | Internal room volume from 1300 lit to 1400 lit | yes | | 1400l | |
| 7 | Temperature: - from minus 25 ° C to minus 30 ° C; - adjustable; - programmable; - accuracy ± 1 ° C | yes | | | minus 40°C |
| 8 | Temperature monitoring: - ensuring the integration of thermal data with local SAP Blood Service; | Yes | | | |
| 9 | Electronic control system of temperature. Information panel: | Yes | | | |
| 10 | will allow the visualization of the temperature parameters in daily / weekly time | Yes | | | |
| 11 | provided with alarm in case of exceeding the programmed parameters | Yes | | | |
| 12 | Interior surfaces made of stainless steel resistant to the action of biocidal substances used for sanitary processing | Yes | | | |
| 13 | External surfaces made of stainless steel resistant to the action of biocidal substances used for sanitary processing. | Yes | | | |
| 14 | Mandatory components present: material of stainless steel drawers | Yes | | | |



FRIGERA, s.r.o.
Hájecká 1304/14a, 618 00 Brno
IČ 28519400 DIČ CZ28519400

-04-

| | | | | |
|----|--|-----|--|--|
| 15 | perforated surface for air circulation | Yes | | |
| 16 | drawers in number of 12 - 14 units, adjustable | Yes | | |
| 17 | secured with telescopic system | Yes | | |
| 18 | Cooling system: Forced air | Yes | | |
| 19 | Automatic defrost system | Yes | | |
| 20 | Refrigerant: Ecological, approved by the EC | Yes | | |
| 21 | EC / MS Declaration of Conformity and / or EC / MS Certificate of Conformity | Yes | | |
| 22 | Declaration from the manufacturer confirming that the medical device is intended for the storage of frozen blood products. | Yes | | |
| 23 | Year of production of the device 2021-2022 | Yes | | |
| 24 | Warranty period of not less than 36 months | Yes | | |



FRIGERA, s.r.o.
Hájecká 1304/14a, 618 00 Brno
IC 28519400 DIC CZ28519400

[Handwritten signature]