

Statie Producere Oxigen Medical Atlas Copco tip OGP 18 PED

Mentenanata pentru 12 luni (PlanA- H4000, PlanB - H8000)

Nr. d/o	KIT SERVICE	COD	CANT.	NOTA
1	GENERATOR OGP18 PED 115/230V MK5	8102318378	1	
2	REZERVOR OXIGEN 1000 L OGP20-35 11B CE	8092352304	1	
3	FILTRU PDp25+ (G1/2)	8102372605	1	
4	COMPRESOR GA22+ P A 10 ME AA 400/50	8153329183	1	
5	USCATOR AER FX90 (E8) 230/50-C-CE	8102227787	1	
6	REZERVOR AER LV 1011L 1000L 12BAR CE	8101021809	2	
7	PURJA EWD 50 230 50/60	8102043174	1	
8	FILTRU DD75+ (G1)	8102372009	1	
9	FILTRU PD75+ (G1)	8102372309	1	
10	FILTRU QDT95 G THREAD	8102296947	1	
11	FILTRU PDp75+ (G1)	8102372609	1	
12	FILTRU STERIL MEDICAL WALKER	O20032MS	1	