

## **Certificate of Analysis**

# 

### **CERTIFIED REFERENCE MATERIAL**

This document is designed and the certified value(s) and uncertainty(ies) are determined in accordance with ISO Guide  $31^{[1]}$ , ISO Guide  $35^{[2]}$ , and Eurachem / CITAC Guides<sup>[3]</sup>

Lot N: F09992		Barcode: 83090882	Certification Date: 13.10.2016 Date of Stability last check:
Description of the Reference Material (CRM):		Solution of: Zinc (Zn) concentration 1.000 g/l in 2% Nitric Acid (HNO3)	
Material Number:		86725.180	
Certified value/ Uncertainty:	Element	Certified Value and Uncertainty [mg/l]:	Metrological traceability:
	Zn	998.9 ± 3.6 (v	NIST SRM No 3168a Lot 120629
Method(s) of certification used:	CRM's calib	ration procedure(s):	Notes:
	(y) WQP 5.15.1.24		The certified value was obtained by a weighted mean of the results of two independent methods among: Classical Volumetric, Primary Gravimetric, Instrumental (ICP/OES, ICP/MS or IC)

#### Concept of Certification and traceability statement:

This certified reference material is produced using a high-purity starting material, acid from sub-boiling and 18 MOhm deionized water. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02

Property of the result of a measurement whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties (ISO VIM<sup>[5]</sup>). The metrological traceability is assured through calibration on AAS . The calibration curve is drawn using a series of standard solutions prepared from a certified reference material traceable to SI of NIST (SRM) and of accredited according to ISO/IEC 17025<sup>[6]</sup> and/or ISO Guide 34<sup>[7]</sup> laboratories/producers. All contributions in relation to the certification of standard solutions are considered when evaluating the uncertainty. The measurement results are traceable to SI. All analytical balances used for the preparation of the solution are calibrated yearly under an in-house procedure with analytical weights, traceable to DKD and are daily checked. Class A laboratory glassware is used. The results from temperature measurement are traceable to SI. The thermometers used for solution's calibration are calibrated from an ISO 17025 accredited from an ISO 17025 accredited laboratory. The ambient conditions are controlled with a hygrometer calibrated from an ISO 17025 accredited laboratory.

#### Starting material :

	Zn	Zn - 83090882	
Density *:	1.013 g/cm <sup>3</sup> at 20 °C		
Minimum shelf-life:	02.2019 (unopened bottle in aluminized bag)		
Date of opening:			
	(Recommended period of use sho	ould not exceed 12 months from date of opening.)	
* These values are not certified			

VWR International LLC, Radnor Corporate Center, Building One, Suite 200, 100 Matsonford Road, Radnor, PA 19087, USA VWR International bvba, Haasrode Research Park Zone 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium

#### Intended use: For Laboratory Use Only

Calibration of AAS

Validation of analytical methods Detection limit and linearity studies

Preparation of "working reference samples"

This statement is not intended to restrict the use for other purposes.

#### Instructions for the correct use of this reference material:

This certified reference material can be used directly or can be diluted in an appropriate high-purity matrix. Only a clean class A glassware should be used. Do not pipet from container. Obtained concentration (in mg/l) after dilution is a result from the multiplication of certified value of CRM concentration and the CRM's volume used for dilution and divided into the flask's volume used for dilution .

#### Stability and storage:

Storage conditions: To be stored under normal laboratory conditions

This CRM is with a guaranteed stability until ±0.5% of the certified concentration within its shelf-life. Stability is guaranteed provided that the solution is kept in its original packaging, tightly closed as written in Storage conditions. According to an in-house procedure the producer will monitor this CRM at appropriate intervals and the purchasers will be notified of any significant changes resulting in recertification or with withdrawal of the CRM during the state period of the validity of the certificate. The original packaging consists of the bottle & its aluminized bag. The aluminized bag is an inseparable part of the unit packing of the

CRM.

#### Hazardous situation:

The normal laboratory safety precautions should be observed when working with this RM. Further details for the handling of this RM are available as safety data sheet.

#### Level of homogeneity:

The material was tested for homogeneity by analyzing randomly selected samples according to an in-house procedure. The level of homogeneity proved satisfactory for a sample volume of 20 ml. The uncertainty incorporates the sample standard deviation combined with the uncertainty calculated from homogeneity and stability studies.

To ensure sufficient homogeneity of the sample prior to use thoroughly mix by inversion.

#### Names of certifying officer:

Saralova Krassimira Taralova Manager:

[1] ISO Guide 31: Reference materials - Contents of certificates and labels [2] ISO Guide 35: Reference materials - General and statistical principles for certification [3] EURACHEM/CITAC Guide: Quantifying Uncertainty in Analytical Measurement [4] ISO/IEC 17025: General requirements for the competence of testing and calibration laboratories [5] ISO Guide 34: General Requirements for the Competence of Reference Material Producers

This certificate relates solely to the lot number given above. All processes (including generating of this certificate) are completely controlled by the specialized Computer-Aided-Manufacturing (CAM) software.

This Certified Reference Material was produced under a quality management system that is:

Registered to ISO 9001 Quality Management System (Lloyd's Register Quality Assurance Ltd Cert No SOF0368072)
Accredited according to ISO/IEC 17025 – Testing (ANAB Cert No AT-1836)

- Accredited according to ISO Guide 34 Reference Material Producer (ANAB Cert No AR-1835)