

Declaration of Conformity

- for EU and UK -

Manufacturer's Name: SHIMADZU CORPORATION
 Analytical & Measuring Instruments Division
Address: 1, Nishinokyo-Kuwabara-cho, Nakagyo-ku,
 Kyoto 604-8511, Japan

declares in sole responsibility that the following product

Product Name Photo Diode Array Detector
Model Name SPD-M40
P/N Depend on configuration. See Appendix 1.

referred to in this declaration conforms with following legislations and standards

EMC Directive 2014/30/EU	EMC Regulation S.I. 2016/1091
EN 61326-1:2013	EN 61326-1:2013
EN 61326-2-6:2013	EN 61326-2-6:2013
EN 61326-2-6:2006	EN 61326-2-6:2006
Low Voltage Directive 2014/35/EU	Electrical Equipment Regulation S.I. 2016/1101
EN 61010-1:2010 / A1: 2019	EN 61010-1:2010 / A1: 2019
RoHS Directive 2011/65/EU	RoHS Regulation S.I. 2012/3032

Note 1) This declaration becomes invalid if technical or operational modifications are introduced without manufacturer's consent.

Note 2) This declaration is valid if this product is used alone or in combination with the accessories of this product which are mentioned in attached Appendix 1 or other instruments which fulfill with the requirement of mentioned legislation.

Note 3) Authorised Representative in EU is as follows:

SHIMADZU EUROPA GmbH

Address: Albert-Hahn-Strasse 6-10,47269 Duisburg, F.R.Germany

Note 4) Authorised Representative in UK is as follows:

SHIMADZU UK Limited

Address: Unit 1A, Mill Court, Featherstone Road Wolverton Mill South,
 Milton Keynes MK12 5RE, U.K

Kyoto, JAPAN 2022-12-22

Place and date of issue



Signature

Toshiyuki Kawano

Name

General Manager of Quality Assurance Department
 Analytical & Measuring Instruments Division

Title



Appendix 1

Product Name: Photo Diode Array Detector

Model No.	P/N
SPD-M40	228-65302-58

Product Options

Model No.	P/N
STANDARD CELL	228-62156-41
UHPLC CELL	228-64724-41
SEMI MICRO CELL	228-64725-41
INERT CELL	228-64728-41
PREPARATIVE CELL	228-64727-41
MICRO CELL	228-64737-41
HIGH PRESSURE CELL	228-64726-41
INERT UH CELL 40 PDA	228-74221-41
INERT LD CELL 40 PDA	228-74222-41



SHIMADZU

新規・改訂内容一覧表

担当部署:分析機器事業部品質保証部

改訂記号	改訂理由	改訂内容	作成者 日付	検討者 日付	承認者 日付
制定	新規作成	-	砂川 2019.4.12	伊藤 2019.4.12	逢坂 2019.4.13
A	規格更新など	<ul style="list-style-type: none"> ・EN61010 A1 : 2019 対応 ・UK 関連文言追加 	藤田 2022.4.25	伊藤 2022.4.25	河野 2022.4.25
B	イナーテル追加	<ul style="list-style-type: none"> ・イナーテル追加 	藤田 2022.6.7	伊藤 2022.6.7	河野 2022.6.7
C	本文変更なし	<ul style="list-style-type: none"> ・SW 電源の代替品追加に伴い TDF に EMC, LVD レポートを追記したため再発行する 			

配布先