

**SCHILLER**

The Art of Diagnostics

SCHILLER AG  
Altgasse 68, Postfach  
6341 Baar, Switzerland  
CHE -105.868.779 MWST

Tel: +41 41 766 42 42

Fax: +41 41 761 08 80

info@schiller.ch

www.schiller.ch

CH-6341 Baar, 18.06.2025

To: Center for Centralized Public Procurement in Health

For tender nr: 21396977

Date of tender 08.04.2025

#### CONFIRMATION LETTER

The company Schiller AG confirms, that the 12- lead ECG Holter, model medilog FD fits and complies with the requirements of the public tender.

Requested technical specification	Confirmation of the requested technical specification
Frequency range Diagnostic 0.05-100 Hz	YES
Input impedance $\geq 100$ Mohm	YES
Common mode rejection range at 50 Hz $> 100$ dB	YES
Automatic button lock required	YES, after the recording starts, the device screen turns off and stopping the recording is only possible if both buttons are pressed at the same time, for no less than 5 seconds. The unintentional disconnection of the device is impossible.

Yours sincerely,

Mircea Dumitrescu

International Sales Manager.