



Sialkot Material Testing Laboratory

UNIT ESTABLISHED BY TRADE DEVELOPMENT AUTHORITY OF PAKISTAN MINISTRY OF COMMERCE GOVT OF PAKISTAN
NEAR SUBLIME CHOWK, MARALA ROAD, SIALKOT - 51310-PAKISTAN
TEL: 0092-52-3562818 simtelsktlab@gmail.com FAX: 0092-52-3554217



SIMTEL/PD/23/1500

June 26, 2023

ANALYTICAL/ TEST REPORT

In respect of : *M/s. Care & Cure Surgico Sialkot*
Sample received from : *Representative*
*Needle/Scalpel/Chisel/Shear/Clamp/Rib Shear/Knife/
Hook/ Curette/Tuning Fork /Forceps /Gouge*
Name of Sample : */Lever/Mallet/Scalpel Handle/Hammer/Osteotome/Mouth
Gag/Square Point/Needle/Bone Raspatory*
Type/Grade of Raw Material : *AISI-410*
Specification against which : *Analytical Results Got on BAIRD
tested DV 6E Direct Emission Spectrometer*

Results are in percentage:

Cr.....		12.561
C	0.143	Mn 0.584
P	0.024	S 0.017
Si	0.275	Ni 0.101

This analytical/test report is based solely on the particular sample(s) supplied by the party and should not be reproduced in part.

Sampling has not been performed by the SIMTEL and the SIMTEL does not accept responsibility that the sample(s) supplied is/are truly representative sample(s) of any batch or stock or entire production.

While the SIMTEL agrees to take every reasonable precaution to ensure validity of its test results, it assumes no liability for the accuracy thereof beyond the amount of the fee charged for the analysis/test.

The party shall assume full responsibility for the ethical use of the results in the analysis/test reports and the laboratory shall be held free from any and all claims which may result from the use of such data by the party or others.

After completion of the report the sample will be preserved for three month until negotiated otherwise.



Engr. Amjad Hussain
Project Director