



MANGO DE BISTURÍ ELECTROQUIRÚRGICO, CONTROL MANUAL

HAND CONTROL ELECTROSURGICAL PENCIL
MANCHE DE BISTOURI ÉLECTROCHIRURGICAL, COMMANDE DIGITALE

CLASE 2b
CLASS 2b
CLASE 2b

CARACTERÍSTICAS


- Fácil manejo
- Ligero
- Cómodo perfil anatómico
- Conector standard de 3 pins

CHARACTERISTICS

- Easy handling
- Light
- Comfortable anatomical shape
- 3-pin standard connector

CARACTÉRISTIQUES

- Facile utilisation
- Léger
- Confortable contour anatomique
- Connecteur standard de 3 pins




MB-100

Mango bisturí electroquirúrgico de control manual de un solo uso, con accesorio AB-80, desmontable.
Single use hand control electrosurgical pencil, with removable AB-80 accessory.
Manche bistouri électrique à usage unique, commande digitale avec accessoire AB-80 amovible.



AL-40

Dispositivo limpiador de electrodo, autoadhesivo.
Adhesive electrode cleaning device.
Dispositif grattoir d'électrode autoadhésive



MB-200

Contiene
Contains
Contient

1 MB-100 + 1 AL-40



MBR-600

Mango de bisturí electroquirúrgico reutilizable, con accesorio AB-80 desmontable.
Reusable electrosurgical pencil, with removable AB-80 accessory.
Manche de bistouri électrochirurgical réutilisable, avec accessoire AB-80 amovible.




LATEX FREE



REF	MEDIDAS MEASUREMENTS DIMENSIONS (mm)	SOPORTE BACKING SUPPORT	GEL	ETIQUETA LABEL ÉTIQUETTE	CONEXIÓN CONNECTION CONNEXION	USO PURPOSE USAGE	U/BOLSA U/POUCH U/POCHE	U/CAJA U/BOX U/CARTON
LF-50	 Ø 50	FOAM MOUSSE	SEMILÍQUIDO SEMILIQUID SEMI-LIQUIDE	X	CORCHETE • STUD • AGRAFE	A	50	1000
LF-36	 36 x 50		SEMILÍQUIDO SEMILIQUID SEMI-LIQUIDE	X		A / P		
SX-50	 Ø 50		SÓLIDO SOLID SOLIDE			A		
SX-36	 36 x 50		SÓLIDO SOLID SOLIDE			A / P		
SF-36	 36 x 42		SÓLIDO SOLID SOLIDE	X		A / P		
SX-30	 Ø 30		SÓLIDO SOLID SOLIDE	X		A / P / N		
LEH-36	 36 x 50		SEMILÍQUIDO SEMILIQUID SEMI-LIQUIDE	X		STRESS		
SP-50	 Ø 50	PAPEL TAPE PAPIER	SÓLIDO SOLID SOLIDE	X	A	50	1000	
LP-50	 Ø 50		SEMILÍQUIDO SEMILIQUID SEMI-LIQUIDE	X	A			
LR-50	 Ø 50	TEXTIL TEXTILE TISSU	SEMILÍQUIDO SEMILIQUID SEMI-LIQUIDE	X	A	50	1000	
EKF-22KT	 22 x 22	FOAM MOUSSE	SÓLIDO SOLID SOLIDE		A / P / N	6	300	

CARDIOLOGÍA • CARDIOLOGY • CARDIOLOGIE

REF	MEDIDAS MEASUREMENTS DIMENSIONS (mm)	SOPORTE BACKING SUPPORT	GEL	ETIQUETA LABEL ÉTIQUETTE	CONEXIÓN CONNECTION CONNEXION	USO PURPOSE USAGE	U/BOLSA U/POUCH U/POCHE	U/CAJA U/BOX U/CARTON
K-140	Ø 30	PE (CLEAR TAPE)	SÓLIDO SOLID SOLIDE	X	Ø1,5 mm CONECTOR SEGURIDAD HEMBRA SAFETY PLUG FEMALE CONNECTEUR DE SURETÉ FEMELLE	P / N	3	99
KS-140	Ø 25			X				
K-150	Ø 30			X				
KS-150	Ø 25			X	Ø4 mm HEMBRA FEMALE FEMELLE			
KF-140	Ø 30	FOAM MOUSSE	SÓLIDO SOLID SOLIDE	X	Ø1,5 mm CONECTOR SEGURIDAD HEMBRA SAFETY PLUG FEMALE CONNECTEUR DE SURETÉ FEMELLE	P / N	3	99
KFS-140	Ø 25			X				
KF-150	Ø 30			X				
KFS-150	Ø 25			X	Ø4 mm HEMBRA FEMALE FEMELLE			




K-140, KS-140, KF-140, KFS-140



K-150, KS-150, KF-150, KFS-150





Telic, S.A.U.

Polígono Industrial Can Barri
C/ Moli d'en Barri, 7-9,
08415 Bigues i Riells, BARCELONA, Spain
Tel: +34 93 865 61 25
Fax: +34 93 865 62 46

EC DECLARATION OF CONFORMITY

TELIC, S.A.U. declares that the products listed in annexes of the present declaration have been manufactured according to requirements of the **Medical Devices Directive 93/42/EEC** and meet requirements set in the Essential Requirements of the Annex I of above mentioned Directive.

Technical documentation, in accordance with the established in the corresponding annexes of Directive 93/42/EEC, is updated and located in our facilities. We are in position to submit these documents in case of Notified Body or Competent Authority requirement.

This declaration applies to design, manufacturing and final control of medical devices. Validity of the present declaration is subject to the expiration of the corresponding EC certificates for different products.

Bigues, on Septiembre 28th 2017

Laura Delgado
Technical Manager

Oscar Lacruz
CEO

EC DECLARATION OF CONFORMITY – ANNEX 1
List of products with EC mark

Defibrillation electrodes without cable	
Description	Set of two adhesive pre-gelled pads with conductive hydrogel for defibrillation. To be used for adult patient use.
Commercial brand	DESFI-DORMO
References	ED-1010
Classification	
Product class I Ib - Non-sterile. According to Rule 9 of Annex IX of the 93/42/EEC Directive	
GMDN	11130
UMDNS	-26915
EC Full Quality Assurance System Certificate	
In accordance to Annex II (except Section 4) of Directive 93/42/EEC	
Certificate number 572.170827	
Issued by: UL International (UK) Ltd.	
Notified Body number 0843	
Valid until: 26/08/2022	
Standards applied	
EN ISO 14971:2012// EN 60601-2-4:2011// EN ISO 10993-1:2009//EN ISO 10993-1:2009/AC:2010//EN ISO 10993-4:2009//EN-ISO 10993-5:2009// EN ISO 10993-7:2008// EN ISO 10993-7:2008/AC:2009// EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN ISO 15223-1:2012 Directiva 2011/65/EU (RoHS 2).	

Return electrodes for electrosurgery	
Description	Pre-gelled electrosurgical plates.
Commercial brand	BLAYCO
References	2125, 2125-5, 2125-C/00, 2125-C/00/5, 2125-C/10, 2125-C/10/5, 2125-C/21 2225, 2225-5, 2225-C/00,2225-C/00/5, 2225-C/10, 2225-C/10/5 2425, 2425-5, 2425-C/00, 2425-C/00/5, 2425-C/10,2425-C/10/5 2925, 2925-5, 2925-C/00, 2925-C/00/5, 2925-C/10, 2925-C/10/5 2500, 2500-5, 2500-C/00,2500-C/00/5, 2500-C/12, 2500-C/12/5 2510, 2510-5, 2510-C/00,2510-C/00/5, 2510-C/12, 2510-C/12/5 2600, 2600-5, 2600-C/00,2600-C/00/5, 2600-C/12, 2600-C/12/5 2700, 2700-5, 2700-C/00,2700-C/00/5, 2700-C/12, 2700-C/12/5 2900, 2900-5, 2900-C/00, 2900-C/00/5, 2900-C/12, 2900-C/12/5
Classification	
Product class I Ib - Non-sterile. According to Rule 9 of Annex IX of the 93/42/EEC Directive	
GMDN	58494
UMDNS	11500
EC Full Quality Assurance System Certificate	
In accordance to Annex II (except Section 4) of Directive 93/42/EEC	
Certificate number 572.170827	
Issued by: UL International (UK) Ltd.	
Notified Body number 0843	

Valid until: 26/08/2022
Standards applied
EN ISO 14971:2012// EN 60601-2-2:2009//EN 60601-2-2:2009/A11:2011//EN ISO 15223-1:2012// EN ISO 10993-1:2009//EN ISO 10993-1:2009/AC:2010//EN ISO 10993-5:2009//EN-ISO 10993-10:2013// EN 60601-1:2006/EN 60601-1:2006/A1:2013 // EN 60601-2-2:2009// EN 60601-2-2:2009/A11:2011. 2011/65/EU Directive (RoHS 2).

Vein strippers	
Description	Single use vein strippers.
Commercial brand	DORMO-STRIP
References	VE-022, VE-025, VE-022 OP, VE-022 BE CO
Classification	
Product class IIa - Sterile. According to Rule 7 of Annex IX of the 93/42/EEC Directive	
GMDN	32321
UMDNS	13828
EC Full Quality Assurance System Certificate	
In accordance to Annex II (except Section 4) of Directive 93/42/EEC Certificate number 572.170827 Issued by: UL International (UK) Ltd. Notified Body number 0843 Valid until: 26/08/2022	
Standards applied	
EN ISO 14971:2012// EN 62366:2008/ EN 62366:2008/A1:2015// EN ISO 11607-1:2009/ EN ISO 11607-1:2009/A1:2014// EN ISO 11607-2:2006/ EN ISO 11607-2:2006/A1:2014// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC:2010// EN ISO 10993-4:2009// EN ISO 10993-5:2009// EN ISO 10993-7:2008/ EN ISO 10993-7:2008/AC:2009//EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN 556-1:2001/ EN 556-1:2001/AC:2006// EN ISO 11135-1:2007// CEN ISO/TS 11135-2:2008// EN ISO 14644-1:2015// EN ISO 14644-3:2005// EN ISO 14644-4:2001// EN ISO 14644-5:2004// EN ISO 14644-8:2013// EN ISO 15223-1:2012// EN 1041:2008+A1:2013.	

Vascular loops	
Description	Two vascular loops for identification, occlusion, retraction.
Commercial brand	DORMO-LOOP
References	405305, 402509, 402505, 402503, 402501, 401909, 401905, 401903, 401901, 400709
Classification	
Product class IIa - Sterile. According to Rule 7 of Annex IX of the 93/42/EEC Directive	
GMDN	61916
UMDNS	-12240
EC Full Quality Assurance System Certificate	
In accordance to Annex V of Directive 93/42/EEC Certificate number 736.170827	

Issued by: UL International (UK) Ltd.

Notified Body number 0843

Valid until: 26/08/2022

Standards applied

EN ISO 14971:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC:2010//EN ISO 10993-4:2009//EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC:2009//EN ISO 10993-10:2013//EN ISO 10993-11:2009//// EN ISO 11135-1:2007// EN ISO 14644-1:2015//EN ISO 14644-2:2015//EN ISO 14644-4:2001//EN ISO 14644-5:2004// EN ISO 15223-1:2012.

Defibrillation electrodes with cable for adult patient

Description Set of two multi-function electrodes. Adult.

Commercial brand DESFI-DORMO

References EDA-1012, EDC-1015, EDC-1020, EDC-1025, EDC-1030, EDC-1035, EDC-1040, EDC-1045, EDC-1050, EDC-1055, EDC-1060, EDC-1065, EDC-1070, EDC-1075.

Description Set of two multi-function electrodes. Adult. Pre-connected.

Commercial brand DESFI-DORMO

References EDC-2015, EDC-2020, EDC-2025, EDC-2030, EDC-2035, EDC-2040, EDC-2045, EDC-2050, EDC-2055, EDC-2060, EDC-2065, EDC-2070, EDC-2075.

Classification

Product class IIb – Non-sterile. According to Rule 9 of Annex IX of the 93/42/EEC Directive

GMDN 45806

UMDNS -3698

EC Full Quality Assurance System Certificate

In accordance to Annex II (except Section 4) of Directive 93/42/EEC

Certificate number 572.170827

Issued by: UL International (UK) Ltd.

Notified Body number 0843

Valid until: 26/08/2022

Standards applied

EN ISO 14971:2012// EN 60601-2-4:2011// EN ISO 10993-1:2009//EN ISO 10993-1:2009/AC:2010//EN ISO 10993-4:2009//EN ISO 10993-5:2009// EN ISO 10993-7:2008// EN ISO 10993-7:2008/AC:2009// EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN ISO 15223-1:2012 Directiva 2011/65/EU (RoHS 2).

Defibrillation electrodes with cable pediatrics

Description Set of two defibrillation electrodes. Paediatrics.

Commercial brand DESFI-DORMO

References EDC-P115, EDC-P120, EDC-P125, EDC-P130, EDC-P135, EDC-P140, EDC-P145, EDC-P150, EDC-P155, EDC-P160, EDC-P165, EDC-P170, EDC-P175.

Description Set of two defibrillation electrodes. Paediatrics. Pre-connected.

Commercial brand DESFI-DORMO

References	EDC-P215, EDC-P220, EDC-P225, EDC-P230, EDC-P235, EDC-P240, EDC-P245, EDC-P250, EDC-P255, EDC-P260, EDC-P265, EDC-P270, EDC-P275.
Classification	
Product class I Ib – Non-sterile. According to Rule 9 of Annex IX of the 93/42/EEC Directive	
GMDN	41587
UMDNS	-26914
EC Full Quality Assurance System Certificate	
In accordance to Annex II (except Section 4) of Directive 93/42/EEC Certificate number 572.170827 Issued by: UL International (UK) Ltd. Notified Body number 0843 Valid until: 26/08/2022	
Standards applied	
EN ISO 14971:2012// EN 60601-2-4:2011// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC:2010//EN ISO 10993-4:2009//EN ISO 10993-5:2009// EN ISO 10993-7:2008// EN ISO 10993-7:2008/AC:2009//EN ISO 10993-10:2013//EN ISO 10993-11:2009// EN ISO 15223-1:2012. ANSI/AAMI DF80:2003 2011/65/EU Directive (RoHS 2).	

Sterile ultrasound gel	
Description	Ultrasound transmission gel. Sterile.
Commercial brand	TRANSONIC
References	G-15E
Classification	
Product class I - Sterile. According to Rule 5 of Annex IX of the 93/42/EEC Directive	
GMDN	58735
UMDNS	-1519
EC Full Quality Assurance System Certificate	
In accordance to Annex II (except Section 4) of Directive 93/42/EEC Certificate number 572.170827 Issued by: UL International (UK) Ltd. Notified Body number 0843 Valid until: 26/08/2022	
Standards applied	
EN ISO 14971:2012// EN ISO 10993-1:2009//EN ISO 10993-1:2009/AC:2010// EN ISO 10993-4:2009//EN ISO 10993-5:2009//EN ISO 10993-10:2013//EN ISO 10993-11:2009// EN ISO 15223-1:2012// EN ISO 11607-1:2009//EN ISO 11607-1:2009/A1:2014//EN ISO 11607-2:2006//EN ISO 11607-2:2006/A1:20//EN ISO 14644-1:2015//EN ISO 14644-3:2005//EN ISO 14644-4:2001//EN ISO 14644-5:2004//EN ISO 14644-8:2013 14// EN 556-1:2001//EN 556-1:2001/AC:2006//EN ISO 11137-1:2015//EN ISO 11137-2:2015// EN ISO 14644-1:2015//EN ISO 14644-3:2005//EN ISO 14644-4:2001//EN-ISO 14644-5:2004//EN ISO 14644-8:2013.	

Cover for surgical light handle	
Description	Cover for surgical light handle.
Commercial brand	BLAYCO
References	LHC-01, LHC-02, LHC-03
Classification	
Product class I - Sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	44977
UMDNS	17977
EC Full Quality Assurance System Certificate	
In accordance to Annex II (except Section 4) of Directive 93/42/EEC	
Certificate number 572.170827	
Issued by: UL International (UK) Ltd.	
Notified Body number 0843	
Valid until: 26/08/2022	
Standards applied	
EN ISO 14971:2012// EN 556-1:2001// EN 556-1:2001/AC: 2006// EN ISO 11135-1:2007// EN ISO 14644-1:2015// EN ISO 14644-3:2005// EN ISO 14644-4:2001// EN ISO 14644-5:2004// EN ISO 15223-1:2012.	

Tungsten electrosurgical accessories	
Description	Disposable sterile electrodes in tungsten
Commercial brand	BLAYCO
References	ABC-45, ABC-55, ABC-65
Classification	
Product class IIb - Sterile. According to Rule 9 of Annex IX of the 93/42/EEC Directive	
GMDN	61869
UMDNS	16860
EC Full Quality Assurance System Certificate	
In accordance to Annex II, section 3 of Directive 93/42/EEC	
Certificate number 685.130511	
Issued by: UL International (UK) Ltd.	
Notified Body number 0843	
Valid until: 10/05/2018	
Standards applied	
ISO 13485:2003// EN 60601-1:2006// EN 60601-2-2:2009// EN ISO 14971:2012// EN ISO 10993-1:2009// EN 1041:2008// EN ISO 15223-1:2012// EN ISO 11135-1:2007// EN 556-1:2001/AC: 2006// EN ISO 11607-1:2009//EN ISO 11607-2:2006. 65/EU Directive (RoHS 2).	

Electrode tip cleaner	
Description	Electrode tip cleaner.
Commercial brand	BLAYCO
References	AL-40
Classification	
Product class I - Sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	37483
UMDNS	15013
EC Full Quality Assurance System Certificate	
In accordance to Annex V, section 3 of Directive 93/42/EEC Certificate number 692.130416 Issued by: UL International (UK) Ltd. Notified Body number 0843 Valid until: 02/10/2016	
Standards applied	
EN ISO 14971:2012//EN 556-1:2001/ EN 556-1:2001/AC: 2006// EN ISO 11135-1:2007//EN 868-5:2009//ISO-2859-1:1999// EN ISO 15223-1:2012.	

Buffered iontophoresis electrodes	
Description	Buffered iontophoresis electrode treatment kit
Commercial brand	DORMO-ION
References	IE-SS15, IE-B20, IE-MS25, IE-LS40
Classification	
Product class IIa – Non-sterile. According to Rule 11 of Annex IX of the 93/42/EEC Directive	
GMDN	45141
UMDNS	-4494 ó 4495
EC Full Quality Assurance System Certificate	
In accordance to Annex II, section 3 of Directive 93/42/EEC Certificate number 572.170827 Issued by: UL International (UK) Ltd. Notified Body number 0843 Valid until: 26/08/2022	
Standards applied	
EN ISO 14971:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-5:2009// EN ISO 10993-10:2013// EN ISO 15223-1:2012// EN 1041:2008+A1:2013. 2011/65/EU Directive (RoHS 2).	

EC DECLARATION OF CONFORMITY – ANNEX 2

List of self-certified products

ECG electrodes and accessories	
Description	ECG electrodes Ag/AgCl.
Commercial brand	DORMO
References	LEH-36, LF-50, LF-36, LP-50, LR-50, SX-50, SX-36, SF-36, SX-30, SP-50
Description	Connection adapter to standard stud.
Commercial brand	DORMO
References	4013-5, 4013-B
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	35035
UMDNS	-26935
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC : 2010// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC : 2009// EN ISO 10993-10:2013. ANSI/AAMI EC12 Disposable ECG electrodes. 2011/65/EU Directive (RoHS 2).	

Neonatal ECG electrodes	
Description	Neonatal ECG electrodes Ag/AgCl.
Commercial brand	DORMO
References	K-140, KS-140, K-150, KS-150, KF-140, KFS-140, KF-150, KFS-150, EKf-22KT, P-40/80 (non pre-gelled)
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	17460
UMDNS	-26935
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC : 2010// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC : 2009// EN ISO 10993-10:2013. ANSI/AAMI EC12 Disposable ECG electrodes. 2011/65/EU Directive (RoHS 2)	

Resting electrodes and accessories	
Description	Resting electrodes.
Commercial brand	DORMO-TAB
References	T-2226
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	35035
UMDNS	17191
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC: 2009// EN ISO 10993-10:2013// EN ISO 15223-1:2012. ANSI/AAMI EC12 Disposable ECG electrodes. 2011/65/EU Directive (RoHS 2)	

TENS electrodes and recharges	
Description	Pre-gelled electrode for electrical stimulation.
Commercial brand	DORMO-TENS
References	DT-30R, DT-50R, DT-30, DT-50, DT-100 RT-30R, RT-50R, RT-30, RT-50, RT-100 T-1005, T-5055
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	35995
UMDNS	-4558 ó 17191
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC: 2009// EN ISO 10993-10:2013// EN ISO 15223-1:2012. ANSI/AAMI EC12:2000 2011/65/EU Directive (RoHS 2)	

Reusable cables for electrosurgery	
Description	Reusable clamp-cables for electrosurgical plates.
Commercial brand	BLAYCO
References	4200, 4200-5, 4210, 4210-5, 4212, 4212-5
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	47487
UMDNS	11496
Standards applied	

EN ISO 14971:2012// EN 60601-2-2:2009/ EN 60601-2-2:2009/A11:2011// EN ISO 15223-1:2012. 2011/65/EU Directive (RoHS 2).

Bite-blocks

Description	Bite block for endotracheal tubes and laryngeal masks.
Commercial brand	MORDEDOR-MO
References	7600, 7650

Classification

Product class I – Non-sterile. According to Rule 5 of Annex IX of the 93/42/EEC Directive

GMDN	10405
UMDNS	10405

Standards applied

EN ISO 9001:2015// EN 46002:1996.

Otoscope speculum

Description	Disposable speculum for otoscope.
Commercial brand	DORMO-SPEC
References	4010, 4020, 4030, 4040, 4050, 4060, 4070, 4080, 4090, 4095

Classification

Product class I – Non-sterile. According to Rule 5 of Annex IX of the 93/42/EEC Directive

GMDN	35348
UMDNS	12849

Standards applied

EN ISO 14971:2012// EN ISO 15223-1:2012.

Protective pad

Description	Protective pad for surgical interventions.
Commercial brand	BLAYCO-PAD
References	AC-3020

Classification

Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive

GMDN	62789
UMDNS	-27909

Standards applied

EN ISO 14971:2012// EN 556-1:2001/ EN 556-1:2001/AC: 2006// EN ISO 11135-1:2007// EN ISO 14644-1:2015// EN ISO 14644-3:2005// EN ISO 14644-4:2001// EN ISO 14644-5:2004// EN ISO 15223-1:2012.

Nasal holder for gastric catheters	
Description	Nasal holder for gastric catheters.
Commercial brand	DORMO-NAS
References	7500, 7525, 7550
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	62581
UMDNS	-28865
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012.	

Cold/hot packs	
Description	Reusable pack for Cold/Hot.
Commercial brand	DORMO, OXD
References	FC-01, FC-02
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	37240
UMDNS	-17863
Standards applied	
EN-ISO 14971:2012// EN ISO 15223-1:2012.	

Ultrasound gels	
Description	Ultrasound transmission gel.
Commercial brand	TRANSONIC GEL, OXD BLUE
References	G-15, G-15/05, G-15/1, G-15/5, G-15/5RB, G-15A US-B250, US-B1, US-B5F, US-B5R
Description	Ultrasound transmission gel.
Commercial brand	TRANSONIC GEL CLEAR, OXD CLEAR
References	GC-15, GC-15/05, GC-15/1, GC-15/5, GC-15/5RB US-C250, US-C1, US-C5F, US-C5R
Classification	
Product class I – Non-sterile. According to Rule 5 of Annex IX of the 93/42/EEC Directive	
GMDN	15321
UMDNS	-17912
Standards applied	
EN ISO 14971:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-4:2009// EN ISO 10993-5:2009// EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN ISO 15223-1:2012.	

ECG Gel	
Description	Conductive gel for electrodes.
Commercial brand	ELECTRO-GEL
References	G-10, G-10A
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	11425
UMDNS	-480
Standards applied	
EN ISO 14971:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-4:2009// EN ISO 10993-5:2009// EN ISO 10993-7:2008/EN ISO 10993-7:2008/AC: 2009// EN ISO 10993-10:2013// EN ISO 10993-11:2009// EN ISO 15223-1:2012. ANSI/AAMI EC12:2000.	

Lubricating gel	
Description	Lubricating water-soluble gel
Commercial brand	DORMO
References	G-20
Classification	
Product class I – Non-sterile. According to Rule 5 of Annex IX of the 93/42/EEC Directive	
GMDN	33587
UMDNS	-21069
Standards applied	
EN ISO 14971:2012// EN ISO 10993-1:2009/EN ISO 10993-1:2009/AC: 2010// EN ISO 10993-5:2009// EN ISO 10993-10:2013// EN ISO 15223-1:2012.	

Paraffin for rehabilitation	
Description	Paraffin 48°C / 118.4°F – Rehabilitation.
Commercial brand	DORMO-PARAFFIN
References	5 kg, 2 kg, 1 kg
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	35232
UMDNS	-25297
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012.	

Paraffin for anatomical pathology	
Description	Paraffin in pearls 58°C / 136.4°F – Anatomical Pathology.
Commercial brand	DORMO-PARAFFIN
References	5 kg
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	57738
UMDNS	-22669
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012.	

Adhesive bandages for kinesiology	
Description	Elastic adhesive bandage (Kinesiologic Tape)
Commercial brand	OXD
References	Blue, Beige, Black, Pink, Green
Classification	
Product class I – Non-sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive	
GMDN	33521
UMDNS	25609
Standards applied	
EN ISO 14971:2012// EN ISO 15223-1:2012.	

EC DECLARATION OF CONFORMITY – ANNEX 3
Article 12 Medical Devices Directive 93/42/EEC for Procedure Pack

In accordance with Article 12 of Directive 93/42/EEC for Procedure Packs of articles with their own CE certificate.

Electrosurgical pencil with electrode tip cleaner	
Description	Electrosurgical pencil, hand control, with 70mm blade electrode and Electrode tip cleaner.
Commercial brand	BLAYCO
References	MB-200
Classification	
	<ul style="list-style-type: none"> • <i>Electrosurgical pencil:</i> Product class IIb – Sterile. According to Rule 9 of Annex IX of the 93/42/EEC Directive • <i>Electrode tip cleaner:</i> Product class I – Sterile. According to Rule 1 of Annex IX of the 93/42/EEC Directive • Pack MB-200 Non-sterile.
GMDN	Electrosurgical pencil: 61869 Electrode tip cleaner: 37483
UMDNS	Electrosurgical pencil: 16860 Electrode tip cleaner: 15013
Standards applied	
EN ISO 14971:2012// EN 60601-1:2006/EN 60601-1:2006/A1:2013// EN 60601-2-2:2009//EN 60601-2-2:2009/A11:2011//EN ISO 15223-1:2012. 2011/65/EU Directive (RoHS 2).	



CERTIFICATE OF REGISTRATION

TELIC, S.A.U.

Polígono Industrial Can Barri
C/ Molí d'en Barri, 7
Bigues i Riells (Barcelona) 08415 SPAIN

UL LLC®(UL) issues this certificate to the Firm named above, after assessing the Firm's quality system and finding it in compliance with:

ISO 13485:2016

EN ISO 13485:2016

Design and manufacture of external defibrillator electrode, paediatric, single-use; multifunction cardiac electrode, adult; external defibrillator electrode pad; electrosurgical return electrode, single-use, non-sterile; vein stripper, single-use, sterile; operating/examination/treatment light handle cover, single-use, sterile; coupling gel, sterile.

Manufacture of sterile surgical vessel loop.



Authorized by

Deborah Jennings-Conner
Global Regulatory Director
Life and Health Sciences, UL LLC



Check Certificate
Status: [here](#)

File Number A17128
Certificate Number 1761.210215
Initial Issue Date July 28, 2018

Cycle Start August 27, 2020
Effective Date February 15, 2021
Expiry Date August 26, 2021

This quality system registration is included in UL's Directory of Registered Firms and applies to the provision of goods and/or services as specified in the scope of registration from the address(es) shown above. By issuance of this certificate the firm represents that it will maintain its registration in accordance with the applicable requirements. This certificate is not transferable and remains the property of UL LLC.



UL LLC
333 Pfingsten Road
Northbrook, IL 60062-2096 USA