

 Shandong Haidike Medical Products Co.,Ltd.	Version number	A/5
	File No.	HDK-CE-001-13
	Effective date	2018.12.28

Declaration of Conformity

Manufacturer:

Name: Shandong Haidike Medical Products Co.,Ltd.
Add: Plant No. 1, Science and Technology
Enterprise Incubator Park, Shan County, Heze City,
Shandong Province, China
TEL: +86 530-4660062
FAX: +86 530-4660055

European Authorized Representative:

Name: CMC Medical Devices & Drugs S.L.
Add: C/ Horacio Lengo Nº 18, CP 29006,
Málaga, Spain
Tel: +34 951 214 054
Contact: Manuel Mateos
Email: info@cmcmedicaldevices.com

Product: Absorbable surgical suture

Medical device: POLYGLYCOLIC ACID

Brand name: RTMED

Specifications:

Absorbable surgical suture	PGA
Sizes	8/0, 7/0, 6/0, 5/0, 4/0, 3/0, 2/0, 0, 1, 2, 3, 4, 5, 6
Raw material of the suture	Polyglycolic acid
Structure	Multifilament
Coated	Polycaprolactone + Calcium Stearate
Needle radian	1/2 circle, 3/8 circle, 4/9 circle, straight
Needle Shape	Round body, cutting, reverse cutting, spatula
Needle diameter × chord length (0.1mm×mm)	(1.5-15)×(4.5-55)

Manufacturer's Name: Shandong Haidike Medical Products Co.,Ltd.

Manufacture's Address: Plant No. 1, Science and Technology Enterprise Incubator Park, Shan County,
Heze City, Shandong Province, China

EU Representative: CMC Medical Devices & Drugs S.L.

Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 concerning medical devices
(MDD 93/42/EEC).

Classification (MDD, Annex IX): PGA is a long-term implanted device belonging to Class III, according to

 Shandong Haidike Medical Products Co.,Ltd.	Version number	A/5
	File No.	HDK-CE-001-13
	Effective date	2018.12.28

Article 8 of the MDD93/42/EEC IX Classification Guidelines.

Rule 8: Implanted tissue for more than 30 days

Conformity Routes: Full Quality Assurance System Medical Devices Directive 93/42/EEC Annex II (4).

For the evaluation of the conformity with this Directive, the following standards have been applied:

EN ISO 13485:2016	EN ISO10993-6-2009
93/42/EEC	EN ISO10993-7-2008 (AC:2009)
MEDDEV 2.12/1 rev.8	EN ISO10993-9-2009
EN ISO 14971:2012	EN ISO10993-10-2013
MEDDEV 2.7.1:2016	EN ISO10993-11-2009
ISO15223-1:2012	YY 1116-2010
EN 1041:2008	ASTM F1980-07 (2011)
ISO15223-2:2012	YY/T 0043-2016
EN ISO10993-1:2018	EP 9.0
EN ISO 11607-1:2009+A1:2014	EN ISO11737-1:2018
EN ISO 11607-2:2006+A1:2014	ISO11737-2:2019
EN ISO10993-3-2014	EN 556-1:2001 (AC:2006)
EN ISO10993-4-2009	EN ISO 11135-1:2014
EN ISO10993-5-2009	

The products are covered by CE Certificate Number: M.2019.106.11727

Identification of Notified Body: UDEM 2292


GMDN: 13908

Registration date: 10.04.2019

Expiry date of the Certificate: 09.04.2024

We herewith declare that the above mentioned products meet the provisions of the following EC Council Directives and Standards.

Name, Surname:



Position/Title: Managing Director

Issued Date: Dec. 28th, 2018

Shandong Haidike Medical Products Co.,Ltd.

