

## Specificatii tehnice

Numărul procedurii de achiziție: 197/GD/T_Biobase din 13.10.2023
Obiectul achiziției: Achiziții de Hepafiltre pentru hotele de siguranță biologică, Biobase

Nr. Lot	Denumirea bunurilor	Specificarea tehnică deplină solicitată de către autoritatea contractantă	Specificarea tehnică deplină propusă de către ofertant
1	2	6	7
1	Filter for BSC class II A2, Biobase BSC-3FA2	BSC serial #EU3FAIA2010002 Biobase; model BSC-3FA2; man. year 2020	<p>Filter set for BSC class II A2, Biobase BSC-3FA2 serial #EU3FAIA2010002, man. year 2020. Setul este format din 2 filtre:</p> <p><b>1 buc. Supply filter H14 dimensions 910*480*69mm</b> Filter Class acc. to ISO 29463: ISO 45 H Filter Class acc. to EN1822: H14 Min. Efficiency at MPPS (%): 99,995 Frame: Aluminium Media: Glass Fiber Pleat Height: 50 Seperator: Hotmelt Max. Operating Temp. (C) : 80 Filtration Area (m²): 11.1 Nominal Airflow (m³/h) : 704 Initial Pressure Drop (Pa) : 125 Faceguard: On clean air side Gasket Pu Foam</p> <p><b>1 buc. Exhaust filter H14 dimensions 600*430*69mm</b> Filter Class acc. to ISO 29463: ISO 45 H Filter Class acc. to EN1822: H14 Min. Efficiency at MPPS (%): 99,995 Frame: Aluminium Media: Glass Fiber Pleat Height: 50 Seperator: Hotmelt Max. Operating Temp. (C) : 80 Filtration Area (m²): 6.5 Nominal Airflow (m³/h) : 416 Initial Pressure Drop (Pa) : 125 Faceguard: On clean air side Gasket Pu Foam</p>

**Semnat:** \_\_\_\_\_

**Numele, prenumele:** Antonova Anna

**În calitate de:** Administrator

**Ofertantul:** MedPunkt SRL

**mun.Chisinau, str.N.Testemiteanu 11, of.59**