

# Technical Specification

Product Name	Transfer pipette			
Code	Length	Drop	Material	Package
MF7110116042S	160mm, 1ml	42ul	PE	Individual pack, 5000pcs/ctn
MF7110115048S	150mm, 2ml	48ul	PE	Individual pack, 5000pcs/ctn
Specification	1ML, 2ML			
Items of inspection	Technical requirements			
Outward appearance	Translucent and smooth. Enough intensity.			
Size	Should conform to the requirements of drawings			
Leakage test	No leaking hole and crack. No leakage after watering.			
Diaphaneity	Water line in the container should be clear by eyes under natural light.			
Sterilization	Yes, Ethylene Oxide			

Ningbo 2022-12-09  
Place, date

宁波蓝柏医疗器械有限公司  
NINGBO MFLAB MEDICAL INSTRUMENTS CO., LTD  
Yazeng Li General Manager  
Name and function

李雅增