

## CERTIFICATE OF ANALYSIS

Certificado de Análisis / Certificat de contrôle / Chargen-Prüfprotokoll

Pourquier® Brucellosis Positive Control[Brucellosis Positive Control]

Product code :	P00130	Expiration date :	22-JUN-2020
Batch :	148		
Tests per kit :	/		
Insert version :	/	Manufacturing date :	JUN-2015

Reconstitute\* with 1ml distilled water

\* After reconstitution, the control serum must be stored at a temperature  $\leq -16^{\circ}\text{C}$  for up to 1 year

He can be frozen and unfrozen up to three times

Control Test	Spécifications	Results
EAT* :	Positive : Agglutination	Positive
SAW* :	Positive ( $\geq 30$ UI/ml)	60 IU/ml [30-120]
FC* :	Positive ( $\geq 20$ ICFTU/ml)	373 ICFTU/ml [187 - 747]
ELISA* :	Positive	Positive

\*Method according to the French standards AFNOR NF U 47-003, NF U 47-004, NF U 47-021, NF U 47-019

This product was performance tested and has met all quality control specifications required for release.

This information is released by :

Title :

Quality Assurance Specialist I, Mauran Rachel

Signature :



Manufacturer: IDEXX Montpellier SAS - 326, rue de la Galéra - Parc Euromédecine - 34090 Montpellier - France  
Tél. : 33(0) 4 99 23 24 25 - Fax : 33(0) 4 67 04 20 25 - info-montpellier@idexx.com  
EN 05-01-17 - Vers : C

Document printed on : 22-Jul-2015