

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : www.himedialabs.com, Email : info@himedialabs.com

Certificate of Analysis , Quality and Conformity

Material Code : SD057	Material Name : Norfloxacin NX 10 mcg	Lot No : 0000418178
AR No.: 040000984541	Date of Report : 2020-01-08	Exp. Date : 2021-12

TEST	SPECIFICATIONS	RESULTS
<u>Appearance</u> 1 Appearance	Filter paper discs of 6 mm diameter with printed "NX 10" on centre of each side of the disc.	Complies
<u>Cultural response</u> 1 Cultural response	Average diameter of zone of inhibition observed on Mueller Hinton Agar (M173) after 18 hours incubation at 35-37°C for standard cultures.	
<u>Escherichia coli ATCC 25922</u> 1 Zone of inhibition (mm)	28mm - 35mm	34mm
<u>Staphylococcus aureus ATCC 25923</u> 1 Zone of inhibition (mm)	17mm - 28mm	27mm
<u>Pseudomonas aeruginosa ATCC 27853</u> 1 Zone of inhibition (mm)	22mm - 29mm	28mm
<u>Staphylococcus aureus ATCC 29213</u> 1 Zone of inhibition (mm)*	18mm - 24mm	22mm
<u>Enterococcus faecalis ATCC 29212</u> 1 Zone of inhibition (mm)*	16mm - 22mm	20mm

The performance of this batch has been tested and standardized as per CLSI as well as EUCAST Standards. *QC Limit as per EUCAST Standard only.

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any

HiMedia Laboratories Private Limited




23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : www.himedialabs.com, Email : info@himedialabs.com

Material Code : SD057	Material Name : Norfloxacin NX 10 mcg	Lot No : 0000418178
AR No.: 040000984541	Date of Report : 2020-01-08	Exp. Date : 2021-12

TEST	SPECIFICATIONS	RESULTS
------	----------------	---------

particulars use, other than that specified in the current HiMedia manual or product sheets.

This document has been produced electronically and is valid
Sukanya Deshmukh**Microbiologist/Sr.Executive Microbiologist**
Ujjwala M. Kokate**Asst./Dy/QC Manager**
Dr. Santosh Kaul**Dy/QA Manager**