## EG-KONFORMITÄTSERKLÄRUNG · EC DECLARATION OF CONFORMITY DÉCLARATION CE DE CONFORMITÉ · DICHIARAZIONE CE DI CONFORMITÀ

Name und Adresse des Herstellers: / Name and address of the manufacturer: / Nom et adresse du fabricant: / Nome e indirizzo del fabbricante: BOEN HEALTHCARE CO., LTD Unit 602, International Center, No.535, Shenxu Road, Suzhou, 215021, Jiangsu, China

Wir erklären in alleiniger Verantwortung, dass / We declare under our sole responsibility that / Nous déclarons sous notre propre responsabilité que / Dichiariamo sotto la sola responsabilità che

das Medizinprodukt: / the medical device: / le dispositif médical: / il dispositivo medico:

### **Gilson Pipette Tips**

der Klasse: / of class: / de la classe: / di classe: Common/Others IVD (Devices of NOT Annex II and NOT self-test)

(IVDD, Artikel 9 Absatz 1) nicht Teil der Liste A und B von Anhang II sein / (IVDD, Article9(1)) not be part of list A & B of annex II (IVDD, article 9, paragraphe 1) ne fait pas partie de la liste A et B de l'annexe II / (IVDD, articolo 9, paragrafo 1) non fanno parte dell'elenco A e B dell'allegato II

den einschlägigen Bestimmungen der Medizinprodukte-Richtlinie 98/79/EG und deren Umsetzungen in nationale Gesetze entspricht. Die Erklärung gilt in Verbindung mit dem zum Produkt gehörigen "Endprüfprotokoll". /

meets the provisions of the directive 98/79/EC and its transpositions in national laws which apply to it. The declaration is valid in connection with the "final inspection report" of the device. /

remplit toutes les exigences de la directive sur les dispositifs médicaux 98/79/EC et de ses transpositions en droit national qui le concernent. La déclaration est valable si elle est associée au «rapport de l'inspection finale» du produit. /

soddisfa tutte le disposizioni della direttiva 98/79/EC e della loro trasposizione nel diritto nazionale che lo riguardano. Questa dichiarazione è valida in congiunzione con il "rapporto di ispezione finale" del prodotto.

Konformitätsbewertungsverfahren: / Conformity assessment procedure: / Procédure d'évaluation de la conformité: / Procedura di valutazione della conformità: Anhang III (voraussichtlicher Punkt 6) der IVDD 98/79 / EG Annex III (expect point 6) of IVDD 98/79/EC Annexe III (sauf le point 6) de l'IVDD 98/79 / CE Allegato III (aspettarsi il punto 6) dell'IVDD 98/79 / CE

Registrier-Nr.: / Registration No.: / N°d'enregistrement: / Numero di registrazione:

Benannte Stelle: / Notified Body: / Organisme notifié: / Organismo notificato:

Suzhou, 201.05.26

Ort, Datum / Place, date / Lieu, date / Luogo, data





## EG-KONFORMITÄTSERKLÄRUNG · EC DECLARATION OF CONFORMITY DÉCLARATION CE DE CONFORMITÉ · DICHIARAZIONE CE DI CONFORMITÀ

Name und Adresse des Herstellers: / Name and address of the manufacturer: / Nom et adresse du fabricant: / Nome e indirizzo del fabbricante: BOEN HEALTHCARE CO., LTD Unit 602, International Center, No.535, Shenxu Road, Suzhou, 215021, Jiangsu, China

Wir erklären in alleiniger Verantwortung, dass / We declare under our sole responsibility that / Nous déclarons sous notre propre responsabilité que / Dichiariamo sotto la sola responsabilità che

das Medizinprodukt: / the medical device: / le dispositif médical: / il dispositivo medico:

### **Pasteur Pipette**

der Klasse: / of class: / de la classe: / di classe: Common/Others IVD (Devices of NOT Annex II and NOT self-test)

(IVDD, Artikel 9 Absatz 1) nicht Teil der Liste A und B von Anhang II sein / (IVDD, Article9(1)) not be part of list A & B of annex II (IVDD, article 9, paragraphe 1) ne fait pas partie de la liste A et B de l'annexe II / (IVDD, articolo 9, paragrafo 1) non fanno parte dell'elenco A e B dell'allegato II

den einschlägigen Bestimmungen der Medizinprodukte-Richtlinie 98/79/EG und deren Umsetzungen in nationale Gesetze entspricht. Die Erklärung gilt in Verbindung mit dem zum Produkt gehörigen "Endprüfprotokoll". /

meets the provisions of the directive 98/79/EC and its transpositions in national laws which apply to it. The declaration is valid in connection with the "final inspection report" of the device. /

remplit toutes les exigences de la directive sur les dispositifs médicaux 98/79/EC et de ses transpositions en droit national qui le concernent. La déclaration est valable si elle est associée au «rapport de l'inspection finale» du produit. /

soddisfa tutte le disposizioni della direttiva 98/79/EC e della loro trasposizione nel diritto nazionale che lo riguardano. Questa dichiarazione è valida in congiunzione con il "rapporto di ispezione finale" del prodotto.

Konformitätsbewertungsverfahren: / Conformity assessment procedure: / Procédure d'évaluation de la conformité: / Procedura di valutazione della conformità: Anhang III (voraussichtlicher Punkt 6) der IVDD 98/79 / EG Annex III (expect point 6) of IVDD 98/79/EC Annexe III (sauf le point 6) de l'IVDD 98/79 / CE Allegato III (aspettarsi il punto 6) dell'IVDD 98/79 / CE

Registrier-Nr.: / Registration No.: / N°d'enregistrement: / Numero di registrazione:

Benannte Stelle: / Notified Body: / Organisme notifié: / Organismo notificato:

Suzhou, 2022.01.01

Ort, Datum / Place, date / Lieu, date / Luogo, data







### according to Directive 98/79/EC of the European Parliament and of the council of 27 October 1998 on *in vitro* diagnostic medical devices

# in combination with article 110 of Regulation (EU) 2017/746 of the European Parliament and of the Council of 5 April 2017

We, DRG Instruments GmbH, Frauenbergstr. 18, 35039 Marburg, Germany

herewith declare under our own responsibility that the products listed in the table below,

### which are classified as miscellaneous products under Directive 98/79/EC

are in accordance with the requirements of the IVD Directive 98/79/EC annex I and III of the European Parliament in regard to the in vitro diagnostic medical devices (IVDs) and therefore are allowed to be CE marked.

#### **Quality Management**

DRG Instruments GmbH has established a Quality Management System for the design/development, production and distribution of in vitro diagnostic according to **EN ISO 13485:2016**.

#### Disclaimer

This Declaration of Conformity was created after the Date of Application of IVDR (May 26<sup>th</sup>, 2022). However, the cause for the creation of this Declaration of Conformity is <u>NOT</u> considered a "significant change" according to IVDR in combination with MDCG 2022-6 and therefore, the device continues to comply to IVDD.

Marburg, 2023-10-23

Dr. Matthias Herkert PRRC

Product Name	DRG CatNo	
FSH ELISA	EIA-1288	
LH-Serum ELISA	EIA-1289	
Prolactin ELISA	EIA-1291	
17-OH Progesterone ELISA	EIA-1292	
C-Peptide ELISA	EIA-1293	
AFP	EIA-1468	
HCG ELISA	EIA-1469	a .
Testosterone ELISA	EIA-1559	
Progesterone ELISA	EIA-1561	



### according to Directive 98/79/EC of the European Parliament and of the council of 27 October 1998 on *in vitro* diagnostic medical devices

Product Name	DRG CatNo
DHEA-S ELISA	EIA-1562
Free Estriol ELISA	EIA-1612
Immunoglobulin E (IgE) ELISA	EIA-1788
Sperm Antibody ELISA	EIA-1826
TGF-B1 ELISA	EIA-1864
CEA ELISA	EIA-1868
Cortisol ELISA	EIA-1887
Anti-GAD ELISA	EIA-1910
ß-HCG ELISA	EIA-1911
Helicobacter pylori IgM ELISA	EIA-2111
TGF-ß2 ELISA	EIA-2396
Estradiol ELISA	EIA-2693
Free Testosterone ELISA	EIA-2924
Insulin ELISA	EIA-2935
SHBG ELISA	EIA-2996
CIC C1q ELISA	EIA-3169
Androstenedione ELISA, 96 well	EIA-3265
Osteocalcin ELISA	EIA-3375
Thyroglobulin ELISA	EIA-3377
Adenovirus IgA ELISA	EIA-3445
Adenovirus IgG ELISA	EIA-3446
Adenovirus IgM ELISA	EIA-3447
Bordetella pertussis IgA ELISA, 96 well	EIA-3449
Bordetella pertussis IgG ELISA	EIA-3450
Bordetella pertussis IgM ELISA	EIA-3451
Brucella IgG ELISA, 96 Well	EIA-3455
Brucella IgM ELISA	EIA-3456
Candida albicans IgA ELISA	EIA-3457
Candida albicans IgG ELISA	EIA-3458
Candida albicans IgM ELISA	EIA-3459
Giardia lamblia Ag (stool) ELISA,	EIA-3477
Helicobacter pylori IgA ELISA	EIA-3483
Helicobacter pylori IgG ELISA	EIA-3484
HSV-1 IgG ELISA	EIA-3485



### according to Directive 98/79/EC of the European Parliament and of the council of 27 October 1998 on *in vitro* diagnostic medical devices

Product Name	DRG CatNo
HSV-1 IgM ELISA	EIA-3486
HSV-2 IgG ELISA	EIA-3487
HSV-2 IgM ELISA	EIA-3488
HSV-1+2 IgG ELISA, 96 well	EIA-3489
HSV-1+2 IgM ELISA	EIA-3490
Mycoplasma pneumoniae IgG ELISA	EIA-3499
Mycoplasma pneumoniae IgM ELISA	EIA-3500
Tetanus toxin IgG ELISA	EIA-3514
Treponema pallidum IgG ELISA	EIA-3517
hGH ELISA	EIA-3552
Anti TPO Ab ELISA	EIA-3561
Anti-Centromere B ELISA	EIA-3569
Anti-Cardiolipin IgG/IgM ELISA	EIA-3587
Anti-Phospholipid Screen IgG/IgM ELISA, 96 well	EIA-3591
Anti-Gliadin IgG ELISA, 96 well	EIA-3605
Anti-Gliadin IgA ELISA	EIA-3606
Anti-Insulin ELISA	EIA-3608
Anti-Tissue-Transglutaminase IgA ELISA	EIA-3611
ACTH (Adrenocorticotropic Hormone) ELISA	EIA-3647
Calcitonin	EIA-3648
Anti TG Ab ELISA	EIA-3708
Free T4 ELISA	EIA-3775
Free T3 ELISA	EIA-3801
Measles Virus IgG ELISA	EIA-3844
Measles Virus IgM ELISA	EIA-3845
Mumps Virus IgG ELISA	EIA-3846
Mumps Virus IgM ELISA	EIA-3847
Mycoplasma pneumoniae IgA ELISA	EIA-3848
Anti-Tissue-Transglutaminase IgG ELISA	EIA-3882
Histamine ELISA	EIA-4005
TSH ELISA	EIA-4171
EBV-EBNA-1 IgG ELISA (qualitative)	EIA-4246
EBV-EBNA-1 IgM ELISA (qualitátive)	EIA-4247
Treponema pallidum IgM ELISA	EIA-4267



### according to Directive 98/79/EC of the European Parliament and of the council of 27 October 1998 on *in vitro* diagnostic medical devices

Product Name	DRG CatNo
Ferritin ELISA	EIA-4408
EBV VCA IgA ELISA (qualitative)	EIA-4472
PLGF ELISA	EIA-4529
Mycoplasma hominis IgG ELISA	EIA-4559
Mycoplasma hominis IgM ELISA	EIA-4560
Ureaplasma urealyticum IgM ELISA	EIA-4561
T4 Total ELISA	EIA-4568
T3 Total ELISA	EIA-4569
NSE ELISA	EIA-4610
Ureaplasma urealyticum IgG ELISA	EIA-4623
TM-CA 15-3 ELISA	EIA-5068
TM-CA 19-9 ELISA	EIA-5069
TM-CYFRA 21-1 ELISA	EIA-5070
TM-CA 72-4 ELISA	EIA-5071
TM-CA 125 ELISA	EIA-5072
Mycoplasma hominis IgA ELISA	EIA-5097
Ureaplasma urealyticum IgA ELISA	EIA-5098
Calprotectin (Serum) ELISA	EIA-5111
Procalcitonin (Human) ELISA	EIA-5291
Aldosterone ELISA	EIA-5298
Anti-CCP ELISA	EIA-5653
Anti-C1q ELISA	EIA-5762
Strongyloides IgG/IgM ELISA	EIA-5812
Folic Acid ELISA	EIA-5847
Hanta Virus IgG ELISA	EIA-5858
Hanta Virus IgM ELISA	EIA-5859
Aspergillus fumigatus IgG ELISA	EIA-6129
Aspergillus fumigatus IgM ELISA	EIA-6130
AMH (Anti-Mullerian Hormone)	EIA-6141
Estriol total ELISA	EIA-6193
EBV (EA) IgG ELISA	EIA-6203
EBV (EA) IgM ELISA	EIA-6204
é end of list	



ФЕДЕРАЛЬНАЯ СЛУЖБА ПО НАДЗОРУ В СФЕРЕ ЗДРАВООХРАНЕНИЯ (РОСЗДРАВНАДЗОР)

# РЕГИСТРАЦИОННОЕ УДОСТОВЕРЕНИЕ НА МЕДИЦИНСКОЕ ИЗДЕЛИЕ

### № ФСР 2011/09957

от 30 октября 2012 года

Настоящее регистрационное удостоверение выдано Закрытое акционерное общество "ЭКОлаб", (ЗАО "ЭКОлаб"), Россия, 142530, Московская область, г. Электрогорск, ул. Буденного, д. 1 и подтверждает, что медицинское изделие Набор реагентов "Антиген кардиолипиновый для реакции микропреципитации "Сифилис-АгКЛ-РМП" по ТУ 9398-016-70423725-2010 в следующей комплектации производства Закрытое акционерное общество "ЭКОлаб", (ЗАО "ЭКОлаб"), Россия, 142530, Московская область, г. Электрогорск, ул. Буденного, д. 1 место производства:

Россия, 142530, Московская область, г. Электрогорск, ул. Буденного, д. 1

класс потенциального риска 26

вид медицинского изделия -

соответствующее регистрационному досье № 33508 от 26.09.2012

В соответствии с приказом Росздравнадзора от 30 октября 2012 года № 2280-Пр/12 и приказом от 23 июля 2013 года № 3428-Пр/13 о замене допущено к обращению на территории Воссийской Федерации.

Приложение: на 1 листе

Врио руководителя Федеральной служов по надзору в сфере здравоохранский

ОКП **93 9817** 

<sup>-</sup>М.А. Мурашко 0001831 ФЕДЕРАЛЬНАЯ СЛУЖБА ПО НАДЗОРУ В СФЕРЕ ЗДРАВООХРАНЕНИЯ (РОСЗДРАВНАДЗОР)

# ПРИЛОЖЕНИЕ К РЕГИСТРАЦИОННОМУ УДОСТОВЕРЕНИЮ НА МЕДИЦИНСКОЕ ИЗДЕЛИЕ № ФСР 2011/09957

Комплект № 1 включает в составе: - антиген кардиолипиновый (АгКЛ); - раствор холин-хлорида в 0,9 % растворе натрия хлорида.

Комплект № 2 в составе: - взвесь АгКЛ.

Приказом от 23 июля 2013 года № 3428-Пр/13 о замене допущено к обращению на территории Российской Федерации.

Врио руководителя Федеральной служи по надзору в сфере здравоохранския

30 октября 2012 года

М.А. Мурашко

0001854



### Page: 1 of 6

## **DECLARATION OF CONFORMITY**

1) <u>Manufacturer</u> (Name, department): **Monobind Inc.** 

### Address: 100 North Pointe, LAKE FOREST, CA 92630. UNITED STATES

and

### 2) European authorized representative: CEpartner4U BV,

Address: **EsdoornLaan 13, 3951DB Maarn, The NetherLands**; (on product labels printed as: CEpartner4U, EsdoornLaan 13, 3951DB Maarn, The NetherLands. www.cepartner4u.com)

3) <u>Product(s)</u> (name, type or model/batch number, etc.):

Immunoassay products; AccuBind® ELISA, AccuLite® CLIA, QSure® Control, Instruments see appendix

### 4) The product(s) described above is in conformity with:

Document No.	Title
98/79/EC	<i>In vitro</i> Diagnostic Medical Devices Directive

5) <u>Additional information</u> (Conformity procedure, Notified Body, CE certificate, Registration nr., etc.):

Conformity assessment procedure for CE marking: *In vitro* Diagnostic Medical Device Directive, Annex III

Registration nr. : NL- CA002-22758 and NL- CA002-22762

Lake Forest, USA; 2021-09-20

AShatola

(Place & date of issue (yyyy-mm-dd))

Tony Shatola; QA Director, Monobind Inc. *(name, function and signature of manufacturer)* 

## <u>Appendix</u>

### List of devices.

#### Item# ltem# ltem# Item# AccuLite® AccuBind® Instru-Risk First date of QSure® Device types EDMS code ELISA CLIA Class CE-marking ment Control Microwells Microwells Allergy & Anemia 2825-300A 2875-300A 12.07.01.02.00 Ferritin Test System I ow 2005-11-11 2825-300B 2875-300B 7525-300A 7575-300A Folate Test System 12.07.01.03.00 Low 2010-06-29 7525-300B 7575-300B 2525-300A 2575-300A 12.02.01.02.00 Immunoglobulin E (IgE) Test System Low 2005-11-11 2525-300B 2575-300B 8625-300A 8675-300A Transferrin Soluble Receptor (sTfR) Test 12.07.01.06.00 Low 2010-06-29 System 8625-300B 8675-300B 7625-300A 7675-300A Vitamin B-12 (Vit B12) Test System 12.07.02.04.00 I ow 2011-09-26 7625-300B 7675-300B 7825-300A 7875-300A Folate, Vitamin B-12 (Anemia Panel VAST) Test 12.07.01.00.00 2013-09-16 Low System 7825-300B 7875-300B Autoimmune 12775-300A 12725-300A Anti-Cyclic Citrullinated Peptide IgG (Anti-CCP 12.11.01.90.00 2019-04-03 Low 12725-300B 12775-300B IgG) Test System 1025-300A 1075-300A 12.10.03.04.00 Anti-Thyroglobulin (Anti-Tg) Test System Low 2005-11-11 1025-300B 1075-300B 1125-300A 1175-300A Anti-Thyroperoxidase (Anti-TPO) Test System 12.10.03.01.00 Low 2005-11-11 1125-300B 1175-300B **Bone Metabolism & Growth** 9325-300A 9375-300A Calcitonin Test System 12.06.03.02.00 2019-04-03 Low 9325-300B 9375-300B 1725-300A 1775-300A 12.06.04.02.00 Growth Hormone (hGH) Test System I ow 2005-11-11 1725-300B 1775-300B 9075-300A 9025-300A Parathyroid Hormone (PTH) Test System 12.06.03.13.00 2011-09-26 Low 9025-300B 9075-300B 10025-300A 10075-300A Parathyroid Hormone (PTH) 3rd & 2nd Gen 12.06.03.13.00 Low 2019-04-03 (VAST) Test System 10025-300B 10075-300B 7725-300A 7775-300A 25(OH) Vitamin D Total Direct (Vit D-Direct) 12.06.03.10.00 2017-07-05 I ow Test System 7725-300B 7775-300B **Cancer Markers** 1925-300A 1975-300A Alpha-Fetoprotein (AFP) Test System 12.03.90.01.00 2005-11-11 Low 1925-300B 1975-300B 3025-300A 3075-300A CA-125 Test System 12.03.01.06.00 2005-11-11 I ow 3025-300B 3075-300B 5625-300A 5675-300A CA 15-3 Test System 12.03.01.02.00 2010-06-29 Low 5675-300B 5625-300B 3925-300A 3975-300A CA 19-9 Test System 12.03.01.03.00 2005-11-11 I ow 3925-300B 3975-300B 1825-300A 1875-300A Carcinoembryonic Antigen (CEA) Test System 12.03.01.31.00 Low 2005-11-11 1825-300B 1875-300B Next Generation Carcinoembryonic Antigen 4625-300A 4675-300A 12.03.01.31.00 Low 2010-06-29

## Date: 2021-09-20



Declaration of Conformity

Document ref.: DoC2021 vs. 12

Page: 3 of 6

Device types	ltem# AccuBind® ELISA	Item# AccuLite® CLIA	ltem# QSure®	Item# Instru- ment	EDMS code	Risk Class	First date of CE-marking
	Microwells	Microwells	Control	mont	I		<b>. .</b>
(CEA-Next Gen) Test System	4625-300B	4675-300B					
Free $\beta$ -Subunit Human Chorionic Gonadotropin (Free Beta hCG) Test System	2025-300A 2025-300B	2075-300A 2075-300B			12.03.01.90.00	Low	2005-11-11
Cardiac Markers							
	2925-300A	2975-300A			40.40.04.00.00	1	0005 44 44
CK-MB Test System	2925-300B	2975-300B			12.13.01.02.00	LOW	2005-11-11
Digovin (DIG) Test System	925-300A	975-300A			12 08 01 01 00	Low	2005-11-11
	925-300B	975-300B			12.00.01.01.00	LOW	2003-11-11
High Sensitivity CRP (hs-CRP) Test System	3125-300A	3175-300A			12.13.01.90.00	Low	2005-11-11
	3125-300B	3175-300B					2000
Myoglobin Test System	3225-300A	3275-300A			12.13.01.05.00	Low	2005-11-11
	3225-300B	3275-300B					
Troponin I (cTnI) Test System	3825-300A	3875-300A			12.13.01.07.00	Low	2005-11-11
Distante a	3025-300B	3075-300B					
Diabetes	2725 2004	2775 2004	r		[	[	<b></b>
C-Peptide Test System	2725-300A	2775-300A			12.06.01.01.00	Low	2005-11-11
	2/25-300A	2175-300B					
Insulin Test System	2425-300A	2475-300A			12.06.01.03.00	Low	2005-11-11
	5825-300A	2110 0000					
Rapid Insulin Test System	5825-300B				12.06.01.03.00	Low	2010-06-29
	7325-300A	7375-300A					
Insulin - C-Peptide (Diabetes Panel VAST)	7325-300B	7375-300B			12.06.01.03.00	Low	2005-11-11
Endocrine	·				·		
ACTH Test System	10625-300	10675-300			12.06.04.01.00	Low	2019-04-03
Aldosterone Test System	10125-300	10175-300			12.06.02.01.00	Low	2019-04-03
Leptin Test System	10925-300	10975-300			12.06.90.17.00	Low	2019-04-03
Fertility & Prenatal						1	
	9725-300A	9775-300A	[	[			
Anti-Müllerian Hormone (AMH) Test System	9725-300B	9775-300B			12.05.02.16.00	Low	2019-04-03
	425-300A	475-300A					
Folicle Stimulating Hormone (FSH) Test System	425-300B	475-300B			12.05.01.04.00	Low	2005-11-11
B-Human Chorionic Gonadotropin (hCG) Test	825-300A	875-300A			40.05.00.05.00		0005 44 44
System	825-300B	875-300B			12.05.02.05.00	LOW	2005-11-11
B-Human Chorionic Gonadotropin Extended	8825-300A	8875-300A			12.05.02.05.00	Low	2013 00 16
Range (hCG-XR) Test System	8825-300B	8875-300B			12.03.02.03.00	LOW	2013-09-10
Rapid B-Human Chorionic Gonadotropin (Rapid	3325-300A				12 05 02 05 00	Low	2005-11-11
-hCG) Test System	3325-300B				12.00.02.00.00	LOW	2000 11 11
Inhibin A Test System	9525-300A	9575-300A			12.05.01.90.00	Low	2019-04-03
, 	9525-300B	9575-300B					
Inhibin B Test System	9625-300A	9675-300A			12.05.01.90.00	Low	2019-04-03
	9025-300A	675-300A					
Luteinizing Hormone (LH) Test System	625-300B	675-300B			12.05.01.05.00	Low	2005-11-11
Pregnancy Associated Plasma Protein – A Mass	12625-300A	12675-300A					
Units (PAPP-A Mass Units) Test System	12625-300B	12675-300B			12.05.02.10.00	Low	2017-07-05
	725-300A	775-300A			40.05.04.00.65	l .	0005 44 43
Prolactin Hormone (PRL) Test System	725-300B	775-300B			12.05.01.08.00	LOW	2005-11-11



Declaration of Conformity

Document ref.: DoC2021 vs. 12

Page: 4 of 6

Device types	ltem# AccuBind® ELISA Microwells	Item# AccuLite® CLIA Microwells	ltem# QSure® Control	ltem# Instru- ment	EDMS code	Risk Class	First date of CE-marking
Prolactin Hormone Sequential (PRLs) Test System	4425-300A 4425-300B	4475-300A 4475-300B			12.05.01.08.00	Low	2005-11-11
Human Chorionic Gonadotropin (hCG) , Human Prolactin (hPRL), Human Luteinizing Hormone (hLH), Follicle Stimulating Hormone (FSH) (Fertility Panel VAST) Test System	8325-300B 8325-300D 8325-300E	8375-300B 8375-300D 8375-300E			12.05.01.90.00	Low	2006-08-24
Alpha-Fetoprotein (AFP), Human Chorionic Gonadotropin (hCG), Unconjugated Estiol (u- E3) Triple Screen (Triple Screen Panel VAST) Test System	8525-300A 8525-300B	8575-300A 8575-300B			12.05.01.90.00	Low	2010-06-29
Infectious Diseases							
Anti-H. Pylori IgG (H. Pylori Ab IgG) Test System	1425-300A 1425-300B	1475-300A 1475-300B			15.01.04.03.00	Low	2005-11-11
Anti-H. Pylori IgM (H. Pylori Ab IgM) Test System	1525-300A 1525-300B	1575-300A 1575-300B			15.01.04.03.00	Low	2005-11-11
Anti-H. Pylori IgA (H. Pylori Ab IgA) Test System	1625-300A 1625-300B	1675-300A 1675-300B			15.01.04.03.00	Low	2005-11-11
Anti-SARS-CoV-2 (COVID-19) IgG Test System	11925-300A 11925-300B	11975-300A 11975-300B			15.04.80.90.00	Low	2020-08-25
Anti-SARS-CoV-2 (COVID-19) IgM Test System	11725-300A 11725-300B	11775-300A 11775-300B			15.04.80.90.00	Low	2020-08-25
Anti-SARS-CoV-2 (COVID-19) IgA Test System	11825-300A 11825-300B	11875-300A 11875-300B			15.04.80.90.00	Low	2020-08-25
Anti-SARS-CoV-2 (COVID-19) S1-RBD IgG Test System	12025-300A 12025-300B	12075-300A 12075-300B			15.04.80.90.00	Low	2021-09-20
D-Dimer Test System	9225-300A 9225-300B	9275-300A 9275-300B			13.02.05.03.00	Low	2020-08-25
Procalcitonin (PCT) Test System	1425-300A 1425-300B	1475-300A 1475-300B			12.06.90.16.00	Low	2017-07-05
Neonatal							
Neonatal 17OHP (N-17OHP) Test System	5525-300A 5525-300B				12.05.01.07.00	Low	2008-02-01
Neonatal (N-T4) Thyroxine Test System	2625-300A 2625-300B				12.04.01.12.00	Low	2005-11-11
Neonatal TBG (N-TBG) Test System	8925-300A 8925-300B				12.04.01.09.00	Low	2013-09-16
Neonatal TSH (N-TSH) Test System	3425-300A 3425-300B 3425-300D 3425-300E				12.04.01.90.00	Low	2005-11-11
Steroid							
Androstenedione (ANST) Test System	12425-300A 12425-300B	12475-300A 12475-300B			12.05.01.01.00	Low	2021-09-20
Cortisol Test System	3625-300A 3625-300B	3675-300A 3675-300B			12.06.02.04.00	Low	2005-11-11
Dehydroepiandrosterone (DHEA) Test System	7425-300A 7425-300B	7475-300A 7475-300B			12.05.01.02.00	Low	2011-09-26
Dehydroepiandrosterone Sulfate (DHEA-S) Test System	5125-300A 5125-300B	5175-300A 5175-300B			12.05.01.02.00	Low	2010-06-29
Estrone (E1) Test System	10325-300A 10325-300B	10375-300A 10375-300B			12.05.02.04.00	Low	2019-04-03

Monobind Inc.

Declaration of Conformity

Document ref.: DoC2021 vs. 12

Page: 5 of 6

	Item#	Item#	ltem#	ltem#			
Device types	AccuBind®	AccuLite®	QSure®	Instru-	EDMS code	Risk	First date of
	ELISA	CLIA	Control	ment		Class	CE-marking
	MICrowells		Control				
Estradiol (E2) Test System	4925-300A	4975-300A			12.05.01.03.00	Low	2010-06-29
	4925-300B	4975-300D					
Unconjugated Estiol (u-E3) Test System	5025-300A	5075-300A			12.05.02.02.00	Low	2010-06-29
	1925-300B	1975-300B					
Progesterone Test System	4625-300A	4075-300A			12.05.01.06.00	Low	2010-06-29
	4023-300B	4075-300D					
17-OH Progesterone (17-OHP) Test System	5225-300A	5275-300A			12.05.01.07.00	Low	2010-06-29
	0025-300A	9975-300A					
System	9925-300A	9975-300A			12.05.01.07.00	Low	2010-10-18
Couldana a Diadian Clabulia (CUDO) Test	9125-300A	9175-300A					
Sex Hormone Binding Globulin (SHBG) Test	9125-300B	9175-300R			12.05.01.09.00	Low	2013-09-16
	3725-300A	3775-300A					
Testosterone Test System	3725-300B	3775-300B			12.05.01.10.00	Low	2007-11-01
	5325-300A	5375-300A					
Free Testosterone Test System	5325-300B	5375-300B			12.05.01.10.00	Low	2010-06-29
Thursid	0020 0008	0010 0000					
Thyroid	125 200 4	175 2004	1				Γ
	125-300A 125-300B	175-300A 175-300B					
Total Triidothyronine (tT3) Test System	125-300D	175-300D			12.04.01.05.00	Low	2005-11-11
	125-300E	175-300E					
	1325-300A	1375-300A					
Free Trijdethyroning (FT2) Test Stystem	1325-300B	1375-300B			12 04 01 01 00	Low	2005 11 11
riee midolityfornine (113) rest Stystem	1325-300A	1375-300D			12.04.01.01.00	LOW	2005-11-11
	1325-300B	1375-300E					
Total Triidothyronine (tT3 SBS) Test System	8125-300A	8175-300A			12 04 01 01 00	Low	2010-06-29
	8125-300B	8175-300B				2011	
Rapid Total Triidothyronine (Rapid -tT3) Test	11225-300A				12.04.01.01.00	Low	2017-07-05
System	11225-300B	575 000 4					
T3-Uptake (T3U) Test System	525-300A	575-300A			12.04.01.06.00	Low	2005-11-11
	525-300B	575-300B					
	225-300A	275-300A					
Thyroxine (tT4) Test System	225-300B 225-300D	275-300B			12.04.01.07.00	Low	2005-11-11
	225-300E	275-300E					
	1225-300A	1275-300A					
	1225-300B	1275-300B					
Free Thyroxine (fT4) Test System	1225-300D	1275-300D			12.04.01.02.00	Low	2005-11-11
	1225-300E	1275-300E					
Tatal Thursving (tT4 SDS) Tast System	8225-300A	8275-300A			12 04 01 01 00	Low	2010 06 20
Total myroxine (114 SBS) Test System	8225-300B	8275-300B			12.04.01.01.00	LOW	2010-06-29
Panid Tatal Thuravina (Panid 174) Taat Suatam	11125-300A				12 04 01 01 00	Low	2017 07 05
Rapid Total Histoxine (Rapid -114) Test System	11125-300B				12.04.01.01.00	LOW	2017-07-03
	325-300A	375-300A					
Thyrotropin (TSH) Test System	325-300B	375-300B			12 04 01 11 00	Low	2005-11-11
	325-300D	375-300D			12.0 1.0 1.1 1.00	2011	2000 11 11
	325-300E	375-300E					
Rapid TSH Test System	0025-300A	6075-300A			12.04.01.11.00	Low	2010-06-29
	0025-300B	0075-300B					· · · · ·
Thyroxine-Binding Globulin (TBG) Test System	3525-300A	3575-300A			12.04.01.09.00	Low	2005-11-11
	3525-300B	3575-300B					
Thyroglobulin (Tg) Test System	2225-300A	2275-300A			12.04.01.08.00	Low	2005-11-11



Declaration of Conformity

Document ref.: DoC2021 vs. 12

Page: 6 of 6

Device types	ltem# AccuBind® ELISA Microwells	Item# AccuLite® CLIA Microwells	ltem# QSure® Control	Item# Instru- ment	EDMS code	Risk Class	First date of CE-marking
	2225-300B	2275-300B					
Total Thyroxine (tT4), Total Triidothyronine (tT3) & Thyroid Stimulating Hormone (TSH) (Thyroid Panel VAST) Test System	8025-300B 8025-300D 8025-300E	8075-300B 8075-300D 8075-300E			12.04.01.01.00	Low	2005-11-11
Free Thyroxine (fT4), Free Triiodothyronine (fT3) & Thyroid Stimulating Hormone (TSH) (Free Thyroid Panel VAST) Test System	7025-300B 7025-300D 7025-300E	7075-300B 7075-300D 7075-300E			12.04.01.01.00	Low	2010-06-29

Miscellaneous Controls				
Anti-H. Pylori Control (IgA, IgG, IgM) – Positive & Negative	HPC-300	12.50.01.16.00	Low	2013-09-16
Anti-Tg & Anti-TPO Control – Positive & Negative	AIT-101	12.50.01.16.00	Low	2010-06-29
Maternal Control – (AFP, uE3, hCG, Free beta hCG) Tri Level	MC-300	12.50.01.16.00	Low	2010-06-29
TBG Control – Tri-Level	TBG-300	12.50.01.16.00	Low	2013-09-16
Tg Control – Tri-Level	TG-300	12.50.01.16.00	Low	2010-06-29
Tumor Marker Control – (CA 125, CA 15-3, CA 19-9) Tri-Level	TMC-300	12.50.01.16.00	Low	2013-09-16

Miscellaneous Instruments					
Autoplex® ELISA & CLIA Analyzer		IN006	21.02.10.01	Low	2010-06-29
Autoplex® G2 ELISA & CLIA Analyzer		IN006-2	21.02.10.01	Low	2013-09-16
Autoplex® G3 ELISA & CLIA Analyzer		IN006-3	21.02.10.01	Low	2017-07-05
NeoEldex® ELISA Analzyer		IN009	21.02.10.01	Low	2011-09-26
Impulse® 3 CLIA Analyzer		IN007	21.02.10.01	Low	2010-06-29
NeoLumax® CLIA Analyzer		IN010	21.02.10.01	Low	2011-09-26
LuMatic® CLIA Analyzer		IN008	21.02.10.01	Low	2011-09-26
PrisMatic® ELISA Analyzer		IN013	21.02.10.01	Low	2013-09-16
PlateWash - Immunoassay Washer		IN002	21.02.10.01	Low	2010-06-29
TITIN® ELISA & CLIA Analyzer		IN015-EC	21.02.10.01	Low	2017-07-05
TITIN® ELISA Analyzer		IN015-E	21.02.10.01	Low	2017-07-05
TITIN-s® ELISA & CLIA Analyzer		IN016-EC	21.02.10.01	Low	2017-07-05
TITIN-s® ELISA Analyzer		IN016-E	21.02.10.01	Low	2017-07-05



# Certificate of Registration of Quality Management System to ISO 13485:2016

Brazil - RDC ANVISA n. 16/2013 RDC ANVISA n. 23/2012 RDC ANVISA n. 67/2009
Canada - Medical Devices Regulations - Part 1- SOR 98/282
United States- 21 CFR 803, 21 CFR 806, 21 CFR 807 - Subparts A to D,
☑ 21 CFR 820 - Quality System Regulation,

The National Standards Authority of Ireland is an MDSAP Recognized Auditing Organization and certifies that:

# Monobind Inc. 100 North Pointe Drive Lake Forest, CA 92630 USA

## Facility ID: F002818

has been assessed and deemed to comply with the requirements of the above standard and regulations in respect of the scope of operations given below:

# The Design, Manufacture, and Distribution of In-Vitro Diagnostic Medical Device Immunoassays and Related Reagents and Controls. The Distribution of Related Washers and Analyzers.

Additional sites covered under this multi-site certification are listed on the Annex (File No. MP19.4585)

Approved by: Kevin Mullaney Director of Certification

Certificate Number: MP19.4585 / Rev 2 Certification Granted: 2019/09/25 Effective Date: 2022/09/25 Expiry Date: 2025/09/24



National Standards Authority of Ireland, 1 Swift Square, Northwood, Santry, Dublin 9, Ireland T +353 1 807 3800 National Standards Authority of Ireland, 20 Trafalgar Square, Nashua, New Hampshire, NH 03063, USA T +1 603 882 4412 All valid certifications are listed on NSAI's website – www.nsaiinc.com The continued validity of this certificate may be verified under "Approved Client Listing



# Annex to Certificate Number: MP19.4585 / Rev 2

# Scope of Registration:

The Design, Manufacture, and Distribution of In-Vitro Diagnostic Medical Device Immunoassays and Related Reagents and Controls. The Distribution of Related Washers and Analyzers.

# Activity

## Location

Headquarters, Design, Manufacture	Monobind Inc. 100 North Pointe Drive Lake Forest, CA 92630 USA File No.: MP19.4585 Facility ID: F002818
Manufacture, Distribution	Monobind Inc. 103 North Pointe Drive Lake Forest, CA 92630 USA File No.: MP19.4585/A Facility ID: F002818

Verified by: Director of Certification



### EG Konformitätserklärung

## **EC Declaration of Conformity**

ORGENTEC Diagnostika GmbH Carl-Zeiss-Straße 49-51, 55129 Mainz, GERMANY

Wir erklären in eigener Verantwortung, dass das ORGENTEC Produkt We declare in our sole responsibility that the ORGENTEC product

## ORG 540A Anti-Tissue-Transglutaminase IgA

zur quantitativen in-vitro-Bestimmung bestimmt ist und entsprechend Art. 9 Abs. Satz 1 der Europäischen Richtlinie 98/79/EG als "Sonstige Produkte" (non-A, non-B, keine Selbstanwendung) klassifiziert ist.

as intended for use in quantitative in vitro determination is classified as "Other Devices" (non-A, non-B, no self-testing device) according to article 9 paragraph 1 sentence 1 of the European directive 98/79/EC.

Das Produkt stimmt mit den Grundlegenden Anforderungen und allen zutreffenden Bestimmungen der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 über in-vitro-Diagnostika überein. Die Konformität zur Richtlinie wurde durch ein Konformitätsbewertungsverfahren nach Anhang III der Richtlinie festgestellt.

This product is conform with the essential requirements and meet the appropriate provisions of the Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on in vitro diagnostic medical devices. Conformity was proved by a conformity assessment procedure referred to in annex III of the directive.

Liste angewendeter Normen:

List of standards applied for CE marking: EN ISO 13485, EN ISO 14971, EN ISO 18113, EN ISO 15223, EN ISO 23640, EN 13612.

Mainz, 2021-02-05

René Betz Head of Regulatory Affairs

Gültig ab / Valid from 2021-02-05 bis / until 2024-02-28

Notification pursuant to §25 Abs. 3 Nr. 3 Medical Devices Act, MPG Type: Reagent EDMS 12-10-90-21-00 GMDN 55223

ORG 540A\_CE declaration of conformity\_QM120359\_2021-02-05\_8

F4.01B Declaration of conformity

**ORGENTEC** Diagnostika GmbH Carl-Zeiss-Straße 49-51 55129 Mainz Deutschland

Telefon: +49 (0) 61 31/92 58-0 Telefax: +49 (0) 61 31/92 58 58 orgentec@orgentec.com www.orgentec.com

Mainzer Volksbank eG IBAN: DE72 5519 0000 0159 8000 10 BIC: MVBMDE55

Commerzbank AG IBAN: DE13 5504 0022 0200 8670 00 BIC: COBADEFFXXX

USt-IdNr. DE149058799 Mainz 14 HRB 4300

agnost

Geschäftsführer Ralf Wehen

#### **ORGENTEC Diagnostika GmbH**

Carl-Zeiss-Straße 49-51 55129 Mainz - Germany Phone: +49 (0) 61 31 / 92 58-0 Fax: +49 (0) 61 31 / 92 58-58 Internet: www.orgentec.com





### ORG 540A Anti-Tissue-Transglutaminase IgA

#### INTENDED PURPOSE

Anti-Tissue-Transglutaminase IgA is an ELISA test system for the quantitative measurement of IgA class autoantibodies to tissue transglutaminase (tTG) in human serum or plasma. This product is intended for professional in vitro diagnostic use only.

Autoantibodies against tissue transglutaminase (tTG) during a gluten containing diet are the most obvious serologic feature of coeliac disease (CD). Determination of serum levels of IgA against tTG is the first choice in suspected CD. In subjects with either primary or secondary humoral IgA deficiency, at least one additional test measuring IgG class CD-specific antibodies is recommended. The clinical relevance of a positive anti-tTG result should be confirmed by histology of the small intestinal mucosa.

#### SYMBOLS USED ON LABELS

IVD	In vitro diagnostic medical device	MICROPLATE	Microplate
	Manufacturer	CALIBRATOR A	Calibrator
		CALIBRATOR B	Calibrator
REF	Catalogue number	CALIBRATOR C	Calibrator
∑ 96	Sufficient for 96 determinations	CALIBRATOR D	Calibrator
LOT	Batch code	CALIBRATOR E	Calibrator
	Baton oodo	CALIBRATOR F	Calibrator
$\leq$	Use by	CONTROL +	Control positive
2°C	Temperature limitation	CONTROL -	Control negative
悉	Keep away from sunlight		
- ©	Do not reuse	DILUENT	Sample Buffer P
(a)	Do hot reuse	CONJUGATE	Enzyme Conjugate
M	Date of manufacture		
( f	CE marked according to 98/79/EC	ТМВ	TMB Substrate
•••	Consult instructions for use	STOP	Stop solution
Li	Consult instructions for use	WASH	Wash Buffer
540A_4	Electronic Instruction For Use: version	RTU	Ready to use

#### PRINCIPLE OF THE TEST

Human recombinant tissue transglutaminase is bound to microwells.

The determination is based on an indirect enzyme linked immune reaction with the following steps:

Specific antibodies in the patient sample bind to the antigen coated on the surface of the reaction wells. After incubation, a washing step removes unbound and unspecifically bound serum or plasma components. Subesquently added enzyme conjugate binds to the immobilized antibody-antigen-complexes. After incubation, a second washing step removes unbound enzyme conjugate. After addition of substrate solution the bound enzyme conjugate hydrolyses the substrate forming a blue coloured product. Addition of an acid stopps the reaction generating a yellow end-product. The intensity of the yellow color

correlates with the concentration of the antibody-antigen-complex and can be measured photometrically at 450 nm.

#### WARNINGS AND PRECAUTIONS

- · All reagents of this kit are intended for professional in vitro diagnostic use only.
- Components containing human serum were tested and found negative for HBsAg, HCV, HIV1 and HIV2 by FDA approved methods. No test can guarantee the absence of HBsAg, HCV, HIV1 or HIV2, and so all human serum based reagents in this kit must be handled as though capable of transmitting infection.
- Bovine serum albumin (BSA) used in components has been tested for BSE and found negative.
- Avoid contact with the substrate TMB (3,3',5,5'-Tetramethyl-benzidine).
- · Stop solution contains acid, classifiaction is non-hazardous. Avoid contact with skin.
- Control, sample buffer and wash buffer contain sodium azide 0.09% as preservative. This concentration is classified as non-hazardous.
- Enzyme conjugate contains ProClin 300 0.05% as preservative. This concentration is classified as non-hazardous.

During handling of all reagents, controls and serum samples observe the existing regulations for laboratory safety regulations and good laboratory practice:

First aid measures: In case of skin contact, immediately wash thoroughly with water and soap. Remove
contaminated clothing and shoes and wash before reuse. If system fluid comes into contact with skin,
wash thoroughly with water. After contact with the eyes carefully rinse the opened eye with running
water for at least 10 minutes. Get medical attention if necessary.

• Personal precautions, protective equipment and emergency procedures:

Observe laboratory safety regulations. Avoid contact with skin and eyes. Do not swallow. Do not pipette by mouth. Do not eat, drink, smoke or apply makeup in areas where specimens or kit reagents are handled. When spilled, absorb with an inert material and put the spilled material in an appropriate waste disposal.

- Exposure controls / personal protection: Wear protective gloves of nitril rubber or natural latex. Wear protective glasses. Used according to intended use no dangerous reactions known.
- Conditions to avoid: Since substrate solution is light-sensitive. Store in the dark.
- · For disposal of laboratory waste the national or regional legislation has to be observed.

Observe the guidelines for performing quality control in medical laboratories by assaying control sera.

		_
CONTENTS O	F THE KI	I
ORG 540A	96 🏹	Sufficient for 96 determinations
MICROPLATE	1	One divisible microplate consisting of 12 modules of 8 wells each. Ready to use. Product code on module: <i>t</i> TG
CALIBRATOR A	1x 1.5 ml	Calibrator A 0 U/ml, containing serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use.
CALIBRATOR B	1x 1.5 ml	Calibrator B 5 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use.
CALIBRATOR C	1x 1.5 ml	Calibrator C 10 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use.
CALIBRATOR D	1x 1.5 ml	Calibrator D 25 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 $0.09\%$ ), yellow. Ready to use.
CALIBRATOR E	1x 1.5 ml	Calibrator E 75 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, NaN3 0.09%), yellow. Ready to use.
CALIBRATOR F	1x 1.5 ml	Calibrator F 200 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use.
CONTROL +	1x 1.5 ml	Control positive, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use. The concentration is specified on the certificate of analysis.

- CONTROL -1x 1.5 ml Control negative, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use. The concentration is specified on the certificate of analysis.
- DILUENT Sample Buffer P. containing PBS, BSA, detergent, preservative sodium azide 0.09%. 20 ml vellow, concentrate (5 x). CONJUGATE
  - 15 ml Enzyme Conjugate: light red, containing anti-human IgA antibodies, HRP labelled: PBS. BSA, detergent, preservative PROCLIN 0.05%, Ready to use.
  - TMB Substrate; containing 3,3', 5,5'- Tetramethylbenzidin, colorless. Ready to use. 15 ml
  - Stop solution; contains acid. Ready to use. 15 ml
    - Wash Buffer, containing Tris, detergent, preservative sodium azide 0.09%; 50 x conc. 20 ml
  - Certificate of Analysis

#### MATERIALS REQUIRED

- Microplate reader capable of endpoint measurements at 450 nm; optional; reference filter at 620 nm
- Data reduction software
- Multi-channel dispenser or repeatable pipette for 100 µl
- Vortex mixer

TMB

STOP

WASH

[]]i

- Pipettes for 10 µl, 100 µl and 1000 µl
- Laboratory timing device
- · Distilled or deionised water
- Measuring cylinder for 1000 ml and 100 ml
- · Plastic container for storage of the wash solution

This ELISA assay is suitable for use on open automated ELISA processors. Each assay has to be validated on the respective automated system. Detailed information is provided upon request.

#### SPECIMEN COLLECTION, STORAGE AND HANDLING

- Collect whole blood specimens using acceptable medical techniques to avoid hemolysis.
- Allow blood to clot and separate the serum or plasma by centrifugation.
- Test serum should be clear and non-hemolyzed. Contamination by hemolysis or lipemia should be avoided, but does not interfere with this assay.
- Specimens may be refrigerated at 2-8°C for up to five days or stored at -20°C up to six months.
- · Avoid repetitive freezing and thawing of serum or plasma samples. This may result in variable loss of antibody activity.
- · Testing of heat-inactivated sera is not recommended.

#### STORAGE AND STABILITY

- Store test kit at 2-8°C in the dark.
- Do not expose reagents to heat, sun, or strong light during storage and usage.
- · Store microplate sealed and dessicated in the clip bag provided.
- · Shelf life of the unopended test kit is 18 months from day of production.
- Unopened reagents are stable until expiration of the kit. See labels for individual batch.
- Diluted Wash Buffer and Sample Buffer are stable for at least 30 days when stored at 2-8°C. We recommend consumption on the same day.

#### PROCEDURAL NOTES

- Do not use kit components beyond their expiration dates.
- Do not interchange kit components from different lots and products.
- All materials must be at room temperature (20-28°C) prior to use.
- Prepare all reagents and samples. Once started, performe the test without interruption.
- Double determinations may be done. By this means pipetting errors may become obvious.
- · Perform the assay steps only in the order indicated.
- · Always use fresh sample dilutions.
- · Pipette all reagents and samples into the bottom of the wells.
- To avoid carryover or contamination, change the pipette tip between samples and different kit controls.
- · Wash microwells thoroughly and remove the last droplets of wash buffer.
- · All incubation steps must be accurately timed.
- · Do not re-use microplate wells.

#### PREPARATION OF REAGENTS

#### WASH

Dilute the contents of one vial of the buffered wash solution concentrate (50x) with distilled or deionised water to a final volume of 1000 ml prior to use.

DILUENT

Sample Buffer P: Prior to use dilute the contents (20 ml) of one vial of sample buffer 5x concentrate with distilled or deionised water to a final volume of 100 ml.

#### Preparation of samples

Dilute patient samples 1:100 before the assay: Put 990 µl of prediluted sample buffer in a polystyrene tube and add 10 µl of sample. Mix well. Note: Calibrators / Controls are ready to use and need not be diluted.

#### **TEST PROCEDURE**

Prepare enough microplate modules for all calibrators / controls and patient samples.

- Pipette 100 µl of calibrators, controls and prediluted patient samples into the wells. Incubate for 30 minutes at room temperature (20-28 °C). Discard the contents of the microwells and wash 3 times with 300 µl of wash solution.
- Dispense 100 μl of enzyme conjugate into each well. Incubate for 15 minutes at room temperature. Discard the contents of the microwells and wash 3 times with 300 μl of wash solution.
- 3. Dispense **100** µl of TMB substrate solution into each well. Incubate for **15 minutes** at room temperature
- 4. Add 100  $\mu I$  of stop solution to each well of the modules
  - Incubate for 5 minutes at room temperature.

Read the optical density at 450 nm (reference 600-690nm) and calculate the results. The developed colour is stable for at least 30 minutes. Read during this time.

Example for a pipetting scheme:

	1	2	3	4	5	6	7	8	9	10	11	12
A	Α	P1										
в	В	P2										
c	С	P3										
D	D											
E	Е											
F	F											
G	C+											
н	C-											

P1, ... patient sample A-F calibrators C+, C- controls

#### VALIDATION

Test results are valid if the optical densities at 450 nm for calibrators / controls and the results for controls comply with the reference ranges indicated on the Certificate of Analysis enclosed in each test kit. If these quality control criteria are not met the assay run is invalid and should be repeated.

#### CALCULATION OF RESULTS

For quantitative results plot the optical density of each calibrator versus the calibrator concentration to create a calibration curve. The concentration of patient samples may then be estimated from the calibration curve by interpolation.

Using data reduction software a 4-Parameter-Fit with lin-log coordinates for optical density and concentration is the data reduction method of choice.

#### **PERFORMANCE CHARACTERISTICS**

#### Calibration

This assay system is calibrated in relative arbitrary units, since no international reference preparation is available for this assay.

#### Measuring range

The calculation range of this ELISA assay is 0 - 200 U/ml

#### Expected values

In a normal range study with samples from healthy blood donors the following ranges have been established with this ELISA assay: Cut-off 10 U/ml

#### Interpretation of results

 Negative:
 < 10 U/ml</th>

 Positive:
 ≥ 10 U/ml

#### Linearity

Patient samples containing high levels of specific antibody were serially diluted in sample buffer to demonstrate the dynamic range of the assay and the upper / lower end of linearity. Activity for each dilution was calculated from the calibration curve using a 4-Parameter-Fit with lin-log coordinates.

Sample	Dilution	Observed	Expected	O/E
Sample	Dilution	Observeu	Lybecieu	UL
		U/ml	U/ml	[%]
1	1:100	<mark>196.9</mark>	196.9	100
	1:200	103.9	98.5	105
	1:400	49.0	49.2	100
	1:800	25.2	24.6	102
	1:1600	11.3	12.3	92
2	1:100	199.3	199.3	100
	1:200	100.8	99.7	101
	1:400	49.8	49.8	100
	1:800	25.8	24.9	104
	1:1600	<mark>11.7</mark>	12.5	94

#### Limit of detection

Functional sensitivity was determined to be: 1 U/mI

#### Reproducibility

Intra-assay precision: Coefficient of variation (CV) was calculated for each of three samples from the results of 24 determinations in a single run. Results for precision-within-assay are shown in the table below. Inter-assay precision: Coefficient of variation (CV) was calculated for each of three samples from the results of 8 determinations in 5 different runs. Results for run-to-run precision are shown in the table below.

			1					
	Intra-Assay			Inter-Assay				
Sample	Mean			Sample	Mean			
	U/ml	CV %			U/ml	CV %		
1	0.9	10.4	]	1	1.3	10.6		
2	47.1	4.4		2	19.4	13.7		
3	112.8	8.7		3	108.5	11.1		

#### Interfering substances

No interference has been observed with haemolytic (up to 1000 mg/dl) or lipemic (up to 3 g/dl triglycerides) sera or plasma, or bilirubin (up to 40 mg/dl) containing sera or plasma. Nor have any interfering effects been observed with the use of anticoagulants (Citrate, EDTA, Heparine). However for practical reasons it is recommended that grossly hemolyzed or lipemic samples should be avoided.

#### Study results

oounto							
Study p	<u>n</u>	<u>n Pos</u>	<u>%</u>				
Coeliac	: disea	se			35	35	100.0
Normal	huma	n sera			90	2	2.2
ORG 540A	POS NEG	Clinical I POS 35 0	Diagnosis NEG 2 88	125			
Sensitivity:	100.0	%	90	125			
Specificity:	97.8	%					
Overall agreement:	98.4	%					

#### LIMITATIONS OF THE PROCEDURE

This assay is a diagnostic aid. A definite clinical diagnosis should not be based on the results of a single test, but should be made by the physician after all clinical and laboratory findings have been evaluated concerning the entire clinical picture of the patient. Also every decision for therapy should be taken individually.

The above pathological and normal reference ranges for antibodies in patient samples should be regarded as recommendations only. Each laboratory should establishe its own ranges according to ISO 15189 or other applicable laboratory guidelines.

#### REFERENCES

- Alessio M, Tonutti E, Brusca I, Radice A, Licini L, Sonzogni A, Florena A, Schiaffino E, Marus W, Sulfaro S, Villalta D: Correlation between IgA tissue transglutaminase antibody ratio and histological finding in celiac disease: A multicentre study. J Pediatr. Gastroenterol. Nutr. 2011.
- Cummins AG, Roberts-Thomson IC: Prevalence of celiac disease in the Asia-Pacific region. J Gastroenterol. Hepatol. 2009, 24:1347-1351.
- Giersiepen K, Lelgemann M, Stuhldreher N, Ronfani L, Husby S, Koletzko S, Korponay-Szabo IR: Accuracy of diagnostic antibody tests for coeliac disease in children: summary of an evidence report. J Pediatr. Gastroenterol. Nutr. 2012, 54:229-241.
- 4. Green PH, Cellier C: Celiac disease. N. Engl. J Med. 2007, 357:1731-1743.
- Gupta R, Reddy DN, Makharia GK, Sood A, Ramakrishna BS, Yachha SK, Thapa BR, Banerjee R, Anuradha S, Dutta U, Puri AS, Jain AK, Mulder CJ, Kumar A, Boindala S: Indian task force for celiac disease: current status. World J Gastroenterol. 2009, 15:6028-6033.
- Husby S, Koletzko S, Korponay-Szabo IR, Mearin ML, Phillips A, Shamir R, Troncone R, Giersiepen K, Branski D, Catassi C, Lelgeman M, Maki M, Ribes-Koninckx C, Ventura A, Zimmer KP: European Society for Pediatric Gastroenterology, Hepatology, and Nutrition guidelines for the diagnosis of coeliac disease. J Pediatr. Gastroenterol. Nutr. 2012, 54:136-160.
- 7. Mahadov S, Green PH: Celiac disease: a challenge for all physicians. Gastroenterol. Hepatol.(N Y) 2011, 7:554 -556.
- 8. Marietta EV, Rashtak S, Murray JA: Correlation analysis of celiac sprue tissue transglutaminase and deamidated gliadin IgG/IgA. World J Gastroenterol. 2009, 15:845-848.
- 9. Murray JA: Serodiagnosis of celiac disease. Clin. Lab. Med. 1997, 17:445-464.
- 10. Rubio-Tapia A, Murray JA: Celiac disease. Curr. Opin. Gastroenterol. 2010, 26:116-122.
- 11. Telega G, Bennet TR, Werlin S: Emerging new clinical patterns in the presentation of celiac disease. Arch. Pediatr. Adolesc. Med 2008, 162:164-168.
- 12. Tjon JM, van BJ, Koning F: Celiac disease: how complicated can it get? Immunogenetics 2010, 62:641-651.
- van der Windt DA, Jellema P, Mulder CJ, Kneepkens CM, van der Horst HE: Diagnostic testing for celiac disease among patients with abdominal symptoms: a systematic review. JAMA 2010, 303:1738-1746.
- 14. Villalta D, Tonutti E, Prause C, Koletzko S, Uhlig HH, Vermeersch P, Bossuyt X, Stern M, Laass MW, Ellis JH, Ciclitira PJ, Richter T, Daehnrich C, Schlumberger W, Mothes T: IgG antibodies against deamidated gliadin peptides for diagnosis of celiac disease in patients with IgA deficiency. Clin. Chem. 2010, 56:464-468.

#### Notice to the user (European Union):

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the EU Member State in which the user and/or the patient is established.

#### Change Control

Former version: ORG 540A\_IFU\_EN\_QM113174\_2016-05-13\_2 Reason for revision: Introduction electronic IFU on homepage





## EG Konformitätserklärung

## EC Declaration of Conformity

ORGENTEC Diagnostika GmbH Carl-Zeiss-Straße 49-51, 55129 Mainz, GERMANY

Wir erklären in eigener Verantwortung, dass das ORGENTEC Produkt

We declare in our sole responsibility that the ORGENTEC product

# ORG 540G Anti-Tissue-Transglutaminase IgG

zur quantitativen in-vitro-Bestimmung bestimmt ist und entsprechend Art. 9 Abs. Satz 1 der Europäischen Richtlinie 98/79/EG als "Sonstige Produkte" (non-A, non-B, keine Selbstanwendung) klassifiziert ist.

as intended for use in quantitative in vitro determination is classified as "Other Devices" (non-A, non-B, no self-testing device) according to article 9 paragraph 1 sentence 1 of the *European directive* 98/79/EC.

Das Produkt stimmt mit den Grundlegenden Anforderungen und allen zutreffenden Bestimmungen der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 über in-vitro-Diagnostika überein. Die Konformität zur Richtlinie wurde durch ein Konformitätsbewertungsverfahren nach Anhang III der Richtlinie festgestellt.

This product is conform with the essential requirements and meet the appropriate provisions of the Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on in vitro diagnostic medical devices. Conformity was proved by a conformity assessment procedure referred to in annex III of the directive.

Liste angewendeter Normen:

List of standards applied for CE marking: EN ISO 13485, EN ISO 14971, EN ISO 18113, EN ISO 15223, EN ISO 23640, EN 13612.

Mainz, 2021-02-05

René Betz



Head of Regulatory Affairs

Gültig ab / Valid from 2021-02-05 bis / until 2024-02-28

Notification pursuant to §25 Abs. 3 Nr. 3 Medical Devices Act, MPG Type: Reagent EDMS 12-10-90-21-00 GMDN 55223

ORG 540G\_CE declaration of conformity\_QM120360\_2021-02-05\_8

F4.01B Declaration of conformity

ORGENTEC Diagnostika GmbH Carl-Zeiss-Straße 49-51 55129 Mainz Deutschland Telefon: +49 (0) 61 31/92 58-0 Telefax: +49 (0) 61 31/92 58 58 orgentec@orgentec.com www.orgentec.com

Mainzer Volksbank eG IBAN: DE72 5519 0000 0159 8000 10 BIC: MVBMDE55 Commerzbank AG IBAN: DE13 5504 0022 0200 8670 00 BIC: COBADEFFXXX

USt-IdNr. DE149058799 0200 8670 00 Mainz 14 HRB 4300 Geschäftsführer Ralf Wehen

#### **ORGENTEC Diagnostika GmbH**

Carl-Zeiss-Straße 49-51 55129 Mainz - Germany Phone: +49 (0) 61 31 / 92 58-0 Fax: +49 (0) 61 31 / 92 58-58 Internet: www.orgentec.com





### ORG 540G Anti-Tissue-Transglutaminase IgG

#### INTENDED PURPOSE

Anti-Tissue-Transglutaminase IgG is an ELISA test system for the quantitative measurement of IgG class autoantibodies to tissue-transglutaminase (tTG) in human serum or plasma. This product is intended for professional in vitro diagnostic use only.

Autoantibodies against tissue transglutaminase (tTG) during a gluten containing diet are the most obvious serologic feature of coeliac disease (CD). Determination of serum levels of IgA against tTG is the first choice in suspected CD. In subjects with either primary or secondary humoral IgA deficiency, at least one additional test measuring IgG class CD-specific antibodies is recommended. The clinical relevance of a positive anti-tTG result should be confirmed by histology of the small intestinal mucosa.

#### SYMBOLS USED ON LABELS

IVD	In vitro diagnostic medical device	MICROPLATE	Microplate
	Manufacturer	CALIBRATOR A	Calibrator
		CALIBRATOR B	Calibrator
REF	Catalogue number	CALIBRATOR C	Calibrator
<u>∑</u> 96	Sufficient for 96 determinations	CALIBRATOR D	Calibrator
LOT	Batch code	CALIBRATOR E	Calibrator
	Bateri code	CALIBRATOR F	Calibrator
$\leq$	Use by	CONTROL +	Control positive
2°C	Temperature limitation	CONTROL -	Control negative
悉	Keep away from sunlight		
- 0	Do not rouso	DILUENT	Sample Buffer P
(a)	Do not reuse	CONJUGATE	Enzyme Conjugate
M	Date of manufacture		
( 6	CE marked according to 98/79/EC	ТМВ	TMB Substrate
~~~		STOP	Stop solution
lli	Consult instructions for use	WASH	Wash Buffer
540G_4	Electronic Instruction For Use: version	RTU	Ready to use

#### PRINCIPLE OF THE TEST

Human recombinant tissue transglutaminase is bound to microwells.

The determination is based on an indirect enzyme linked immune reaction with the following steps:

Specific antibodies in the patient sample bind to the antigen coated on the surface of the reaction wells. After incubation, a washing step removes unbound and unspecifically bound serum or plasma components. Subesquently added enzyme conjugate binds to the immobilized antibody-antigen-complexes. After incubation, a second washing step removes unbound enzyme conjugate. After addition of substrate solution the bound enzyme conjugate hydrolyses the substrate forming a blue coloured product. Addition of an acid stopps the reaction generating a yellow end-product. The intensity of the yellow color

correlates with the concentration of the antibody-antigen-complex and can be measured photometrically at 450 nm.

#### WARNINGS AND PRECAUTIONS

- · All reagents of this kit are intended for professional in vitro diagnostic use only.
- Components containing human serum were tested and found negative for HBsAg, HCV, HIV1 and HIV2 by FDA approved methods. No test can guarantee the absence of HBsAg, HCV, HIV1 or HIV2, and so all human serum based reagents in this kit must be handled as though capable of transmitting infection.
- Bovine serum albumin (BSA) used in components has been tested for BSE and found negative.
- Avoid contact with the substrate TMB (3,3',5,5'-Tetramethyl-benzidine).
- · Stop solution contains acid, classifiaction is non-hazardous. Avoid contact with skin.
- Control, sample buffer and wash buffer contain sodium azide 0.09% as preservative. This concentration is classified as non-hazardous.
- Enzyme conjugate contains ProClin 300 0.05% as preservative. This concentration is classified as non-hazardous.

During handling of all reagents, controls and serum samples observe the existing regulations for laboratory safety regulations and good laboratory practice:

First aid measures: In case of skin contact, immediately wash thoroughly with water and soap. Remove
contaminated clothing and shoes and wash before reuse. If system fluid comes into contact with skin,
wash thoroughly with water. After contact with the eyes carefully rinse the opened eye with running
water for at least 10 minutes. Get medical attention if necessary.

• Personal precautions, protective equipment and emergency procedures:

Observe laboratory safety regulations. Avoid contact with skin and eyes. Do not swallow. Do not pipette by mouth. Do not eat, drink, smoke or apply makeup in areas where specimens or kit reagents are handled. When spilled, absorb with an inert material and put the spilled material in an appropriate waste disposal.

- Exposure controls / personal protection: Wear protective gloves of nitril rubber or natural latex. Wear protective glasses. Used according to intended use no dangerous reactions known.
- · Conditions to avoid: Since substrate solution is light-sensitive. Store in the dark.
- · For disposal of laboratory waste the national or regional legislation has to be observed.

Observe the guidelines for performing quality control in medical laboratories by assaying control sera.

CONTENTS O	F THE KI	Т
ORG 540G	∑ 96	Sufficient for 96 determinations
MICROPLATE	1	One divisible microplate consisting of 12 modules of 8 wells each. Ready to use. Product code on module: <b>tTG</b>
CALIBRATOR A	1x 1.5 ml	Calibrator A 0 U/ml, containing serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use.
CALIBRATOR B	1x 1.5 ml	Calibrator B 5 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use.
CALIBRATOR C	1x 1.5 ml	Calibrator C 10 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use.
CALIBRATOR D	1x 1.5 ml	Calibrator D 25 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use.
CALIBRATOR E	1x 1.5 ml	Calibrator E 75 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, NaN3 $0.09\%),$ yellow. Ready to use.

- CALIBRATOR F 1x 1.5 ml Calibrator F 200 U/ml, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use.
- CONTROL + 1x 1.5 ml Control positive, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use. The concentration is specified on the certificate of analysis.
- CONTROL -1x 1.5 ml Control negative, containing tTG antibodies in a serum/buffer matrix (PBS, BSA, detergent, NaN3 0.09%), yellow. Ready to use. The concentration is specified on the certificate of analysis.
- Sample Buffer P, containing PBS, BSA, detergent, preservative sodium azide 0.09%, DILUENT 20 ml vellow, concentrate (5 x).
- CONJUGATE Enzyme Conjugate containing anti-human IgG antibodies, HRP labelled; PBS, BSA, 15 ml detergent, preservative PROCLIN 0.05%, light red, Ready to use.
  - TMB Substrate; containing 3,3', 5,5'- Tetramethylbenzidin, colorless. Ready to use. 15 ml
  - Stop solution; contains acid. Ready to use. 15 ml
    - Wash Buffer, containing Tris, detergent, preservative sodium azide 0.09%; 50 x conc. 20 ml
  - Certificate of Analysis

#### MATERIALS REQUIRED

- · Microplate reader capable of endpoint measurements at 450 nm; optional: reference filter at 620 nm
- Data reduction software
- Multi-channel dispenser or repeatable pipette for 100 µl
- Vortex mixer

TMB

STOP

WASH

[]]i

- Pipettes for 10 µl, 100 µl and 1000 µl
- Laboratory timing device
- · Distilled or deionised water
- Measuring cylinder for 1000 ml and 100 ml
- · Plastic container for storage of the wash solution

This ELISA assay is suitable for use on open automated ELISA processors. Each assay has to be validated on the respective automated system. Detailed information is provided upon request.

#### SPECIMEN COLLECTION, STORAGE AND HANDLING

- · Collect whole blood specimens using acceptable medical techniques to avoid hemolysis.
- Allow blood to clot and separate the serum or plasma by centrifugation.
- Test serum should be clear and non-hemolyzed. Contamination by hemolysis or lipemia should be avoided, but does not interfere with this assay.
- Specimens may be refrigerated at 2-8°C for up to five days or stored at -20°C up to six months.
- · Avoid repetitive freezing and thawing of serum or plasma samples. This may result in variable loss of antibody activity.
- · Testing of heat-inactivated sera is not recommended.

#### STORAGE AND STABILITY

- Store test kit at 2-8°C in the dark.
- Do not expose reagents to heat, sun, or strong light during storage and usage.
- · Store microplate sealed and dessicated in the clip bag provided.
- · Shelf life of the unopended test kit is 18 months from day of production.
- Unopened reagents are stable until expiration of the kit. See labels for individual batch.
- Diluted Wash Buffer and Sample Buffer are stable for at least 30 days when stored at 2-8°C. We recommend consumption on the same day.

#### PROCEDURAL NOTES

- Do not use kit components beyond their expiration dates.
- Do not interchange kit components from different lots and products.
- All materials must be at room temperature (20-28°C) prior to use.
- Prepare all reagents and samples. Once started, performe the test without interruption.
- Double determinations may be done. By this means pipetting errors may become obvious.
- · Perform the assay steps only in the order indicated.
- · Always use fresh sample dilutions.
- · Pipette all reagents and samples into the bottom of the wells.
- To avoid carryover or contamination, change the pipette tip between samples and different kit controls.
- · Wash microwells thoroughly and remove the last droplets of wash buffer.
- · All incubation steps must be accurately timed.
- · Do not re-use microplate wells.

#### PREPARATION OF REAGENTS

#### WASH

Dilute the contents of one vial of the buffered wash solution concentrate (50x) with distilled or deionised water to a final volume of 1000 ml prior to use.

DILUENT

Sample Buffer P: Prior to use dilute the contents (20 ml) of one vial of sample buffer 5x concentrate with distilled or deionised water to a final volume of 100 ml.

#### Preparation of samples

Dilute patient samples 1:100 before the assay: Put 990 µl of prediluted sample buffer in a polystyrene tube and add 10 µl of sample. Mix well. Note: Calibrators / Controls are ready to use and need not be diluted.

#### TEST PROCEDURE

Prepare enough microplate modules for all calibrators / controls and patient samples.

- 1. Pipette 100 µl of calibrators, controls and prediluted patient samples into the wells Incubate for **30 minutes** at room temperature (20-28 °C). Discard the contents of the microwells and wash 3 times with 300 µl of wash solution.
- 2. Dispense 100 µl of enzyme conjugate into each well. Incubate for 15 minutes at room temperature. Discard the contents of the microwells and wash 3 times with 300 µl of wash solution.
- 3. Dispense **100 µ**I of TMB substrate solution into each well Incubate for 15 minutes at room temperature
- 4. Add 100 µl of stop solution to each well of the modules
  - Incubate for 5 minutes at room temperature.

Read the optical density at 450 nm (reference 600-690nm) and calculate the results. The developed colour is stable for at least 30 minutes. Read during this time.

Example for a pipetting scheme:

	1	2	3	4	5	6	7	8	9	10	11	12
A	А	P1										
в	В	P2										
c	С	P3										
D	D											
E	Е											
F	F											
G	C+											
н	C-											

P1, ... patient sample A-F calibrators C+, C- controls

#### VALIDATION

Test results are valid if the optical densities at 450 nm for calibrators / controls and the results for controls comply with the reference ranges indicated on the Certificate of Analysis enclosed in each test kit. If these quality control criteria are not met the assay run is invalid and should be repeated.

#### CALCULATION OF RESULTS

For quantitative results plot the optical density of each calibrator versus the calibrator concentration to create a calibration curve. The concentration of patient samples may then be estimated from the calibration curve by interpolation.

Using data reduction software a 4-Parameter-Fit with lin-log coordinates for optical density and concentration is the data reduction method of choice.

#### PERFORMANCE CHARACTERISTICS

#### Calibration

This assay system is calibrated in relative arbitrary units, since no international reference preparation is available for this assay.

#### Measuring range

The calculation range of this ELISA assay is 0 - 200 U/ml

#### Expected values

In a normal range study with samples from healthy blood donors the following ranges have been established with this ELISA assay: Cut-off 10 U/ml

#### Interpretation of results

Negative: < 10 U/ml Positive: ≥ 10 U/ml

#### Linearity

Patient samples containing high levels of specific antibody were serially diluted in sample buffer to demonstrate the dynamic range of the assay and the upper / lower end of linearity. Activity for each dilution was calculated from the calibration curve using a 4-Parameter-Fit with lin-log coordinates.

Sample	Dilution	Observed	Expected	O/E
		U/ml	U/ml	[%]
1	1:100	188.0	188.0	100
•	1:200	101.7	94.0	108
•	1:400	48.4	47.0	103
•	1:800	24.1	23.5	103
•	1:1600	<mark>10.9</mark>	11.8	93
2	1:100	<mark>196.0</mark>	196.0	100
•	1:200	97.2	98.0	99
•	1:400	49.4	49.0	101
•	1:800	24.8	24.5	101
	1:1600	11.9	12.3	97

#### Limit of detection

Functional sensitivity was determined to be: 1 U/ml

#### Reproducibility

Intra-assay precision: Coefficient of variation (CV) was calculated for each of three samples from the results of 24 determinations in a single run. Results for precision-within-assay are shown in the table below. Inter-assay precision: Coefficient of variation (CV) was calculated for each of three samples from the results of 8 determinations in 5 different runs. Results for run-to-run precision are shown in the table below.

	Intra-Assay		]	Inter-Assay				
Sample	Mean			Sample	Mean			
	U/ml	CV %			U/ml	CV %		
1	4.2	6.5	]	1	4.5	10.9		
2	20.7	6.7		2	22.4	9.2		
3	63.7	6.1		3	74.0	8.8		

#### Interfering substances

No interference has been observed with haemolytic (up to 1000 mg/dl) or lipemic (up to 3 g/dl triglycerides) sera or plasma, or bilirubin (up to 40 mg/dl) containing sera or plasma. Nor have any interfering effects been observed with the use of anticoagulants (Citrate, EDTA, Heparine). However for practical reasons it is recommended that grossly hemolyzed or lipemic samples should be avoided.

#### Study results

Study p	<u>n</u>	<u>n Pos</u>	%					
Coeliac	Coeliac disease							
Coeliac	: disea	se (IgA	deficient	t)	8	8	100.0	
Normal	huma	n sera			90	0	0.0	
		Clinical	Diagnosis	S				
		POS	NEG					
ORG 540G	POS	31	0					
	NEG	12	90					
		43	90	133				
Sensitivity:	72.1	%						
Specificity:	100.0	%						
Overall agreement:	91.0	%						

#### LIMITATIONS OF THE PROCEDURE

This assay is a diagnostic aid. A definite clinical diagnosis should not be based on the results of a single test, but should be made by the physician after all clinical and laboratory findings have been evaluated concerning the entire clinical picture of the patient. Also every decision for therapy should be taken individually.

The above pathological and normal reference ranges for antibodies in patient samples should be regarded as recommendations only. Each laboratory should establishe its own ranges according to ISO 15189 or other applicable laboratory guidelines.

#### REFERENCES

- Alessio M, Tonutti E, Brusca I, Radice A, Licini L, Sonzogni A, Florena A, Schiaffino E, Marus W, Sulfaro S, Villalta D: Correlation between IgA tissue transglutaminase antibody ratio and histological finding in celiac disease: A multicentre study. J Pediatr. Gastroenterol. Nutr. 2011.
- Cummins AG, Roberts-Thomson IC: Prevalence of celiac disease in the Asia-Pacific region. J Gastroenterol. Hepatol. 2009, 24:1347-1351.
- Giersiepen K, Lelgemann M, Stuhldreher N, Ronfani L, Husby S, Koletzko S, Korponay-Szabo IR: Accuracy of diagnostic antibody tests for coeliac disease in children: summary of an evidence report. J Pediatr. Gastroenterol. Nutr. 2012, 54:229-241.
- 4. Green PH, Cellier C: Celiac disease. N. Engl. J Med. 2007, 357:1731-1743.
- Gupta R, Reddy DN, Makharia GK, Sood A, Ramakrishna BS, Yachha SK, Thapa BR, Banerjee R, Anuradha S, Dutta U, Puri AS, Jain AK, Mulder CJ, Kumar A, Boindala S: Indian task force for celiac disease: current status. World J Gastroenterol. 2009, 15:6028-6033.
- Husby S, Koletzko S, Korponay-Szabo IR, Mearin ML, Phillips A, Shamir R, Troncone R, Giersiepen K, Branski D, Catassi C, Lelgeman M, Maki M, Ribes-Koninckx C, Ventura A, Zimmer KP: European Society for Pediatric Gastroenterology, Hepatology, and Nutrition guidelines for the diagnosis of coeliac disease. J Pediatr. Gastroenterol. Nutr. 2012, 54:136-160.
- 7. Mahadov S, Green PH: Celiac disease: a challenge for all physicians. Gastroenterol. Hepatol.(N Y) 2011, 7:554 -556.
- 8. Marietta EV, Rashtak S, Murray JA: Correlation analysis of celiac sprue tissue transglutaminase and deamidated gliadin IgG/IgA. World J Gastroenterol. 2009, 15:845-848.
- 9. Murray JA: Serodiagnosis of celiac disease. Clin. Lab. Med. 1997, 17:445-464.
- 10. Rubio-Tapia A, Murray JA: Celiac disease. Curr. Opin. Gastroenterol. 2010, 26:116-122.
- 11. Telega G, Bennet TR, Werlin S: Emerging new clinical patterns in the presentation of celiac disease. Arch. Pediatr. Adolesc. Med 2008, 162:164-168.
- 12. Tjon JM, van BJ, Koning F: Celiac disease: how complicated can it get? Immunogenetics 2010, 62:641-651.
- van der Windt DA, Jellema P, Mulder CJ, Kneepkens CM, van der Horst HE: Diagnostic testing for celiac disease among patients with abdominal symptoms: a systematic review. JAMA 2010, 303:1738-1746.
- 14. Villalta D, Tonutti E, Prause C, Koletzko S, Uhlig HH, Vermeersch P, Bossuyt X, Stern M, Laass MW, Ellis JH, Ciclitira PJ, Richter T, Daehnrich C, Schlumberger W, Mothes T: IgG antibodies against deamidated gliadin peptides for diagnosis of celiac disease in patients with IgA deficiency. Clin. Chem. 2010, 56:464-468.

#### Notice to the user (European Union):

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the EU Member State in which the user and/or the patient is established.

#### Change Control

Former version: ORG 540G\_IFU\_EN\_QM113175\_2016-05-13\_2 Reason for revision: Introduction electronic IFU on homepage



СИСТЕМА ДОБРОВОЛЬНОЙ СЕРТИФИКАЦИМ 0117161 «ПРОМТЕХСТАНДАРТ»

№ РОСС RU.32001.04ИБФ1 в едином реестре зарегистрированных систем добровольной сертнфикации

# СЕРТИФИКАТ СООТВЕТСТВИЯ

ПО



Регистрационный номер РОСС RU.04ИБФ1.ОС23.0000308

Срок действия с 30

30.06.2022

29.06.2025

ОРГАН ПО СЕРТИФИКАЦИИ № РОСС RU, 32001.04ИБФ1.0С23 Общество с ограниченной ответственностью «ЕВРАЗИЙСКИЙ СОЮЗ СЕРТИФИКАЦИИ» 192289, город Санкт-Петербург, улица Олеко Дундича, дом № 35, корпус 1, литера А, 2-Н, офис 4, тел. +7 (812) 649-93-88, email: info@essert.ru

# выдан

Закрытому акционерному обществу «ЭКОлаб» ИНН 5035025076 ОГРН 1035007106958 Адрес: 142530, РФ, Московская область, г. Электрогорск, ул. Буденного, д. 1

# НАСТОЯЩИЙ СЕРТИФИКАТ УДОСТОВЕРЯЕТ, ЧТО СИСТЕМА МЕНЕДЖМЕНТА КАЧЕСТВА МЕДИЦИНСКИХ ИЗДЕЛИЙ

применительно к работам согласно приложению № 1 к настоящему сертификату

## СООТВЕТСТВУЕТ ТРЕБОВАНИЯМ ГОСТ ISO 13485-2017 (ISO 13485:2016)

Выдан на основании решения экспертной комиссии, протокол № РОСС RU.04ИБФ1.0С23.0000308П от 30.06.2022

пля CEPTNOUKATOR уководитель органа

редседатель комиссии

А.В. Арендарь

А.А. Акимов HHIDRARL davouro

тоящий страциных с цогостствия обязывает организацию поддерживать состояние выполняемых работ в соответствие с вышеуказанным стандартом, что будет находиться под контролем органа по сертификации системы добровольной сертификации «ПромТехСтандарт» и подтверждаться при прохождении ежего люто пистемы

АО «ОПЦИОН», Москва, 2021 г., «В». ТЗ № 113

# СИСТЕМА ДОБРОВОЛЬНОЙ СЕРТИФИКАЦИМ<sup>0117162</sup> «ПРОМТЕХСТАНДАРТ»

№ РОСС RU.32001.04ИБФ1 в едином ресстре зарегистрированных систем добровольной сертификации



Настояний сертификат соответствия

# **РАЗРЕШЕНИЕ**

НА ПРИМЕНЕНИЕ ЗНАКА СООТВЕТСТВИЯ СИСТЕМЫ ДОБРОВОЛЬНОЙ СЕРТИФИКАЦИИ «ПРОМТЕХСТАНДАРТ-

по

Регистрационный номер РОСС RU.04ИБФ1.ОС23.0000308Р

Срок действия с

30.06.2022

29.06.2025

### ОРГАН ПО СЕРТИФИКАЦИИ

№ РОСС RU.32001.04ИБФ1 ОС23 Общество с ограниченной ответственностью «ЕВРАЗИЙСКИЙ СОЮЗ СЕРТИФИКАЦИИ» 192289. город Санкт-Петербург, улица Олско Дундича, дом № 35. корпус 1, литера А, 2-Н, офис 4, тел. :+7 (812) 649-93-88, email: info@essert.ru

# выдано

Закрытому акционерному обществу «ЭКОлаб» ИНН 5035025076 ОГРН 1035007106958 Адрес: 142530, РФ, Московская область, г. Электрогорск, ул. Буденного, д. 1

на основании сертификата № РОСС RU.04ИБФ1.ОС23.0000308

# Настоящее разрешение предоставляет право применения знака соответствия системы добровольной сертификации «ПРОМТЕХСТАНДАРТ»:

при маркировке продукции, при оказании работ (услуг), на бланках организации, в рекламно-информационных материалах, печатных изданиях, вывесках, выставочных стендах и т.д., на сайтах организации в сети Интернет, в соответствии с правилами применения знака соответствия системы добровольной сертификации "ПромТехСтандарт"

Руководитель органа

А.В. Арендарь

Председатель комиссии

А.А. Акимов

HOLO KOHTBO.39

обязывает организацию поддерживать состояние выполняемых работ в соответствие с вышеуказанным стандартом, что будет находиться под контролем органа по сертификации системы добровольной сертификации «ПромТехСтандарт» и подтверждаться при прохождении ежегодного инспекцию

# СИСТЕМА ДОБРОВОЛЬНОЙ СЕРТИФИКАЦИМ 0117163 «ПРОМТЕХСТАНДАРТ»

№ РОСС RU.32001.04ИБФ1 в едином реестре зарегистрированных систем добровольной сертификации



# СЕРТИФИКАТ СООТВЕТСТВИЯ ЭКСПЕРТА

Регистрационный номер РОСС RU.04ИБФ1.ОС23.000720Э

Срок действия с 30.06.2022

по 29.06.2025

ОРГАН ПО СЕРТИФИКАЦИИ

№ РОСС RU. 32001.04ИБФ1.ОС23 Общество с ограниченной ответственностью «ЕВРАЗИЙСКИЙ СОЮЗ СЕРТИФИКАЦИИ» 192289, город Санкт-Петербург. улица Олеко Дундича, дом № 35, корпус 1, литера А. 2-Н, офис 4, тел.: +7 (812) 649-93-88, email: info@essen.ru

# НАСТОЯЩИЙ СЕРТИФИКАТ УДОСТОВЕРЯЕТ, ЧТО

### Овинникова Светлана Сергеевна

соответствует требованиям

предъявляемым системой добровольной сертификации "ПромТехСтандарт"

к экспертам-аудиторам внутренних проверок

системы менеджмента качества медицинских изделий

на соответствие требованиям стандарта

ГОСТ ISO 13485-2017 (ISO 13485:2016)

Выдан на основании решения экспертной комиссии, протокол № РОСС RU.04ИБФ1.ОС23.000720ПЭ от 30.06.2022

и зарегистрирован в реестре экспертов системы добровольной сертификации "Пром Тех Стандарт".



Руководитель органа

А.В. Арендарь

А.А. Акимов шналы, фамиция

Председатель комиссии

Настоящий сертификат соответствия обязывает организацию поддерживать состояние вышодниемых работ в соответствие с вышеуказанным стандартом, что будет находиться под контролем органа по сертификации системы добровольной сертификации «ПромТехСтандарт» и полтверждаться при прохождении ежегодного инспекци





# СЕРТИФИКАТ СООТВЕТСТВИЯ ЭКСПЕРТА

Регистрационный номер РОСС RU.04ИБФ1.ОС23.000721Э

Срок действия с

30.06.2022 по

29.06.2025

ОРГАН ПО СЕРТИФИКАЦИИ № РОСС RU.32001.04ИБФ1.ОС23 Общество с ограниченной ответственностью «ЕВРАЗИЙСКИЙ СОЮЗ СЕРТИФИКАЦИИ» 192289, город Санкт-Петербург, улица Олеко Дундича, дом № 35, корпус 1, литера А. 2-Н. офис 4, тел.: +7 (812) 649-93-88, email: info@essent.ru

# НАСТОЯЩИЙ СЕРТИФИКАТ УДОСТОВЕРЯЕТ, ЧТО

### Котляр Марина Анатольевна

соответствует требованиям

предъявляемым системой добровольной сертификации "ПромТехСтандарт"

к экспертам-аудиторам внутренних проверок

системы менеджмента качества медицинских изделий

на соответствие требованиям стандарта

FOCT ISO 13485-2017 (ISO 13485:2016)

Выдан на основании решения экспертной комиссии, протокол № РОСС RU.04ИБФ1.ОС23.000721ПЭ от 30.06.2022

и зарегистрирован в реестре экспертов системы добровольной сертификации "ПромТехСтандарт".



Председатель комиссии

А.В. Арендарь нишалы, фамини

А.А. Акимов unundanti damunu

Настоящий сертификат соответствия обязывает организац

обязывает организацию поддерживать состояние выполниемых работ в соответствие с выписуказанным стандартом, что будет находиться ст контролем органа по сертификации системы добровольной сертификации «Пром ГехСтмидарт» и подтверждаться при прохожлении ежегодного инспекции.

СИСТЕМА ДОБРОВОЛЬНОЙ СЕРТИФИКАЦИМ<sup>0117165</sup> «ПРОМТЕХСТАНДАРТ»

№ РОСС RU.32001.04ИБФ1 в едином ресстре зарегистрированных систем добровольной сертификации



# СЕРТИФИКАТ СООТВЕТСТВИЯ ЭКСПЕРТА

Регистрационный номер РОСС RU.04ИБФ1.0С23.000722Э

Срок действия с

30.06.2022 по

29.06.2025

ОРГАН ПО СЕРТИФИКАЦИИ № РОСС RU.32001.04ИБФ1.ОС23

Общество с ограниченной ответственностью «ЕВРАЗИЙСКИЙ СОЮЗ СЕРТИФИКАЦИИ» 192289, город Санкт-Петербург, улица Олеко Дундича, дом № 35, корпус 1, литера А. 2-Н, офис 4, тел.: +7 (812) 649-93-88, email: info@essent.ru

# НАСТОЯЩИЙ СЕРТИФИКАТ УДОСТОВЕРЯЕТ, ЧТО

Трошенкова Елена Петровна

соответствует требованиям

предъявляемым системой добровольной сертификации "ПромТехСтандарт"

к экспертам-аудиторам внутренних проверок

системы менеджмента качества медицинских изделий

на соответствие требованиям стандарта

FOCT ISO 13485-2017 (ISO 13485.2016)

Выдан на основании решения экспертной комиссии, протокол № РОСС RU.04ИБФ1.0С23.000722ПЭ от 30.06.2022

и зарегистрирован в реестре экспертов системы добровольной сертификации "ПромТехСтандарт".



Руководитель органа

Председатель комиссии

А.В. Арендарь

А.А. Акимов лы, фамилия

Настонций сертификат соответствия обязывает организацию поддерживать состояние вынолниемых работ в соответствие с вышеуказанным стандартом, что будет изходиться под контролем органа по сертификации системы добровольной сертификации «ПромТехСтандарт» и полтверждаться при прохождении екстодного инсискци





# СЕРТИФИКАТ СООТВЕТСТВИЯ ЭКСПЕРТА

Регистрационный номер РОСС RU.04ИБФ1.0С23.000723Э

Срок действия с

30.06.2022 ПО

29.06.2025

ОРГАН ПО СЕРТИФИКАЦИИ № POCC RU.32001.04ИБФ1.ОС23 Общество с ограниченной ответственностью «ЕВРАЗИЙСКИЙ СОЮЗ СЕРТИФИКАЦИИ» 192289, город Санкт-Петербург, улица Олеко Дундича, дом № 35, корпус 1, литера А, 2-Н, офис 4, тел.: +7 (812) 649-93-88, email: info@essert.ru

# НАСТОЯЩИЙ СЕРТИФИКАТ УДОСТОВЕРЯЕТ, ЧТО

### Нищакова Наталья Евгеньевна

соответствует требованиям

предъявляемым системой добровольной сертификации "ПромТехСтандарт"

к экспертам-аудиторам внутренних проверок

системы менеджмента качества медицинских изделий

на соответствие требованиям стандарта

FOCT ISO 13485-2017 (ISO 13485:2016)

Выдан на основании решения экспертной комиссии, протокол № РОСС RU.04ИБФ1.0С23.000723ПЭ от 30.06.2022

и зарегистрирован в реестре экспертов системы добровольной сертификации "Пром Тех Стандарт".



Руководитель органа

Председатель комисси

А.В. Арендарь инициалы, фамюция

А.А. Акимов

OFO KOHTDO.TR

Настоящий сертификат соответствия

обязывает организацию поддерживать состояние вынолниемых работ в соответствие с вышеуказанным станда ем органа по сертификации системы добровольной сертификации «ПромТехСтандарт» и полтверждаться при прохо нуказанным стандартом, что будет находиться

СИСТЕМА ДОБРОВОЛЬНОЙ СЕРТИФИКАЦИМ 0117167 «ПРОМТЕХСТАНДАРТ»

№ РОСС RU.32001.04ИБФ1 в едином реестре зарегистрированных систем добровольной сертификации



# СЕРТИФИКАТ СООТВЕТСТВИЯ ЭКСПЕРТА

Регистрационный номер РОСС RU.04ИБФ1.ОС23.000724Э

Срок действия с

30.06.2022 по

29.06.2025

ОРГАН ПО СЕРТИФИКАЦИИ № РОСС RU.32001.04ИБФ1.ОС23 Общество с ограниченной ответственностью «ЕВРАЗИЙСКИЙ СОЮЗ СЕРТИФИКАЦИИ» 192289, город Санкт-Петербург. улица Олеко Дундича, дом № 35, корпус 1, литера А. 2-Н, офис 4, тел.: +7 (812) 649-93-88, email: info@essert.ru

# НАСТОЯЩИЙ СЕРТИФИКАТ УДОСТОВЕРЯЕТ, ЧТО

Королева Татьяна Александровна

соответствует требованиям

предъявляемым системой добровольной сертификации "ПромТехСтандарт"

к экспертам-аудиторам внутренних проверок

системы менеджмента качества медицинских изделий

на соответствие требованиям стандарта

FOCT ISO 13485-2017 (ISO 13485:2016)

Выдан на основании решения экспертной комиссии, протокол № РОСС RU 04ИБФ1 ОС23.000724ПЭ от 30.06.2022

и зарегистрирован в реестре экспертов системы добровольной сертификации "Пром Тех Стандарт".



Руководитель органа

А.В. Арендарь

А.А. Акимов ниналы, фамития

KOHT BO.TR

Председатель комиссии

онный сонзификат соответствия облакнает арганизацию подлерживать состояние выполниемых работ в соответствие с вышеуказациым ставдартом, что будет находиться под контролем органа по сертификации системы добровольной сертификации «ПромТехСтандарт» в полтверждаться при прохождении ежегодного виспекцию

# СИСТЕМА ДОБРОВОЛЬНОЙ СЕРТИФИКАЦИМ 0117168 «ПРОМТЕХСТАНДАРТ»

№ РОСС RU.32001.04ИБФ1 в едином ресстре зарегистрированных систем добровольной сертификации

# ПРИЛОЖЕНИЕ № 1



К сертификату соответствия № РОСС RU.04ИБФ1.ОС23.0000308 (является неотъемлемой частью сергификата соответствия)

Срок действия с

30.06.2022 по

29.06.2025

### ОРГАН ПО СЕРТИФИКАЦИИ

№ POCC RU.32001.04ИБФ1.0C23

Общество с ограниченной ответственностью «ЕВРАЗИЙСКИЙ СОЮЗ СЕРТИФИКАЦИИ» 192289, город Санкт-Петербург, улица Олеко Дундича, дом № 35, корпус 1, литера А, 2-Н, офис 4, тел. : +7 (812) 649-93-88, email: info@essert.ru

### Применительно к видам работ:

### ОКВЭД:

46.46.2 - Торговля оптовая изделиями, применяемыми в медицинских целях

### ОКВЭД 2:

21.20.2 Производство материалов, применяемых в медицинских целях 72.19 Научные исследования и разработки в области естественных и технических наук прочие 86.90.9 Деятельность в области медицины прочая, не включенная в другие группировки



Руководитель органа

А.В. Арендарь ниналы, фамил

А.А. Акимов иналы, фамилия

Настоящий сертификат соответствия

обязывает организацию поддерживать состояние выполняемых работ в соответствие с вышеуказанным стан под контролем организации системы добровольной сертификации «ПромТехСтандарт» и подтвержаться при прох зать состояние выполняемых работ в соответствие с вышеуказанным стандартом, что будет ваходиться

АО «ОПЦИОН», Москва, 2021 г., «В». ТЗ № 1130.