



EDAN

EDAN INSTRUMENTS, INC.

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 98/79/EC

MANUFACTURER: Edan Instruments, Inc.
#15 Jinhui Road, Jinsha Community, Kengzi Sub-District,
Pingshan District, 518122 Shenzhen, P.R.China

EUROPEAN REPRESENTATIVE: Shanghai International Holding Corp. GmbH
Eiffestrasse 80, 20537 Hamburg Germany

PRODUCT/ MODEL: Hematology Analyzer/ H60, H60S, H66, H66S, H68, H68S,
H69, H69S
Reagents for Hematology Analyzer/ HD600 Diluent, HL600
Lyse, HC600 Cleaner, ED-60D Control H, ED-
60D Control N, ED-60D Control L Hematology
Controls, ED-CAL PLUS Hematology Calibrator.

The accessories are used together with the product.

EDMA[Name/Code]: CC Hardware + accessories + consumables + software/23.01.10.01.00
CBC-Reagents(Cleaning-/Diluting-/Lysing-/Sheat-fluids)/13.01.01.01.00
Blood Multilevel Controls/ 13.01.50.03.00
Whole Blood Calibrators/ 13.01.50.07.00

CLASSIFICATION:General/other device, devices other than those covered by Annex II and devices for performance evaluation, non-self-testing, according to article 9 of IVDD.

CONFORMITY ASSESSMENT ROUTE: Annex III

WE, EDAN INSTRUMENTS, INC., HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT(S) MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 98/79/EC OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL OF 27 OCTOBER 1998 ON IN VITRO DIAGNOSTIC MEDICAL DEVICES.

ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURER.

STANDARDS APPLIED: EN ISO 14971: 2012, IEC 61010-1:2017, IEC 61010-2-101:2018, EN 61326-1:2013, EN 61326-2-6:2013, EN 62304:2006 +A1:2015, EN 62366-1:2015, EN 1041: 2008+A1:2013, EN ISO 15223-1:2016, EN ISO 18113-1:2011, EN ISO 18113-2:2011, EN ISO 18113-3:2011, EN 13612:2002, EN 13641:2002, EN ISO 17511:2003, EN ISO 23640:2015

CE MARK



START OF CE-MARKING:

2021.8.10

PLACE, DATE OF ISSUE:

SHENZHEN, 2021.8.10

SIGNATURE:

NAME LIU YONGYING

MANAGEMENT REPRESENTATIVE