



EMSAS®

MEDICAL COOLING & HEATING SYSTEMS

www.emsas-as.com.tr

QUALITY & SAFETY PROTECTION

BLOOD BANK REFRIGERATOR

**LABORATORY - PHARMACY
REFRIGERATOR**

INCUBATOR - AGITATOR

PLASMA THAWER

DEEP FREEZER

ULTRA LOW FREEZERS

**BREAST MILK STORAGE
REFRIGERATOR & FREEZER**

BABY BOTTLE WARMERS

**ETUVE (LABORATORY OVEN) &
WATER BATH**

COOLED INCUBATOR



www.emsas-as.com.tr





EKT Serie (0°C / +15°C)

- The outer surface of the device is made of galvanized steel, painted with electrostatic paint to protect against rust. The inner surface is produced of stainless steel.
- The device operates with 195-230 V/50 Hz network voltage.
- The door of the device has double thermal glazing, it can be locked and has magnetic seal rings. Thus, it provides an opportunity to trace the stocks.
- There is a plastic covered wire shelves in the device that can be adjusted according to the user request.
- Operates between +0/+15°C and it is set to +4°C.
- Provides equal temperature distribution with the strengthened fan system it contains..
- In order to prevent the temperature rising up speedily , Interior fans shut down when door is opened
- Works smoothly and quietly.
- In order to protect the productivity of the device cooling evaporator, there is full-automatic defrosting system.
- The lightening of the inner cabin of the device is carried out with LED lightening.
- A user-friendly digital control panel with a micro processor has been used on the device. This system can keep the data in memory for 30 days.
- There is a USB socket to transfer the cabinet temperature information to the computer when requested. In this way, the 10 year old temperature records can be transferred to the PC in excel format.
- The thermostat at the control panel of the device can make measurements with 0.1 degree sensitivity.
- There is an accumulator system charged automatically at the control panel of the device. This system allows the control panel and thermal printer (if available) to operate for 24 hours when the electricity cuts off.
- While the refrigerator is working the device gives alerts with the visual and audio signal when the lower and upper temperature limits are exceeded, the door of the device is left open, and there is electricity cut off or low voltage problems.
- The cooling system and the insulation system of the device do not contain CFC gas that is harmful for OZONE.
- High quality well known brand compressor is used which is hermetically sealed and air-cooled
- There are two braked and two regular castors under the device that provide easy portability.
- Our devices can be connected to heat tracking system.
- All medical devices produced by EMSAS A.S. are guaranteed for 2 years.
- As per the request of the user, a thermal printer can be installed at the device. The numerical and graphical printout of the data recorded with the thermal printer can be taken.
- All of our products are produced in line with the ISO 9001:2008, TSE service place competence certificates. The device has CE certificates and barcode.

EKT 425



**OPTIONAL BLOCK
3XGLASS DOOR**



**OPTIONAL CHROME
SHELVES**




**OPTIONAL GRAPHICAL
CHART RECORDER**



**OPTIONAL THERMAL
PRINTER**



**OPTIONAL TOUCH
SCREEN DISPLAY**



**OPTIONAL VACCINE TRACKING
SYSTEM**

EKT-B Serie (0°C / +15°C)

- EKT-B Serie has same features with EKT Serie only the difference; door system is designed as polyurethane pressed block (solid) door.



MEDICAL COOLING & HEATING SYSTEMS



www.emsas-as.com.tr
www.kandolabi.com
www.asidolabi.com
ISO 9001:2008
ISO 13485



Muradiye Sanayi Bölgesi Muradiye Mahallesi
28 Sokak No:6 Yunusemre - MANİSA / TURKEY
Tel: +90 (236) 214 03 96 - 97 - 98 Fax: +90 (236) 214 07 06
izmir Tel: +90 (232) 479 55 22 - 479 55 23
mail: export1@emsas-as.com.tr
www.emsas-as.com.tr



EKT SERIES
PHARMACY
REFRIGERATOR
MANUAL BOOK

GENERAL INFORMATION

Please read the manual book carefully and keep the book for forward before using your device.

If there can be any defect contact with the shipping company to the report damage. According to ICC regulations, the responsibility belongs to the consumer.

When operating the device;

1. Follow the warning labels
2. Do not remove the labels
3. Do not operate the damaged device
4. Do not operate the device with a damaged cable system.
5. Do not move the device while the device operating.

- * In case of any problem please contact with EMSAŞ A.Ş technical service.
- * Read the manual book carefully before using the device, this ensures you to get high efficiency.
- * The EMSAŞ A.Ş has the right of changing and developing the device structure.
- * Please keep the product package until be sure the device is full, complete, and correct.
- * The warranty is valid until the precautions described in this manual are followed.

***Informations in this document belongs to EMSAŞ A.Ş
It can not be reproduced or distributed without any permission.***

WARRANTY CONDITIONS

1. The warranty period starts from the delivery date of the goods and is two years.
2. The entire product, including all parts, is under our company's warranty.
3. In case of malfunction of the product within the warranty period, the period of repair is added to the warranty period. The repair period of the goods is maximum of thirty working days. This period starts from the date of notification of the goods to the service station. In case there is no service station, respectively; it starts from the date of notification to one of the seller, dealer, agency, representative, importer or manufacturer of the goods. If the defect of the industrial goods cannot be fixed within 15 working days, the manufacturer or the importer; it has to allocate another industrial good with similar characteristics to the use of the consumer until the repair of the good is completed.
4. If the product fails during the warranty period due to material and workmanship or assembly errors, it will be repaired without any charge under any other name, such as labor cost, replacement part price, or any other name.
5. Despite the consumer's use of the right to repair, the goods;
 - The fact that the same fault is repeated more than twice within a year from the date of delivery to the consumer, or that different faults occur more than four times, or that the sum of different faults is more than six within the specified warranty period, and that these faults perpetuate the inability to benefit from the goods,
 - Exceeding the maximum time required for repair,
 - In cases where it is determined that the repair of the fault is not possible with the report to be issued by the company's service station, if the service station is not available, respectively, one of the dealers, agents, representatives, importers, or manufacturers, the consumer may request a free replacement of the goods, a refund or a price reduction at the rate of the defect.
6. Defects arising from the use of the product contrary to the terms in the user manual are not covered by the warranty.
7. For problems that may arise regarding the warranty certificate, the Ministry of Industry and Trade, General Directorate of Protection of Consumer and Competition can be applied.

CUSTOMER MATTERS REGARDING THE WARRANTY

Issues that our customers should pay attention to regarding the warranty

1. Malfunctions arising from the use of the device other than its normal use,
2. Defects arising from exposure to substances with physical or chemical effects,
3. Damages and malfunctions that occur in case of non-compliance with the points in the user manual,

4. Damages due to lower excess voltage faulty installation or using a voltage different from the voltage written on the label of the device,
5. Damages that will occur during loading, unloading, and transportation under the responsibility of the customer,
6. The device has been disassembled for various reasons or its functions have been changed by people other than service officials,
7. Malfunctions and damages that may occur due to fire and lightning,
8. Responsibility for the completion of the warranty certificate and the delivery to the consumer belongs to the seller, dealer, agency, or representative office from which the consumer purchased the goods,
9. This warranty is void if the warranty document has been tampered with, the original serial number on the product has been removed or tampered with.

THIS MANUAL MUST BE READ CAREFULLY BEFORE OPERATING THE DEVICE.

WARRANTY CONDITIONS APPLY IF THE INSTRUCTIONS AND WARNINGS DESCRIBED IN THE DEVICE'S USER MANUAL ARE COMPLIED.

THE INFORMATION CONTAINED IN THIS MANUAL BELONGS TO EMSAS. THE INFORMATION IN THE MANUAL CANNOT BE REDUCED OR DISTRIBUTED WITHOUT EMSAS' PERMISSION.

In the above-mentioned cases, troubleshooting is done for a fee.

The product's installation and transportation to the place of use are not included in the product price.

* The useful life of the goods announced by the Ministry of Industry and Trade in accordance with the Law on the Protection of the Consumer No. 4077 is 10 years.

TABLE OF CONTENTS

- 1. DESCRIPTION OF DEVICE, FUNCTION, AND OPERATING PRINCIPLE**
- 2. TECHNICAL SPECIFICATIONS**
 - 2.1. Technical Specifications Table
 - 2.2. Technical Specifications Description
- 3. INSTALLING THE DEVICE**
 - 3.1. Environmental Conditions
 - 3.2. Handling and Shipping
 - 3.3. Unpacking
 - 3.4. Determination and Placement of Settlement
- 4. INSTRUCTIONS FOR USE**
 - 4.1. Device Operation
 - 4.2. Introduction Image
 - 4.3. Control Display Panel and its Functions
 - 4.4. Electronic Wiring Diagram
- 5. CLEANING AND MAINTENANCE**
 - 5.1. Maintenance
 - 5.2. Cleaning
- 6. SAFETY**
- 7. DISPOSAL(battery, compressor, mainboard and etc.)**
- 8. WARNINGS & PRECAUTIONS**
- 9. MEANING OF SYMBOLS**
 - 9.1. Meaning of Symbols on Top of the Device
- 10. SERVICE**
 - 10.1. Troubleshooting Table
- 11. SPARE PARTS & ACCESSORIES**
 - 11.1. SPARE PARTS
 - 11.2. ACCESSORIES

1. INTRODUCTION

1.1. DESCRIPTION OF DEVICE, FUNCTION, AND OPERATING PRINCIPLE

Products such as vaccines, kits, drugs and serums are designed to be used in hospitals, laboratories, pharmacies, health inspection centers and departments where products such as drugs, kits and serum are stored.

Cabinets with a cooling system that allows them to be stored at 0°C / +15°C for up to 12 months are called "Kit Storage Cabinets".

PURPOSE:

In all health centers, products such as vaccines, medical kits, drugs and serums are at temperatures between 0°C / +15°C.

It is to be kept at +4 °C for up to 12 months. Our kit storage cabinets are designed to serve for many years in health centers.

WORKING PRINCIPLES:

The device is designed to control temperatures in the range of 0°C / +15°C.

All of our cabinets in the EKT series can be adjusted to the desired temperature between 0°C / +15°C.

However, the factory set value is +4°C.

The cooling system in our EKT series cabinets works with the on/off (open / close) principle.

In all our EKT series cabinets, air circulation is provided with the help of a fan in order to ensure equal heat distribution in the inner cabinet.

Microprocessor-controlled electronic cards are used in our devices.

CFC-free polyurethane foam insulation material is used so that the device can maintain the temperature in the interior for a longer period of time.

Our EKT series cabinets are divided into 5 main groups:

EKT Series: Double Glass Door Model.

EKT-A Series: Double Glass Door & Simple Card Model

EKT-B Series: Block Door Model

EKT-AB Series: Block Door & Simple Card Model

EKT VK Series: Thermal Printer Model

Matters to be considered by the User During Use:

- When putting blood bags in your cooler, do not stack products in the cabinet in a way that prevents air circulation. This will cause unwanted temperature differences in the upper and lower parts of your cooler.
- There is 1 sensor in the blood storage cabinets that measures the temperature inside the cabin. On the electronic card, this temperature is reflected on the screen and shown to the user.
- After starting your cooler, it can give an alarm visually and audibly until it reaches the ideal operating temperature on the digital display. When the cooler reaches the ideal temperature (between 0°C and +15°C), the visual and audible alarm will stop. After that, you can load your cabinet.

External Probe Input;

The plug on the right side of the cabinet is removed and the probe is placed (If the thickness of the probe is high, the plastic can be expanded.) Take care that the inserted probe tip is inserted into the cabinet and fixed next to the sensor inside the cabinet.

2.2 Technical Specifications Description

- The outer surface of the device is made of corrosion-resistant electrostatic painted galvanized sheet material, and the inner surface is made of chrome-nickel alloy sheet material.
- The device is 195-230 V/50 Hz. It works with mains voltage.
- The door of the device is double glazed, lockable and magnetically sealed. In this way, it allows stock tracking.
- Inside of the device there is a user arrangeable plastic coated shelf. The device works in the range of 0°C / +15°C, which is the ideal storage temperature of products such as kits, vaccines, drugs, and serum, and can be adjusted to +4°C.
- Equal heat distribution is provided in the cabinet with the reinforced fan system in the cabinet.
- The device works silently and smoothly without separating kits components.
- The device has full automatic defrost system for protecting evaporators efficient work.
- Cabinet interior lightning realized with LED lighting module.
- Eco friendly microprocessor controlled panel is used in our device.
- The device has a USB port that enables the transfer of temperature information to the computer environment at any time.
- The thermostat in the cabinet control panel can measure with an accuracy of 0.1°C.
- The device has a sensor for measuring the cabin temperature inside.
- In control panel there is a accumulator system that can be charged automatically. In case of a electricity cut this system ensures that the digital control panel and thermal printer continue to work for 24 hour long.
- The device gives an audio and visual warning signal when the device's doors open, in power cut, when the upper and lower limits are exceeded, and low voltage situations while the device working.
- The insulation material does not include CFC that harmful for environment and ozon layer.
- 2 pieces of braked and 2 pieces of regular castors are used at the bottom of our EKT series devices, except EKT 80, EKT 80 include 4 pieces of ball joint.
- All medical devices produce by Emsaş A.Ş are guaranteed by 2 years.
- As per the request of the user, a thermal printer can be installed at the device. The numerical and graphical printout of the data recorded with the thermal printer can be taken.
- All of our products are produced in accordance with ISO 9001:2008, TSE service place qualification requirements. Our blood cabinets that we have produced have CE, UBB registration and barcode.

OPTIONAL FEATURES:

USB MEMORY: Electronic card system can be sent as USB memory upon user request. Up to 10 years, degrees can be stored and transferred to the computer environment. The transfer is provided in excel format.

VK (THERMAL PRINTER SYSTEM): The thermal printer system can be optionally attached to our devices. It can give output at intervals between 5 minutes and 60 minutes. It has the feature of printing the temperature of the device inside the cabinet on paper.

MAINBOARD WITHOUT USB MEMORY: In line with the user's request and requirement, a card that does not keep the temperature in the cabinet in its memory can also be inserted.

3.INSTALLING THE DEVICE

3.1. Environmental Conditions

Your device works efficiently under the following environmental conditions:

- Using only indoor area
- Storage and Transportation conditions: +25°C / +45°C
- Relative Humidity: %65
- Temperature Values: 0°C / +15°C

3.2. Handling and Shipping

- Because of, the device weight all the shipping should apply with the proper equipment and experienced staff.
- During transportation it must be supported from underneath and must not be turned upside down.
- Interior hardware (shelf rails, supports, wire shelves, drawers) is shipped inside the unit. The device is fixed to a wooden base by means of screws or packaged with cardboard, Styrofoam, and burst nylon.

3.3. Unpacking

- Open the device cardboard package.
- Remove the stretch and styrofoam, which is the second packaging on the device.
- Remove the wooden base
- Check is there any transportation damage on the device with eye.
- If is there any transportation damage firstly take a picture of damage and secondly Inform the transport company officer to keep a report.

3.4. Determination and Placement of Settlement

- Check that the placement spot is suitable
- Avoid direct sunlight and heat sources
- Remove the accessory from the unit
- Care should be taken to carry the device vertically during the relocation phase and during the transport phase. Otherwise, the device will not work or will not work properly and cannot be used because the balance of the device will be disturbed.
- Check that the ambient temperature where the device is installed is between 15°C and 20°C. Otherwise, the cooling performance of your device will decrease.
- Installation should be on flat ground and with four legs. Otherwise, it may not work properly.
- The placement of the device shall not be wet or damp. Otherwise, it can cause damage to devices electrical systems.
- The device shall not be close to any other dissipate heat device. Otherwise, it can cause damage to devices electrical systems.
- Check that the user has the opportunity to monitor the device if he/she does other work.
- When placing your device, you should pay attention to the fact that there must be a minimum of 10 cm space between the side, back and top surfaces and the place to be placed, and 20 cm from the top. There should be no plates behind the device that may cause evaporation.
- The hose in the compressor compartment of your cooler; for evacuation of dripping water during defrosting. Before operating the cabinet, be sure to check whether there is water in the hose in the container.
- When your device needs to be moved or relocated, have it done through the service channel if necessary.

4. INSTRUCTIONS FOR USE

4.1 Device Operation


- Plug the device in (grounded plug)
- Press the on/off switch, enter the password and turn on the device.
- The operating range of the device is between 0°C ve +15°C.
- Before operating your device, have technical personnel (electrician) check whether the socket to be used is grounded or not and the mains voltage. Do not operate if it is not grounded.
- Wait before operating your device at least 45 minutes after unpacking your device. This will prevent the

- Service representative at first start-up; Information about the device; technical personnel and health personnel (nurse - Dr) of the relevant customer should transfer.

4.2 Introduction Image



PIC1. EKT Pharmacy Refrigerator





PIC2. Control Panel



PIC3. Control Panel Without USB Port

DISPLAY BUTTONS	1		ON/OFF BUTTONS
	2		LIGHT ON/OFF BUTTONS
	3		UP ADJUSTMENT BUTTON
	4		MENU AND SETTINGS BUTTON
	5		DOWN ADJUSTMENT BUTTON
DIGITAL DISPLAY BUTTONS	6		RED LIGHT (LCD DISPLAY)
	7		GREEN LIGHT(WHEN COMPRESSOR IS ON)
	8		YELLOW LIGHT (POWER LIGHT)
	9		ALARM SILENCE BUTTON
	10		PRINTER BUTTON

1		ENTER BUTTON
2		UPPER ADJUSTMENT BUTTON
3		LOWER ADJUSTMENT BUTTON


Factory Output Settings

EKT SERIES

- | | |
|----------------------------------|-------------|
| 1) Operating Temperature Range | : 0 / +15°C |
| 2) Set Temperature Value | : +4°C |
| 4) Compressor Delay Time | : 30 sec. |
| 5) Door Open Alarm Waiting Time | : 2 min |
| 6) Alarm Silence Time | : 30 sec. |
| 7) Thermal Printer Printing Time | : 8 hour |


Set Parameters

By pressing the SET Button, the parameter entry screen is accessed. There are 3 separate passwords, these passwords have separate functions.



CALIBRATION MENU
PASSWORD: 3113

EKT SERIES




ERROR CODES

1. H1: Sensor Errors
 2. H2: Door Open Error
 3. H3: No Electricity Error
 4. H4: Alarm Upper Limit Exceed Alarm
 5. H5: Alarm Lower Limit Exceed Alarm
 6. H6: The Agitator Is Not Working Alarm
- Supply Voltage: 220Vac/50Hz


Control Panel Without USB Port



1		ENTER BUTTON
2		UPPER ADJUSTMENT BUTTON
3		LOWER ADJUSTMENT BUTTON




CONTROL AND FUNCTIONS OF TOUCH SCREEN PANEL



PIC5. Start Screen

CONTROL PANEL



PIC6. Touch Screen Control Panel

CONTROL BUTTONS	1		ON/OFF BUTTON
	2		SETTINGS BUTTON, ITS USED FOR TO ENTER THE MENU
	3		ALARM QUIET BUTTON
	4		LIGHT ON/OFF BUTTON, NO INTERIOR LIGHTING USAGE FOR FREEZERS
	5		USB DOES NOT READ / NOT READ SYMBOL
	6		TIME ADJUSTMENTS BUTTON
	7		PRINTER BUTTON
DIGITAL MONITOR			THE TEMPERATURE SYMBOL, SHOWS THE SET VALUE
			INSTANT VALUE OF CABIN TEMPERATURE
			CABIN TEMPERATURE/TIME GRAPHIC
			220 VOLT SUPPLY SYMBOL
			ALARM CONDITIONS SYMBOL

ON / OFF BUTTON



- Use for on/off the system.
- When press this button (For on and off)will request for password. (Depends on us).
- The device won't work directly after plug in, The password section will appear on the screen. The system will start after key-in the password.
- Password: ***11#**

BUTTON FOR HOME PAGE

- User can go back home page directly from another page by pressing this button.
- This button can use for resetting the main page.

SETTING MANU



- The device will request for password after pressing this button.
- The system will open the menu based on the entered password.
- The settings in the entered passwords may vary depending on the operating mode

USER SETTINGS PASSWORD: 2112



- SET TEMPERATURE: (Between 0C and 15C) **4C set**
- HYSTERESIS VALUE : (Between 0,0 and 2) **2C set**
- ALARM MIN. LIMIT :(Between 0 and +/- 10) **2C set**
- ALARM MAX. LIMIT :(Between 0 and +/-10) **8C set**
- DEFROST RANGE: (HOUR) **8 set**
- DEFROST TIME: (MINUTE) **1 set**
- Note: Setting for time and date by pressing this icon.
- User must key in the password 2112 to set the date and time.



SERVICE SETTING PASSWORD:1111



- Only EMSAS A.S authorized technician can access this menu.
- This menu will be accessible with a password by an authorized technician. The password will be shared only with authorized technicians.
- Password is **1111**.

Parameters to be applied in this menu:

- LANGUAGE: TURKISH/ **ENGLISH**
- USB: ON / OFF (Optional)
- PRINTER: ON / OFF (Optional)
- SAMPLE LIQUID: ON/OFF
- PRINT FREQUENCY DURATION: 05 Mins– 60 Mins (Optional)
- DEFROST TYPE: GAS / **RESISTANCE**
- AUTOMATIC PRINTER: ON / OFF (Optional)
- RECORD TYPE: **CABINET TEMPERATURE** / SAMPLE LIQUID / AVERAGE

SERVICE SETTINGS BUTTON PASSWORD: 3113



- The calibration menu will appear on the system after key in the password 3113.

The parameters are:

- Calibration 1: **+/-00,0**
- Calibration 2 : **+/-00,0**

SERVICE SETTINGS BUTTON PASSWORD: 4114



•When the service setting button is pressed and the 4114 password is entered, the system will enter the operating mode menu. Parameters to be included in the operating mode menu:

•4 – 8C Degree (Vaccine/ Medicine Bank Refrigerator)

- Deep Freezer (EE and EF Serie)
- Double Compressor (EKT – D Serie)
- Incubator (ECI Serie)
- Climatic Cabinet

Note: This system is prepared for the Vaccine/ Medicine Bank Refrigerator EKT series.

SILENT ALARM BUTTON



- This button mutes the alarm when the device starts to give an audible alarm for any reason.
- After the button is pressed, the device does not ring for a certain period of time when it is in alarm. This period depends on our request.
- After the specified time has passed, the alarm starts to ring again.

PRINT BUTTON



- This button is used for the thermal printer system.
- This is the button used if we want to record outside of the automatic recording time in the system setting.
- This button is used to manually record from the thermal printer.

LIGHT ON / OFF BUTTON



- To on/off the LED light manually inside the cabinet.
- When the freezer part is selected in the operating mode, this icon does not need to be seen on the main screen since deep freezers do not use LED light.

ALARM INDICATOR

- We can see two alarms at the same time on the main screen of the touch screen.
- When the device gives an alarm normally, it should appear as red stripes on the whole screen and instead of these two icons, it shows the error code and error name of the alarm.
- The entire screen turns red when the device alarms.

USB INDICATOR




- Indicates whether a USB Flash Disk is inserted in the device.
- After USB is plugged in, it will turn green and have a tick on it.
- When the USB is not plugged in, a small red cross may appear in its original color and above it.

COMPRESSOR IS ON



- Since our system works on / off, a sign should appear on the main screen when the compressor is activated. This sign should disappear after the compressor is deactivated. This cycle should continue as the compressor starts up and enters.
- Compressor icon can also be used instead of this icon.
- In addition, we can show whether the fan is active or not with this sign.

4.3 Electronic Wiring Diagram




5. CLEANING AND MAINTENANCE

5.1 Maintenance:

Periodic maintenance should be carried out once a month depending on the ambient conditions in which the device is located, and once every 6 months under normal conditions.

5.2 Cleaning

- The device is cleaned in our factory before delivery to the customer.
- Clean the device at ambient temperature after disconnecting the power cable from the device.
- Wipe your device with a damp and dry cloth against dirt and dust. Finally, plug in your cabinet and start it from the on-off button.
- Depending on the environmental conditions of your Medical Device, clean it regularly once a month or once every 6 months.
- In order to clean the condenser of your device, it is necessary to clean it from dust first.
- The condenser compartment of your device should be cleaned by a technical personnel with a vacuum cleaner or a soft paint brush, and the electrical components should not be damaged.
- Clean the condenser under your device with a damp cloth at least once every 4 months by pulling out the electrical outlet.



- The glass doors can clean with proper glass cleaner.
- Do not clean the engine compartment of your cooler with water, etc. It will damage the electrical components of your water closet and the user may be exposed to electric shocks.
- Wipe the plastic drainage bowl next to the compressor, where the defrosting water reaches, with a damp cloth.
- Do not use abrasive and purifying chemicals while cleaning inside and outside your cooler with a damp cloth, it can cause damage. Faults and errors that may arise from this reason will be treated out of warranty.

6. SAFETY

- Before doing cleaning and maintenance please review and read your safety instructions.
- If the device needs to be relocated or moved for maintenance, be sure to check the brakes of the wheels. Otherwise, the wheels may be damaged and broken. In this case, the balance of your device may deteriorate and may not work efficiently.
- Do not touch the unit with wet hand and/or feet
- Do not insert screwdrivers or other pointed objects between the moving parts of the device.
- The device cannot use by an unauthorized person.
- Before doing any kind of cleaning or maintenance operations please cut the electricity.
- Disconnect the main switch from the switching power before unplugging the appliance when cleaning or maintaining the appliance.
- Check that the device placed where it will be used is on flat ground and with four legs. It's dangerous to do any repair or service operations about device by unauthorized persons.
- When your device plug in do not any kind of maintenance and reparation.
- We strongly recommend you to use a voltage regulator in order not to damage the compressor in places where there may be larger changes.
- Do use your device according to the principles of the user manual.
- When you have a service request regarding your product, please contact our Service Center at the phone numbers above.

7. DISPOSAL

The user is responsible for the proper disposal of each part as waste. Electronic parts used, except power and current sources, should be disposed of according to the rules specified for the waste of electronic devices. When the product is to be disposed of, hospital, laboratory and national legislation should be followed.

8. WARNINGS & PRECAUTIONS

Pay attention to the warnings down below

- Do not use the device out of usage aim
- The device should operate after reading the manual book by authorized and educated personnel. In case of any problem, only authorized personnel can interfere to the device.
- The voltage in the electrical network must be suitable for the power of the device and must be grounded.
- Only spare parts and accessories provided by Emsaş A.Ş should be used.
- There should be no substances in the usable volume that may be affected by the operating temperature of the device and damage the device.
- We strongly recommend you to use a voltage regulator in order not to damage the compressor in places where there may be larger changes.
- When putting the vaccines in your device, do not stack them on the back to prevent air circulation. This will cause unwanted temperature differences in the upper and lower parts of your cooler.
- After operating your cooler, your cooler can give audible and visual alarms until it reaches the ideal operating temperature (0°C/+15°C). When your cooler reaches ideal operating temperature, the alarms shut down. Then you can load the products.
- When you receive your product confirm the warranty certificate your authorized dealer.

9. MEANING OF SYMBOLS

9.1. Meaning of Symbols on top of the Device



CE Certificate



Manufacturer Name and Address



Date of Manufacture

Indicates the date of manufacture of the medical device.



Catalog Number

Indicates the manufacturer's catalog number by which the medical device can be identified.



Serial Number

Indicates the manufacturer's serial number by which a particular medical device can be identified.



Fragile

Indicates that a medical device may break or be damaged if not handled carefully. It is located on the parcel.



Keep away from the sun.



Protect from heat and radioactive sources.

The device is affected by heat. For this reason, the device must be protected from heat sources.



Protect from moisture. Located on the parcel.



Storage conditions: 20°C / 45°C



Usage Instruction

A synonym for "refers to using instructions" is "refer to operating instructions".



This symbol draws attention to safety warnings and potentially dangerous situations. Failure to follow these warnings could result in damage to the device and injury to users.



Indicates that the user should refer to the instructions for use for important warning information such as warnings and precautions that cannot be presented on the medical device for many reasons.



Grounded Socket

Operate in a place with protective grounding



Alternative Current



Dangerous Voltage!

It is a symbol indicating the current available on the device. The place where there is electricity is also the symbol that should be specified.




Do not touch!



Do not touch with wet hand and/or feet!



Keep Upright. It is located on the parcel.



10. SERVICE

Your device can give audible and visual alarms in first operation. In case of this situation press the alarm silent button located on the display
Do not allow the repair of the device to anyone other than the authorized service during the warranty period. In case of interventions other than authorized service, the device will be subject to operation out of warranty.

When you encounter a problem with the device, call the authorized service or Emsaş numbers that made the first installation.

10.1 Troubleshooting Table

FAULT	CAUSE /CORRECTION
THE COOLER DOES NOT WORK THERE IS NO ELECTRICITY	<ul style="list-style-type: none"> *Make sure it is plugged in. *Make sure there is electricity in the mains. *If there is voltage outside the tolerances given in the network, the device it doesn't work. *Make sure the power button is on.
LIGHT IS NOT WORKING WHEN THE DOOR OPEN	<ul style="list-style-type: none"> * The lamp can be broken or it may have moved from the lamp socket. If its broken inform the service * The lamp switch may be damaged. Inform the service
THE TEMPERATURE DISTRIBUTION IN THE CABINET IS NOT HOMOGENEOUS ENOUGH, THE TOP AND BOTTOM ARE AT SEPARATE TEMPERATURES.	<ul style="list-style-type: none"> * Material may have been placed in front of the indoor fan. For better air circulation materials should not be placed in front of the fan. * The inner fan may be damaged, inform the service. * The fan switch may be damaged, inform the service.
THE DEVICE IS NOT COOLING	<ul style="list-style-type: none"> * Keep condenser clean. Do control fan in the back is working or not * If the fan is not working, inform the service * If the fan is working and the cabinet is not getting cold, inform the service
THE PRINTER IS NOT WORKING	Check the battery current. If the battery is 12V low, the battery is replaced or charged in a separate place. The paper is checked.
THE DEVICE MAKE NOISE	Check the battery voltage. If it is less than 10V, replace the battery and charge it separately. Check the printer settings as described in Chapter "Troubleshooting".
BATTERY IS NOT WORKING/ COMPRESSOR IS NOT WORKING	Do call your authorized service immediately

11. SPARE PARTS & ACCESSORIES

11.1 Spare Parts

There are no spare parts on the device that you can replace. Please notify our authorized service in case of malfunction.

11.2 Accessories

Warranty Certificate and Manual Book.

2 pieces shelves for EKT 80 model

2 pieces shelves for EKT 150 model

3 pieces shelves for EKT 175 model

4 pieces shelves for EKT 250 model

5 pieces shelves for EKT 425 and EKT 725

There is no other accessories than these

Contact

For any service request or training or questions about your device please use this guide kalite@emsas-as.com.tr.

For general information use info@emsas-as.com.tr or to see <http://www.emsas-as.com.tr>