

Certificate of Analysis

PrepFiler Express BTA™ Forensic DNA Extraction Kit

Product No. **4441351**
Lot No. **2408319**
Date of Manufacture **19AUG2024**
Expiration Date **09DEC2025**

DESCRIPTION

The PrepFiler Express BTA™ Forensic DNA Extraction Kit contains reagents and plastics for use in lysis of cells, removal of PCR inhibitors and elution of bound DNA.

KIT CONFIGURATION

Part No.	Component Description	Lot No.
4466367	PrepFiler™ Express Cartridges	2408365
4441396	PrepFiler BTA™ Lysis Buffer (pouch)	2406146
	4441404 BTL, PrepFiler BTA Lysis Buffer 14mL	2405082
4441401	Proteinase K (pouch)	2406160
	4441400 Tube Proteinase K	2406154
4468236	PrepFiler Express BTA™ Plastics (box)	2311019
100108928	AutoMate Express™ Tips and Tip Holders (bag)	JJ7230605

KIT SPECIFICATIONS

RESULT

Positive sample extraction test:

Diluted human DNA (corresponding to 100ng DNA per sample) is processed using the AutoMate Express™ automated extraction system and the recovered DNA quantified by RT-PCR.

PASS

*A passing result is achieved based on a detectable quantity of DNA as determined by RT-PCR.

No detectable human DNA contamination:

PrepFiler BTA™ Lysis Buffer and Proteinase K without the addition of any sample or sample DNA is processed using the AutoMate Express™ automated extraction system and the recovered material is tested for the presence of human DNA contamination using a sensitive STR profiling assay.

PASS

*A passing result is achieved if Globalfiler™ STR profiling @ 31 cycles meets the specifications as detailed in ISO18385 Section A.4.3 using an instrument-specific analytical threshold of 1 x LOQ (mean signal + 10SD) within the read region.

For Research, Forensic or Paternity Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 18385 and ISO 13485 certified quality management system. Site: Warrington, UK

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com.


Quality Assurance
Issued 06SEP2024

Certificate of Analysis

PrepFiler Express BTA™ Forensic DNA Extraction Kit

Product No. **4441351**
Lot No. **2408319**
Date of Manufacture **19AUG2024**
Expiration Date **09DEC2025**

KIT SPECIFICATIONS

RESULT

PrepFiler Express™ Cartridge Analytical specifications:

- Lysis Buffer: pH 7.25 – 7.45 Conductivity 19.80 – 24.20 mS/cm
- Wash Buffer A: Density 0.85 – 0.87 g/mL
- Wash Buffer B: Density 0.89 – 0.91 g/mL
- Elution Buffer: pH 7.90 – 8.10, Conductivity 400.0 – 650.0 mS/cm, DNase/RNase free

PASS

Additional checks:

- Cartridges undergo visual inspection to ensure the following:
 - No evidence of leakage or penetrating damage to foil surface
 - Foil alignment and placement is checked for consistency relative to label and top well
 - Cartridge label is checked to ensure it is visually straight with consistent placement
 - Sampling of cartridges is subjected to fill volume and vacuum leak testing checks during manufacture
- Kit and kit components undergo further visual inspection at both intermediate and final stages of kit manufacture

PASS

For Research, Forensic or Paternity Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 18385 and ISO 13485 certified quality management system. Site: Warrington, UK

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com.


Quality Assurance
Issued 06SEP2024