



DICHIARAZIONE DI CONFORMITA' DEL COSTRUTTORE

La società:

RIMSA P. LONGONI S.r.l.

Via Monterosa, 18/20/22 - 20831 SEREGNO (MB) - ITALIA

dichiara sotto la propria responsabilità che i Dispositivi Medici prodotti nell'anno 2019:

Modello:

PENTALED 28

Versionsi	mod. Pentaled 28 SO (dispositivo a soffitto)
	mod. Pentaled 28 PI (dispositivo a piantana)
	mod. Pentaled 28 PA (dispositivo a parete)
	mod. Pentaled 28+28 (dispositivo a soffitto doppio)

costruiti da RIMSA P.LONGONI S.r.l., sono conformi:

- *all'Allegato VII della Direttiva 93/42/CEE del 14/05/1993, applicata in Italia dal Decreto Legislativo N.46 del 24 Febbraio 1997 e successive varianti (inclusa la Direttiva 2007/47/CE del 05/09/2007, applicata in Italia dal Decreto Legislativo N.37 del 25 Gennaio 2010) e alle seguenti norme:*
 - IEC 60601-1 (Parte 1: Prescrizioni generali relative alla sicurezza fondamentale e alle prestazioni essenziali)
 - IEC 60601-2-41 (Parte 1: Norme particolari relative alla sicurezza fondamentale e alle prestazioni essenziali delle lampade scialitiche per uso chirurgico e per la diagnosi)
 - IEC 60601-1-2 (Parte 2: Prescrizioni generali per la sicurezza fondamentale e prestazioni essenziali- Norma collaterale: Compatibilità elettromagnetica – Prescrizioni e prove)
 - Classificazione in riferimento all'articolo 9 e allegato IX della Direttiva 93/42/CEE e 2007/47/CE
Durata: Breve termine (Allegato IX, Par.1 "Definizioni", art.1, comma 1.1)
Descrizione: Disp. Medico non invasivo (Allegato IX, Par.1 "Definizioni", art.1, comma 1.2)
Disp. Medico attivo (Allegato IX, Par.1 "Definizioni", art.1, comma 1.4)
CLASSE: Classe I (Allegato IX, par.3 "Classificazione", art.3, comma 3.3, Regola 12) e (Allegato IX Par.3 "Classificazione", art.1, comma 1.1 Regola 1)
 - Riferimento fascicolo tecnico Cod. RIM-FT019.
 - La valutazione di conformità è sviluppata in riferimento all'articolo 11 della Direttiva 93/42/CEE e 2007/47/CE.

Il Sistema Qualità di RIMSA è conforme alle norme UNI EN ISO 9001 e UNI CEI EN ISO 13485 ed è certificato da CSQ (certificato CSQ n.9120.RMS1 e 9124.RMS2).

Seregno, 07/01/2019

Luogo e data

Timbro e firma del Consigliere Delegato
(Paolo Longoni)



MANUFACTURER'S DECLARATION OF CONFORMITY

The company

RIMSA P. LONGONI S.r.l.

Via Monterosa, 18/20/22 - 20831 SEREGNO (MB) - ITALY

declares on its own responsibility that the Medical Devices made in the year 2019:

Model:

PENTALED 28

Versions	mod. PentaLED 28 SO (ceiling device)
	mod. PentaLED 28 PI (mobile device)
	mod. PentaLED 28 PA (wall device)
	mod. PentaLED 28+28 (double-ceiling device)

constructed by RIMSA P.LONGONI S.r.l., conforms:

- *to the Attachment VII of the 93/42/EEC Directive of May 14th 1993, and subsequent variations (including the 2007/47/EC Directive of September 5th 2007); moreover, it is in conformity with the following safety regulation:*

- IEC 60601-1 (Part 1: General requirements for basic safety and essential performance)
- IEC 60601-2-41 (Part 2: Particular requirements for basic safety and essential performance of surgical luminaires and luminaires for diagnosis)
- IEC 60601-1-2 (Part 1: General requirements for basic safety and essential performance – collateral standard: Electromagnetic compatibility - Requirements and tests)
- Classification with reference to article 9 and Attachment IX of 93/42/EEC Directive and 2007/47/EC
Duration : Short term (Annex IX, Par.1 "Definitions", art.1, sub-section 1.1)
Description: Non invasive Medical Device (Annex IX, Par.1 "Definitions", art.1, sub-section 1.2)
Active Medical Device (Annex IX, Par.1 "Definitions", art.1, sub-section 1.4)
Class: Class I (Annex IX, Par.3 "Classification", art.3, sub-section 3.3, Rule 12)
(Annex IX, Par.3 "Classification", art.1, sub-section 1.1, Rule 1)
- Reference to technical file Code RIM-FT019.
- The conformity assessment is developed in reference to article 11 of the 93/42/EEC Directive and 2007/47/EC .

The RIMSA Quality System conforms to the UNI EN ISO 9001 and UNI EN ISO 13485 regulations and is certified by the CSQ (certified CSQ n.9120.RMS1 and 9124.RMS2).

Seregno, 07/01/2019

Place and date



Mark and sign of Managing Director
(Paolo Longoni)