## DECLARATION OF CONFORMITY TO REGULATION(EU) 2017/745 ON MEDICAL DEVICES



CONTEC MEDICAL SYSTEMS CO., LTD

NO.112 QINHUANG WEST STREET, ECONOMIC & TECHNICAL

DEVELOPMENT ZONE, QINHUANGDAO, HEBEI PROVINCE,

PEOPLE'S REPUBLIC OF CHINA

SRN of Manufacturer: CN-MF-000007715



SHANGHAI INTERNATIONAL HOLDING CORP. GMBH(EUROPE)
EIFFESTRASSE 80, 20537 HAMBURG GERMANY

SRN of Authorised Representative: DE-AR-00000001

This EU declaration of conformity is issued under the sole responsibility of the manufacturer.

We keep all supporting documentation and ensure that the authorised representative has the necessary documentation permanently available.

69450401IGN0008C2; 69450401IGN0009C4;

69450401IGN0010BM; 69450401IGN0058CH;

**BASIC UDI-DI:** 69450401IGN0050BZ; 69450401IGN0060C4;

69450401IGN0004BS; 6945401IGN000594; 6945401IGN000579P; 69450401IGN0003BQ;

69450401IGN0006BW

PRODUCT AND TRADE NAME: CUFF

IGN0008; IGN0009; IGN0010; IGN0058;

CATALOGUE NUMBER/MODEL: IGN0050; IGN0060; IGN0004; IGN0005;

IGN0057; IGN0003; IGN0006

RISK CLASS OF THE DEVICE: Class I according to rule 13 Annex VIII

We, (CONTEC MEDICAL SYSTEMS CO., LTD) herewith declare that the stated medical devices meet REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on medical devices.

CONFORMITY ASSESSMENT PROCEDURE: Regulation (EU) 2017/745, Annex II + III

PLACE, DATE OF ISSUE: QINHUANGDAO, 2021/10/11

HUKUN, Chairman/ manufacturer

NAME AND FUNCTION, SIGNATURE:



MDR210930-DOC	Ver: A
Page 1 of 1	