



EC Declaration of Conformity according to Annex III of the IVDD

This is to certify that following IVD products:

Product name	Product code
Rota-Strip	C-1001
Rota Uni-Strip	C-1501
Combi-Strip	C-1004
Combi K-SeT	K-1204
Combi K-SeT	K-1504
RSV Respi-Strip	C-1006
RSV K-SeT	K-1206
RSV K-SeT	K-1506
Adeno Respi-Strip	C-1009
Adeno Respi K-SeT	K-1209
Adeno Respi K-SeT	K-1509
Influ A+B K-SeT	K-1212
Influ A+B K-SeT	K-1512
Giardia Strip	C-1013
Giardia K-SeT	K-1513
Legionella K-SeT	K-1215
Legionella K-SeT	K-1515
GastroVir K-SeT	K-1516
Crypto/Giardia Duo-Strip	C-1018
Pylori-Strip	C-1019
Pylori K-SeT	K-1519
Clostridium K-SeT	K-1220
Clostridium K-SeT	K-1520
COVID-19 Ag Respi-Strip	C-1023
COVID-19 Ag Respi-Strip	C-1123
COVID-19 Ag Respi-Strip	C-1223
COVID-19 Ag Respi-Strip	C-1323
COVID-19 Sero NP/RBD	K-1224



COVID-19 Ag K-SeT	K-1525
HAT Sero K-SeT	K-12S2
HAT Sero K-SeT	K-15S2
OXA-48 K-SeT	K-15R1
RESIST-3 O.O.K. K-SeT	K-15R4
RESIST-3 O.O.K. K-SeT	K-15R4/XBRI
RESIST-3 O.O.K. K-SeT	K-11R4/XBRI
RESIST-3 O.K.N. K-SeT	K-15R5
OXA-23 K-SeT	K-15R7
O.K.N.V.I. RESIST-5	K-15R11
O.K.N.V.I. RESIST-5	K-15R11/XBRIT
O.K.N.V.I. RESIST-5	K-11R11/XBRIT
RESIST ACINETO	K-15R13
RESIST CTX-M	K-15R14
IMP K-SeT	K-15R10
BL-RED 25	RED-0001
SARS-CoV-2 RT-LAMP	L-0001
Flu A & FLU B RT-LAMP	L-0002
Negative control	CTR-1000
Rotavirus Positive Control	P-1001
Adenovirus Positive Control	P-1002
RSV Positive Control	P-1006
Influenza A Positive Control	P-1010
Giardia Positive Control	P-1013
Legionella Positive Control	P-1015
Pylori Positive Control	P-1019
C difficile Positive Control	P-1020
RESIST penta O.K.N.V.I. control	P-10R11
OXA-163 Positive Control	P-10R4-1
OXA-23 Positive Control	P-10R7
COVID-19 Ag Positive Control	P-1023
Proguanil / Malarone – Strip	C-10T1
Mefloquine / Lariam - Strip	C-10T2

are manufactured and sold by

Coris BioConcept
CREALYS Science Park
Rue Guillaume Fouquet, 11
5032 Gembloux, BELGIUM





EC Declaration of Conformity according to Annex III of the IVDD

These products:

1. Belong to the Class “Others/General” as they are not for self-testing and do not belong to List A or List B of Annex II of IVDD (98/79 EC).
2. Comply with all Essential Requirements (Annex I) of the IVDD (98/79 EC)
3. This compliance has been properly documented using a checklist created from Annexes I and III of the IVDD, linked to all supporting Technical Documentation. This documentation included both product specific and process (Quality System) specific documents.
4. Have a Quality System in place based on ISO 13485.
5. This Declaration is issued by Coris BioConcept and has an unlimited time validity.
6. This Declaration of Conformity is signed below, certifying these requirements have been met and documented.

For Coris BioConcept, made in Gembloux the 09th of May 2023

Thierry LECLIPTEUX, PhD
CEO
Coris BioConcept



Certificate BE23/00000111

The management system of

Coris Bioconcept

CREALYS Science Park - Rue Guillaume Fouquet, 11, 5032 Gembloux, Belgium

has been assessed and certified as meeting the requirements of

ISO 13485:2016

EN ISO 13485:2016

For the following activities

Design, development, manufacture and distribution of in vitro diagnostic tests for the detection of pathogens in the diagnosis of respiratory, gastric, enteric and parasitic diseases, the detection of resistance to antibiotics and the detection in urine of therapeutics, used for the treatment of these infectious diseases.

Distribution of instrument for electrochemical detection to be used with Coris' kit.

This certificate is valid from 05 May 2023 until 20 August 2024 and remains valid subject to satisfactory surveillance audits.

Issue 1. Certified since 05 May 2023

Organization certified since 07 April 2021 and first certified by SGS under UKAS since 05 May 2023.



Authorised by
Jonathan Hall
Global Head - Certification Services

SGS United Kingdom Ltd
Rossmore Business Park, Ellesmere Port, Cheshire, CH65 3EN, UK
t +44 (0)151 350-6666 - www.sgs.com



This document is an authentic electronic certificate for Client' business purposes use only. Printed version of the electronic certificate are permitted and will be considered as a copy. This document is issued by the Company subject to SGS General Conditions of certification services available on [Terms and Conditions | SGS](#). Attention is drawn to the limitation of liability, indemnification and jurisdictional clauses contained therein. This document is copyright protected and any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful.



O.K.N.V.I. RESIST-5



www.corisbio.com
IFU-58R11/TB/08

Manufacturer:

Coris BioConcept
CREALYS Science Park
Rue Guillaume Fouquet, 11
5032 GEMBLoux
BELGIUM
Tel.: +32(0)81.719.917
Fax: +32(0)81.719.919
info@corisbio.com

Produced in BELGIUM

When the provided buffer containing the resuspended bacteria comes into contact with the membrane, the solubilised conjugates migrate with the sample by passive diffusion, while conjugates and sample material come into contact with the immobilised respective antibodies that are adsorbed onto the nitrocellulose strip. If the sample contains an OXA-48, KPC, NDM, VIM or IMP carbapenemase, the respective complexes made of the conjugates and either OXA-48, or KPC, or NDM or VIM or IMP will remain bound to their respective specific lines (OXA-48: "O" line; KPC: "K" line; NDM: "N" line, VIM: "V" line, IMP: "I" line). The migration continues by passive diffusion and both conjugates and sample material come into contact with the (upper) line control reagent that binds a control conjugate ("C" line), thereby producing a red line.

The result is visible within 15 minutes in the form of red lines on the strip.

III. REAGENTS AND MATERIALS

1. O.K.N.V.I. RESIST-5 (2x20 cassettes)

20 sealed pouches containing two lateral-flow cassettes and one desiccant. Each cassette contains one sensitised strip.

2. LY-D buffer vial (7 mL)

Tris-EDTA solution containing NaN_3 (<0.1%) and a detergent.

3. Instruction for use (1)

4. Disposable collection tubes (20)

5. Disposable transfer pipettes (20)

Materials to be ordered separately:

- RESIST-BC (S-1001): reagents kit for use with blood culture
- ReSCape (S-1002): reagents kits for use with rectal swab

IV. SPECIAL PRECAUTIONS

- All operations linked to the use of the test must be performed in accordance with good laboratory practices.

- All reagents are for *in vitro* diagnostic use only.

- Pouch must be opened with care.

- Avoid touching nitrocellulose with your fingers.

- Wear gloves when handling samples.

- Never use reagents from another kit.

- Green or blue lines indicate immunoreagents adsorption sites. Green or blue colour disappears during the test.

- The quality of the reagents cannot be guaranteed beyond their shelf-life dates or if reagents are not stored under required conditions as indicated in the insert.

V. WASTE DISPOSAL

- Dispose of gloves, swabs, test tubes and used devices in accordance with GLP.

- Each user is responsible for the management of any waste produced, and must ensure that it is disposed of in accordance with the applicable legislation.

VI. STORAGE

- An unopened pouch may be kept at between 4 and 30°C and used until the shelf-life date indicated on the packaging. Once the pouch is opened, run the test immediately.

- Avoid freezing devices and buffer.

VII. SPECIMEN HANDLING AND COLLECTION

Specimens to be tested should be obtained and handled by standard microbiological methods. Make sure that the specimens are not treated with solutions containing formaldehyde or its derivatives.

The performance of was established with colonies from Mueller Hinton agar, Columbia agar + 5% sheep blood, CLED agar, UriSelect™ 4; Mac Conkey agar, Drigalski agar, CHROMagar™ mSuperCarba®, ChromID® ESB� agar, ChromID®Carba Smart, ChromID® OXA-48, ChromID® CPS® Elite agar, CHROMagar™ KPC, Brilliance™ CRE agar, Brilliance™ ESB� agar, CHOCOLATE PolyViteX™ agar, TSA+5% sheep blood.

VIII. PROCEDURE

PREPARATIONS OF THE TEST:

Allow kit components, in unopened packaging, and specimens (in the event that the plate containing colony to be tested was kept at 4°C) to equilibrate at room temperature (15-30°C) before performing a test.

Open the pouch and remove the device. Once opened, run the test immediately. Indicate the patient's name or specimen number on the device (one device per sample).

SPECIMEN PREPARATION PROCEDURE:

Performance claims with regard to sample types other than bacterial colonies have been established for rectal swabs and blood cultures.

With rectal swabs and blood cultures, the preparation procedure has to be followed as described in the respective kits (S-1002, ReSCape and S-1001, RESIST-BC)

With bacterial colonies, we recommend the use of fresh agar cultures for optimal test performance and as followed:

1. Prepare one collection tube and add **11 drops** of LY-D buffer in the tube
2. Harvest bacteria by taking **3 colonies** with a disposable bacteriological loop and dip the loop in the bottom of the tube containing the buffer. The same bacteriological loop can be used to collect the 3 colonies.
3. Stir thoroughly before removing the loop.
4. Close the tube and vortex the preparation to homogenize.
5. Use the transfer pipette provided in the kit and add 100 µL of diluted sample into the sample well of each of the two cassettes labelled (i) NDM, KPC and OXA-48 and (ii) IMP and VIM (**diluted sample must reach the black line indicated on the transfer pipette to accurately aspirate 100 µL**).
6. Allow to react for 15 minutes and read the result

In vitro rapid diagnostic test for the detection of OXA-48, KPC, NDM, VIM and IMP carbapenemases in bacterial culture

FOR IN VITRO DIAGNOSTIC USE FOR PROFESSIONAL USE ONLY

EN

References: K-15R11, 2x20 cassettes, buffer, 20 tubes and transfer pipettes

(EN) For Instructions For Use in your language : (FR) Pour obtenir les notices dans la langue de votre choix : (ES) Para las instrucciones de uso en su idioma : (PT) Para Instruções de Uso na sua língua : (IT) Per le Istruzioni di Uso nella sua lingua : (DE) Für Gebrauchsanleitungen in Ihrer Sprache : (NL) Voor Gebruiksaanwijzing in uw eigen taal : (PL) Instrukcja obsługi w języku użytkownika:	website: ifu.corisbio.com key-code: COR58R1104
	(EU) +800 135 79 135 (non-EU) +31 20 794 7071 (CA) +1 855 805 8539 (AR, CO, UY, AU, NZ) +800 135 79 135

I. INTRODUCTION

Carbapenemase-producing Organisms (CPO), and more specifically, Carbapenem-resistant Enterobacteriaceae (CRE) represent a major public health concern worldwide due to their broad spectrum of resistance to antibiotics including, besides carbapenems, most classes of antimicrobial agents, and thus leaving very few options for the management of infected patients. Besides CREs, CPOs also include nonfermenting Gram-negative bacilli (NFGNB), such as *Pseudomonas aeruginosa* and *Acinetobacter baumannii* that exhibit resistance not only to beta lactam and other groups of antibiotics, but also to carbapenems. The rapid spread of CPOs and genes encoding these resistances has led to nosocomial outbreaks and endemic situations worldwide.

Development of new rapid diagnostic tests to track antimicrobial resistance patterns is considered as one of the priority core actions by international experts and health authorities. NDM and KPC represent two of the most increasing and prevalent carbapenemases in many countries. On the other hand, class D OXA-48 type carbapenemases are the most challenging resistance mechanisms to be detected by clinical laboratories. VIM is not only present in Enterobacteriaceae but is also highly prevalent in non-fermenting bacteria. IMP should be regarded as a potential problem since they degrade not only C3G but also carbapenem antimicrobial drug like Imipenem. IMP prevalence is the lowest, apart from Japan where it is more prevalent. Inhibitor-based phenotypic confirmatory tests exist for the confirmation of class A (KPC) and class B (VIM, IMP, NDM) carbapenemases. Nowadays, definitive confirmation of CPO resistance mechanism relies on molecular assays. These tests are expensive and can only be performed in dedicated environment and by skilled personnel, hence limiting their more generalized usage.

O.K.N.V.I. RESIST-5 test is part of Coris BioConcept RESIST range of antimicrobial resistance diagnostic tests.

II. PRINCIPLE OF THE TESTS

These tests are ready to use and are based on a membrane technology with colloidal gold nanoparticles. Our kit is aimed to detect and identify the carbapenemases from a bacterial colony isolate of Enterobacteriaceae or NFGNB growing on agar plate. Each pouch contains: 2 lateral-flow cassettes for the identification of (i) OXA-48, KPC, NDM and (ii) VIM and IMP.

Identification of OXA-48, KPC and NDM. A nitrocellulose membrane is sensitised with:


- (1) a monoclonal antibody directed against OXA-48 carbapenemase and variants (except OXA-163-like enzymes) ("O" line)
- (2) a monoclonal antibody directed against KPC carbapenemase ("K" line)
- (3) a monoclonal antibody directed against NDM carbapenemase ("N" line)
- (4) a control capture reagent (upper "C" line).

Four different colloidal gold nanoparticles conjugates are dried on a membrane: a conjugate directed against a second epitope of the OXA-48 carbapenemase, a conjugate directed against a second epitope of the KPC carbapenemase, a third conjugate specific to NDM carbapenemase and a control conjugate to validate the test conditions.

Identification of VIM and IMP. A nitrocellulose membrane is sensitised with:

- (1) a monoclonal antibody directed against VIM carbapenemase ("V" line),
- (2) a monoclonal antibody directed against IMP carbapenemase ("I" line)
- (3) a control capture reagent (upper "C" line).

Three different colloidal gold nanoparticles conjugates are dried on a membrane: a conjugate directed against VIM carbapenemase, a conjugate directed against IMP carbapenemase and a control conjugate.



Positive results may be reported as soon as the test and control lines become visible. Do not take the appearance of new lines into account after the reaction time has passed.

The result must be read on still wet strip.

IX. INTERPRETING RESULTS

The results are to be interpreted as follows for each of the two cassettes:

Negative test result: a reddish-purple line appears across the central reading window at the Control line (C) position. No other line is present.

Positive test result: in addition to a reddish-purple line at the Control line (C), a visible reddish-purple line appears at one of the Test lines position ("N" or "K" or "O") on cassette labelled (i) NDM, KPC, OXA-48 or at one of the Test lines position ("I" or "V") on cassette labelled (ii) IMP and VIM. Intensity of the test line may vary according to the quantity of antigens as well as of the variant type present in the sample. Any reddish-purple test line (OXA-48, KPC, NDM, VIM and IMP), even weak, should be considered as a positive result.

If a positive test line appears beside of the "O" mark, the sample contains OXA-48 or OXA-48-like variants. If it appears beside the "K" mark, the sample contains KPC variants; beside the "N" mark, the sample contains NDM; the "V" mark, the sample contains VIM; and beside of the "I" mark, IMP is present in the sample. Combinations of positive test lines can occur.

In this case the sample contains several carbapenemases.

Invalid test result: The absence of a Control line indicates a failure in the test procedure. Repeat invalid tests with a new test device.

Note: during the drying process, a very faint shadow may appear at the Test line positions. It should not be regarded as a positive result.

Molecular method	Positive	Negative	Total
NDM test			
Positive	40	0	40
Negative	0	140	140
Total	40	140	180

95 % Confidence Interval ¹

Sensitivity:	100 %	(89.1 to 100 %)
Specificity:	100 %	(96.7 to 100 %)
Positive Predictive value:	100 %	(89.1 to 100 %)
Negative predictive value:	100 %	(96.7 to 100 %)
Agreement:	100 %	(180/180)

Molecular method	Positive	Negative	Total
VIM test			
Positive	43	0	43
Negative	3	134	137
Total	46	134	180

95 % Confidence Interval ¹

Sensitivity:	93.5 %	(81.1 to 98.3 %)
Specificity:	100 %	(96.5 to 100 %)
Positive Predictive value:	100 %	(89.8 to 100 %)
Negative predictive value:	97.8 %	(93.2 to 99.4 %)
Agreement:	98.3 %	(177/180)

Molecular method	Positive	Negative	Total
IMP test			
Positive	19	0	19
Negative	0	161	161
Total	19	161	180

95 % Confidence Interval ¹

Sensitivity:	100 %	(79.1 to 100 %)
Specificity:	100 %	(97.1 to 100 %)
Positive Predictive value:	100 %	(79.1 to 100 %)
Negative predictive value:	100 %	(97.1 to 100 %)
Agreement:	100 %	(180/180)

The O.K.N.V.I. RESIST-5 kit was also validated with rectal swabs and blood cultures.

C. Repeatability and reproducibility

To check intra-batch accuracy (repeatability), the same positive samples and a buffer solution were processed 15 times on kits of the same production batch in the same experimental conditions. All observed results were confirmed as expected.

To check inter-batch accuracy (reproducibility), some samples (positive and buffer) were processed on kits from three different production batches. All results were confirmed as expected.

XI. LIMITS OF THE KIT

The test is qualitative and cannot predict the quantity of antigens present in the sample. Clinical presentation and other test results must be taken into consideration to establish diagnosis. A positive test does not rule out the possibility that other antibiotic resistance mechanisms may be present.

XII. TECHNICAL PROBLEMS / COMPLAINTS

If you face a technical problem or if performances do not correspond with those indicated in this package insert:

- Record the lot number of the kit concerned.
- If possible, keep the sample in the appropriate storage condition during the complaint management.
- Contact Coris BioConcept (client.care@corisbio.com) or your local distributor.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

XIII. BIBLIOGRAPHIC REFERENCES

- J. Wesley MacDonald and V. Chibbhai Evaluation of the RESIST-4 O.K.N.V immunochromatographic lateral flow assay for the rapid detection of OXA-48, KPC, NDM and VIM carbapenemases from cultured isolates *Access Microbiology* 2019;1
- T. Pilate, S. Desmet Detection of carbapenemase-producing *Pseudomonas aeruginosa* in a tertiary care centre Annual Meeting of the Royal Belgian Society of Laboratory Medicine November 15th, 2019 Belgium
- Oueslati S, Iorga BI, Tiliu L, Exilie C, Zavala A, Dortet L, Jousset AB, Bernabeu S, Bonnin RA, Naas T. Unravelling ceftazidime/avibactam resistance of KPC-28, a KPC-2 variant lacking carbapenemase activity. *J Antimicrob Chemother.* 2019 Aug 1;74(8):2239-2246
- Brolund A, Lagerqvist N, Byfors S, Struelens MJ, Monnet DL, Albiger B, Kohlenberg A. Worsening epidemiological situation of carbapenemase-producing Enterobacteriaceae in Europe, assessment by national experts from 37 countries, July 2018. *Euro Surveill.* 2019 Feb; 24 (9) 1560-7917
- Oliveira J, Reygaert WC. Gram Negative Bacteria. *StatPearls Publishing; 2019 Jan-2019*
- Baeza LL, Pfennigwerth N, Greissl C, Götting S, Saleh A, Stelzer Y, Gatermann SG, Hamprecht A. Comparison of five methods for detection of carbapenemases in Enterobacteriales with proposal of a new algorithm. *Clin Microbiol Infect.* 2019 Mar 18. pii: S1198-743X(19)30104-1
- Rösner S, Kamalanabhaiah S, Küsters U, Kolbert M, Pfennigwerth N, Mack D. Evaluation of a novel immunochromatographic lateral flow assay for rapid detection of OXA-48, NDM, KPC and VIM carbapenemases in multidrug-resistant Enterobacteriaceae. *J Med Microbiol.* 2019 Mar;68(3):379-381.
- Glupczynski Y, Evrard S, Huang TD, Bogaerts P. Evaluation of the RESIST-4 K-SeT assay, a multiplex immunochromatographic assay for the rapid detection of OXA-48-like, KPC, VIM and NDM carbapenemases. *J Antimicrob Chemother.* 2019 Feb 6. doi: 10.1093

Last update 26 AUGUST 2025

REF	Catalogue number		Manufacturer
IVD	In vitro diagnostic medical device		Temperature limits
	Contains sufficient for <n> tests	LOT	Batch code
	Consult instructions for use		Do not reuse
	Keep dry		Use by
DIL SPE	Diluent specimen	CONT NaN ₃	Contains Sodium azide
UDI	Unique device identifier		

X. PERFORMANCE

A. Detection Limit

The detection limit determined with purified recombinant proteins of OXA-48, KPC, NDM, VIM and IMP have been evaluated at 0.25 ng/mL, 0.5 ng/mL, 0.0625 ng/mL, 0.23 ng/mL and 0.1 ng/mL, respectively.

B. Retrospective study

The test cassettes were validated by comparison with reference molecular method (validated in house multiplex PCR including sequencing) in a retrospective study performed on 180 non duplicated, consecutive suspected CPE clinical isolates collected between 2012 and 2021 from Belgian hospitals.

Molecular method	Positive	Negative	Total
OXA-48 test			
Positive	41	0	41
Negative	0	139	139
Total	41	139	180

95 % Confidence Interval ¹

Sensitivity:	100 %	(89.3 to 100 %)
Specificity:	100 %	(96.6 to 100 %)
Positive Predictive value:	100 %	(89.3 to 100 %)
Negative predictive value:	100 %	(96.7 to 100 %)
Agreement:	100 %	(180/180)

Molecular method	Positive	Negative	Total
KPC test			
Positive	24	0	24
Negative	0	156	156
Total	24	156	180

95 % Confidence Interval ¹

Sensitivity:	100 %	(82.8 to 100 %)
Specificity:	100 %	(97.0 to 100 %)
Positive Predictive value:	100 %	(82.8 to 100 %)
Negative predictive value:	100 %	(97.0 to 100 %)
Agreement:	100 %	(180/180)

¹ Newcombe, Robert G. "Two-Sided Confidence Intervals for the Single Proportion: Comparison of Seven Methods," *Statistics in Medicine*, 17, 857-872 (1998).

RESIST ACINETO



www.corisbio.com
IFU-58R13/EN/04

Manufacturer:
Coris BioConcept
CREALYS Science Park
Rue Guillaume Fouquet, 11
5032 GEMBLOUX
BELGIUM
Tel.: +32(0)81.719.917
Fax: +32(0)81.719.919
info@corisbio.com
Produced in BELGIUM

***In vitro* rapid diagnostic test for the detection of OXA-23, OXA-40, OXA-58 and NDM carbapenemases in bacterial culture**

FOR IN VITRO DIAGNOSTIC USE FOR PROFESSIONAL USE ONLY



References: K-15R13, 20 cassettes, buffer, 20 tubes and transfer pipets

I. INTRODUCTION

Acinetobacter baumannii is an important opportunistic and multidrug-resistant Gram-negative bacteria responsible for nosocomial infections in health facilities. If left untreated, this infection can lead to septicemia and death. The carbapenem-hydrolysing oxacillinases (OXAs) are the most commonly reported carbapenem-resistance determinants in *Acinetobacter* spp., particularly in *A. baumannii*, with OXA-23 as the most prevalent carbapenem-resistance determinant observed in these isolates. OXA-40 (=24) and OXA-58 are also often encountered while other OXAs are less frequent. Recently *Acinetobacter* spp. harbouring OXA's together with NDM have emerged, particularly because of mobile genetic elements co-harboring NDM and OXA encoding genes. Mobile genetic elements (incl. plasmids) constitute reservoirs for horizontal transmission of these resistance factors. Detection of these resistance factors, not only in resistant species but also in carrier species, is therefore of paramount importance in the control of antibiotic resistance in the hospital. Nowadays, definitive confirmation of OXA-23, OXA-40, OXA-58, other less frequent OXAs and NDM relies on molecular amplification analysis and DNA sequencing. These tests are expensive and can only be performed in dedicated environment and by skilled staff, hence limiting their more generalized usage. The development of new rapid diagnostic tests to track antimicrobial resistance patterns is considered as one of the priority core actions by international experts and health authorities. The RESIST ACINETO test aiming at a rapid identification of the OXA-23, OXA-40/58 and related OXAs and NDM carbapenemases ensures effective treatment of patients and prevention of spread of carbapenemases *Acinetobacter* spp. carrier, especially in hospitals.

II. PRINCIPLE OF THE TEST

This test is ready to use and is based on a membrane technology with colloidal gold nanoparticles. Our kit is aimed to detect and identify the carbapenemases from bacterial colony isolate of *Enterobacteriaceae* or NFGNB growing on agar plate.

A nitrocellulose membrane is sensitized with:

- (1) a monoclonal antibody directed against NDM carbapenemase ("NDM" line)
- (2) a monoclonal antibody directed against OXA-23 carbapenemase ("O23" line)
- (3) a monoclonal antibody directed against OXA-40 and OXA-58 carbapenemases ("O40/58" line)
- (4) a control capture reagent (upper "C" line).

There are different conjugates coupled to colloidal gold particles which are dried on a membrane: a conjugate directed against a second epitope of the NDM carbapenemase, a conjugate directed against a second epitope of the OXA-23 carbapenemase, a third conjugate specific to OXA-40 carbapenemase, a fourth conjugate specific to OXA-58 carbapenemase and a control conjugate to valid the test conditions.

This test is aimed at the detection of NDM, OXA-23, OXA-40 and OXA-58 carbapenemases (and related variants) on colonies of *Enterobacteriaceae* isolates growing on agar plate.

When the provided buffer containing the resuspended bacteria comes into contact with the membrane, the solubilised conjugates migrate with the sample by passive diffusion, while conjugates and sample material come into contact with the immobilised respective antibodies that are adsorbed onto the nitrocellulose strip. If the sample contains an NDM, OXA-23, OXA-40 or OXA-58 carbapenemase, the respective complexes made of the conjugates and their specific targets will remain bound to their respective specific lines (NDM: "NDM" line; OXA-23: "O23" line; OXA-40 or OXA-58: "O40/58" line). The

migration continues by passive diffusion and both conjugates and sample material come into contact with the (upper) line control reagent that binds a control conjugate ("C" line), thereby producing a red line.

The result is visible within 15 minutes in the form of red lines on the strip.

III. REAGENTS AND MATERIALS

1. RESIST ACINETO (20)

20 sealed pouches containing one device and one desiccant. Each device contains one sensitized strip.

2. LY-A buffer vial (15 mL)

Saline solution buffered to pH 7.5 containing TRIS, Na₃N (<0,1%) and a detergent.

3. Instruction for use (1)

4. Disposable collection tubes (20)

5. Disposable transfer pipettes (20)

Materials to be ordered separately:

- RESIST-BC (S-1001): reagents kit for use of blood culture
- ReSCape (S-1002): reagents kits for use of rectal swab

IV. SPECIAL PRECAUTIONS

- All operations linked to the use of the test must be performed in accordance with good laboratory practices.
- All reagents are for *in vitro* diagnostic use only.
- Pouch must be opened with care.
- Avoid touching nitrocellulose with your fingers.
- Wear gloves when handling samples.
- Never use reagents from another kit.
- Green lines indicate immunoreagents adsorption sites. Green colour disappears during the test.
- Reagents' quality cannot be guaranteed beyond their shelf-life dates or if reagents are not stored under required conditions as indicated in the insert.

V. WASTE DISPOSAL

- Dispose of gloves, swabs, test tubes and used devices in accordance with good laboratory practices.
- Each user is responsible for the management of any waste produced, and must ensure that it is disposed of in accordance with the applicable legislation.

VI. STORAGE

- An unopened pouch may be kept at between 4 and 30°C and used until the shelf-life date indicated on the packaging. Once the pouch is opened, perform the test immediately.
- Avoid freezing devices and buffer.

VII. SPECIMEN HANDLING AND COLLECTION

Specimens to be tested should be obtained and handled by standard microbiological methods.

Make sure that the specimens are not treated with solutions containing formaldehyde or its derivatives.

Culture media tested and validated with Coris BioConcept RESIT kits are listed on the website: <https://www.corisbio.com/products/oknvi-resist-5/faq>

VIII. PROCEDURE

PREPARATIONS OF THE TEST:

Allow kit components, in unopened packaging, and specimens (in case the plate containing colony to be tested was kept at 4°C) to reach room temperature (15-30°C) before performing a test.

Open the pouch and remove the device. Once opened, perform the test immediately. Indicate the patient's name or specimen number on the device (one device per sample).


SPECIMEN PREPARATION PROCEDURE:

Performance claims with regard to sample types other than bacterial colonies have been established for rectal swabs and blood cultures.

With rectal swabs and blood cultures, the preparation procedure to be followed is that described in the respective kits (S-1002, ReSCape; S-1001, RESIST-BC).

With bacterial colonies, we recommend the use of fresh agar cultures for optimal test performance and as followed:

1. Prepare one tube and add **6 drops** of LY-A buffer in the tube.
2. Harvest bacteria by taking **1 colony** with a disposable bacteriological loop and dip the loop in the bottom of the collection tube containing the buffer.
3. Stir thoroughly before removing the loop.
4. Close the tube and vortex the preparation to homogenize.
5. Use transfer pipette and add 100 µL of diluted sample into the sample well of the cassette labelled NDM, OXA-23 and OXA-40/58 (**diluted sample must reach the black line indicated on the transfer pipette to accurately aspirate 100 µL**).
6. Allow to react for 15 min max and read the result.




Positive results may be reported sooner the moment the test and control lines become visible.


Do not take the appearance of new lines into account after the reaction time is passed.

The result must be read on still wet strip.

IX. INTERPRETING RESULTS



Positive



Negative

Invalid

The results are to be interpreted as follows:

Negative test result: a reddish-purple line appears across the central reading window at the Control line ("C") position. No other band is present.

Positive test result: in addition to a reddish-purple line at the Control line (C), a visible reddish-purple line appears at one of the Test lines position ("NDM" or "O23" or "O40/58") on cassette.

Intensity of the test line may vary according to the quantity of antigens as well as of the variant type present in the sample. Any reddish purple test line ("NDM", "O23" and "O40/58"), even weak, should be considered as a positive result.

If a positive test line appears beside of the "O40/58" mark, the sample contains OXA-40 or OXA-58 variants*. If it appears beside the "O23" mark, the sample contains OXA-23; beside the "NDM" mark, the sample contains NDM. Combinations of positive test lines can occur. In this case the sample contains several carbapenemases.

Invalid test result: The absence of a Control line indicates a failure in the test procedure. Repeat invalid tests with a new test device.

Note: during the drying process, a very faint shadow may appear at the Test line positions. It should not be regarded as a positive result.

The RESIST ACINETO test can generate positive results on the OXA-40/58 line for other related families of OXA in *A. baumannii* as well as non-*baumannii* *Acinetobacter* species, a.o. *A. pittii*. This has been documented for OXA-143 or related variants (OXA-255, OXA-499 and like) or OXA-213. This positive signal will depend on homology between these OXA and the OXA-40 or OXA-58 and on the expression level in the strain. It is important to identify *Acinetobacter* at the species level in parallel of carrying out the RESIST ACINETO test and to perform antibiogram. Confirming an *Acinetobacter* species is important since OXA-23 and OXA-58 may be (rarely) present in *P. mirabilis* as well as in *E. coli*, without any therapeutic involvement.

X. PERFORMANCE

A. Detection Limit

The detection limit determined with purified recombinant proteins of OXA-23, OXA-40, OXA-58 and NDM have been evaluated at 1.5 ng/ml, 0.099 ng/ml, 0.104 ng/ml and 0.144 ng/ml respectively.

B. Validation on collection of reference strains

The RESIST ACINETO test was evaluated on a collection of 297 clinical isolates with fully characterized resistance mechanisms to beta-lactams antibiotics by phenotypic and molecular tests (Germany).

OXA-23 status	Positive	Negative	Total
RESIST ACINETO			
Positive	189	0	189
Negative	2	106	108
Total	191	106	297

95 % Confidence Interval^H

Sensitivity:	99 %	(95.9 to 99.8 %)
Specificity:	100 %	(95.6 to 100 %)
Positive Predictive value:	100 %	(97.5 to 100 %)
Negative predictive value:	98.1 %	(92.8 to 99.7 %)
Agreement:	99.3 %	(295/297)

OXA-40/58 status	Positive	Negative	Total
RESIST ACINETO			
Positive	100	3	103
Negative	0	194	194
Total	100	197	297

95 % Confidence Interval^H

Sensitivity:	100 %	(95.4 to 100 %)
Specificity:	98.5 %	(95.3 to 99.6 %)
Positive Predictive value:	97.1 %	(91.1 to 99.2 %)
Negative predictive value:	100 %	(97.6 to 100 %)
Agreement:	99 %	(294/297)

NDM status	Positive	Negative	Total
RESIST ACINETO			
Positive	13	0	13
Negative	0	284	284
Total	13	284	297

95 % Confidence Interval^H

Sensitivity:	100 %	(71.7 to 100 %)
Specificity:	100 %	(98.3 to 100 %)
Positive Predictive value:	100 %	(71.7 to 100 %)
Negative predictive value:	100 %	(98.3 to 100 %)
Agreement:	100 %	(297/297)

The RESIST ACINETO test was also validated with rectal swabs and blood cultures.

C. Repeatability and reproducibility

To check intra-batch accuracy (repeatability), the same positive samples and a buffer solution were processed 15 times on kits of the same production batch in the same experimental conditions. All observed results were confirmed as expected.

To check inter-batch accuracy (reproducibility), some samples (positive and buffer) were processed on kits from three different production batches. All results were confirmed as expected.

D. Variants Detected by the RESIST test

A summary of variants reported in publications as detected is listed in an FAQ on the Coris BioConcept website <https://www.corisbio.com/faq/>. This list is not exhaustive of all the enzymatic variants that can be detected by the RESIST test.

XI. LIMITS OF THE KIT

The test is qualitative and cannot predict the quantity of antigens present in the sample. Clinical presentation and other test results must be taken into consideration to establish diagnosis.

A positive test does not rule out the possibility that other antibiotic resistance mechanisms may be present.

XII. TECHNICAL PROBLEMS / COMPLAINTS

If you face a technical problem or if performances do not correspond with those indicated in this package insert:

- Record the lot number of the kit concerned.
- If possible, keep the sample in the appropriate storage condition during the complaint management.
- Contact Coris BioConcept (client.care@corisbio.com) or your local distributor.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

XIII. BIBLIOGRAPHIC REFERENCES

- Alatraqchi AG, Mohd Rani F, A Rahman NI, Ismail S, Cleary DW, Clarke SC, Yeo CC.** Complete Genome Sequencing of *Acinetobacter baumannii* AC1633 and *Acinetobacter nosocomialis* AC1530 Unveils a Large Multidrug-Resistant Plasmid Encoding the NDM-1 and OXA-58 Carbapenemases. *mSphere*. 2021 Jan 27;6(1):e01076-20
- Chen Y, Guo P, Huang H, Huang Y, Wu Z, Liao K.** Detection of co-harboring OXA-58 and NDM-1 carbapenemase producing genes resided on a same plasmid from an *Acinetobacter pittii* clinical isolate in China. *Iran J Basic Med Sci*. 2019 Jan;22(1):106-111
- AIAMri AM, AIQurayan AM, Sebastian T, AINimr AM.** Molecular Surveillance of Multidrug-Resistant *Acinetobacter baumannii*. *Curr Microbiol*. 2020 Mar;77(3):335-342
- Kumar S, Patil PP, Singhal L, Ray P, Patil PB, Gautam V.** Molecular epidemiology of carbapenem-resistant *Acinetobacter baumannii* isolates reveals the emergence of blaOXA-23 and blaNDM-1 encoding international clones in India. *Infect Genet Evol*. 2019 Nov;75:103986
- Ogbolu DO, Alii OAT, Oluremi AS, Ogunjimi YT, Ojebode DI, Dada V, Alaka OO, Foster-Nyarko E, Webber MA.** Contribution of NDM and OXA-type carbapenemases to carbapenem resistance in clinical *Acinetobacter baumannii* from Nigeria. *Infect Dis (Lond)*. 2020 Sep;52(9):644-650
- Musila L, Kyany'a C, Maybank R, Stam J, Oundo V, Sang W.** Detection of diverse carbapenem and multidrug resistance genes and high-risk strain types among carbapenem non-susceptible clinical isolates of target gram-negative bacteria in Kenya. *PLoS One*. 2021 Feb 22;16(2):e0246937
- Monnheimer M, Cooper P, Amegbletor HK, Pello T, Groß U, Pfeifer Y, Schulze MH.** High Prevalence of Carbapenemase-Producing *Acinetobacter baumannii* in Wound Infections, Ghana, 2017/2018. *Microorganisms*. 2021 Mar 5;9(3):537.
- Newcombe, Robert G.** "Two-Sided Confidence Intervals for the Single Proportion: Comparison of Seven Methods," *Statistics in Medicine*, 17, 857-872 (1998).

Last update: 12 FEBRUARY 2024

REF	Catalogue number		Manufacturer
IVD	In vitro diagnostic medical device		Temperature limits
Σ	Contains sufficient for <n> tests	LOT	Batch code
	Consult instructions for use		Do not reuse
	Keep dry		Use by
DIL SPE	Diluent specimen	CONT Na ₂ S	Contains Sodium azide
UDI	Unique device identifier		