

## Certificate of Analysis

### General Product Information

Product Name: **Rb a Hu CD117, c-Kit**

Code Number: **A450229-2**

Storage data: **See Product label**

### Lot Specific Data

Lot Number: **11813687**

Expiration date: **2026FEB28**

### Specifications

Test	Specification	Result
Immunohistochemistry Specific Stain	Membrane and/or cytoplasmic staining of gastrointestinal stromal tumors	Pass
Immunohistochemistry Equivalent	Intensity of specific and background staining must be equivalent to corresponding control lot $\pm$ 0.5 grade	Pass
Immunohistochemistry Specific Stain	GIST specimen 2.0 - 4.0 membrane and/or cytoplasmic staining	Pass
Immunohistochemistry Background	$\leq 1.0$ background staining	Pass
Immunohistochemistry Specific Stain	Negative area/tissue must have no specific staining and non-specific staining $\leq 1.0$	Pass

QC approval	<u>Eric Smith</u>	Date	<u>2025MAR10</u>
QA approval	<u>Hector Cristobal</u>	Date	<u>2025MAR11</u>

# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	p120 Catenin (EP66), RMab, 0.1 mL Concentrate
<b>Catalog Number</b>	BSB 2080
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Synaptophysin (EP158), RMab, 0.5mL Concentrate
<b>Catalog Number</b>	BSB 2241
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	CDK4 (EP180), RMab, 0.1 mL Concentrate
<b>Catalog Number</b>	BSB 2464
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Tinto OLIG2 (EP112), RMab, 7mL Predilute
<b>Catalog Number</b>	BSB 2560
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe  
**Name** Estelle Anuwe  
**Department** Quality Assurance





**Diagnostic  
BioSystems**

**Certificate of Analysis**

**GMP-F-0081-F**  
**Release Date: 10/4/2022**

<b>Product Name:</b>	Elite PAP Pen
<b>Catalog number:</b>	K039
<b>Lot #</b>	2303-B
<b>Manufacturing Date:</b>	N/A
<b>Expiration Date:</b>	07/17/2026
<b>Solution Inspection:</b>	Pass
<b>Performance Inspection:</b>	Pass
<b>Visual Inspection:</b>	Pass
<b>Appearance:</b>	Pen with black cap
<b>Storage:</b>	Store in room temperature

**Created By:** Jacqueline Yonathan  
QC Technician

**Date:**

3/12/25

**Approved By:** Neena Gupta  
QA/RA Manager

**Date:**

3/12/2025

# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Tinto PAX-2 (EP235), RMab, 7mL Predilute
<b>Catalog Number</b>	BSB 2567
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	SOX-10 (EP268), RMab, 0.5mL Concentrate
<b>Catalog Number</b>	BSB 2584
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance





# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	GATA3 (L50-823), MMab, 0.5mL Concentrate
<b>Catalog Number</b>	BSB 2674
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	MDM2 (BSB-64), MMab, 0.5mL Concentrate
<b>Catalog Number</b>	BSB 2982
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Tinto MUC4 (8G 7), MMab, 7mL Predilute
<b>Catalog Number</b>	BSB 2986
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Tinto SALL4 (6E3), MMab, 15mL Predilute
<b>Catalog Number</b>	BSB 3185
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	SATB2 (EP281), RMAb, 0.5mL Concentrate
<b>Catalog Number</b>	BSB 3201
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Tinto ATRX (BSB-108), MMab, 7mL Predilute
<b>Catalog Number</b>	BSB 3294
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Tinto Napsin A (BSB-112), MMab, 15mL Predilute
<b>Catalog Number</b>	BSB 3394
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Tinto Mesothelial Cell (HBME-1), MMab, 15mL Predilute
<b>Catalog Number</b>	BSB 3457
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance





# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	INSM1 (BSB-123), MMab, 0.5mL Concentrate
<b>Catalog Number</b>	BSB 3557
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	p63 (4A4), MMab, 1 mL Concentrate
<b>Catalog Number</b>	BSB 3607
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Actin Smooth Muscle (BSB-15), MMab, 0.1 mL Concentrate
<b>Catalog Number</b>	BSB 5032
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe

**Name** Estelle Anuwe

**Department** Quality Assurance



# Certificate of Analysis

The undersigned at Bio SB, Inc., certifies that the following product was manufactured and tested according to good manufacturing practices and conforms to Bio SB's quality control procedures. Product components have been tested by immunohistochemistry using normal and/or cancer tissue microarray controls. Analysis showed acceptable sensitivity and specificity compared to the controls.

<b>Product Name</b>	Tinto Alpha-Fetoprotein (BSB-23), MMab, 7mL Predilute
<b>Catalog Number</b>	BSB 5051
<b>Storage</b>	2-8°C

<u>Lot Number</u>	<u>Expiry Date</u>	<u>Manufacturing Date</u>	<u>Quantity</u>
xxxxxxx			0

**Signature** Estelle Anuwe  
**Name** Estelle Anuwe  
**Department** Quality Assurance

