



Declaration of Conformity

Reference No. : LCS200515002AE

Applicant : Zhengzhou AiQURA Intelligent Technology Co., LTD.

Address : Floor 4, building 26, smart phone industrial park, intersection of xingang avenue and renmin road, zhengzhou airport economic comprehensive experimental zone

Trade Mark : AiQURA

Product : Finger Pulse Oximeter

Model : AD805

Standard : FCC Part 15, Subpart B, Class B(sDoC)

Important note:

Is herewith confirmed to comply with the requirements set out in the FCC Rules and Regulations Part 15 Subpart B and CISPR PUB. 22 and the measurement procedures were according to ANSI C63.4-2014. The said equipment in the configuration described in this report shows the maximum emission levels emanating from equipment are within the compliance requirements.

This device is in conformance with Part 15 of the FCC Rules and Regulations for radio frequency devices. Operation of this product is subject to the following two conditions:

- (1) this device may not cause harmful interference
- (2) this device must accept any interference received, including interference that may cause undesired operation.

The test was carried out on May 15, 2020 ~ May 18, 2020 at Shenzhen LCS Compliance Testing Laboratory Ltd.



Date of issue: May 18, 2020



Shenzhen LCS Compliance Testing Laboratory Ltd.

Room 101, 201, Building A and Room 301, Building C, Juji Industrial Park, Yabianxueziwei, Shajing Street, Bao' an District, Shenzhen, Guangdong, China

Tel: (86)755-82591330

Fax: (86)755-82591332

Http://www.lcs-cert.com

Email: webmaster@lcs-cert.com

Declarația de conformitate

Nr. de referință: LCS200515002AE

Aplicant: *Zhengzhou AiQURA Intelligent Technology Co., LTD.*

Marca: AiQURA

Produs: Pulsoximetru

Model: AD805

Standard: FCC Part 15, Subpart B, Class B (sDoc)

Notă importantă:

Se confirm, prin prezenta, că se respectă cerințele stabilite în Regulamentele și Reglementările FCC, Part 15 Subpart B and CISPR PUB. 22, iar procedurile de măsurare au fost conform ANSI C63.4-2014. În configurația descrisă în acest raport se arată că nivelul maxim de emisii ale echipamentului se află în normele solicitate.

Acest dispozitiv este în conformitate cu Partea 15 a FCC Regulamentele și Reglementările pentru dispozitivele cu frecvență radio. Funcționarea acestui produs este supusă următoarelor două condiții:

- (1) Acest utilaj nu poate crea interferențe dăunătoare
- (2) Acest dispozitiv trebuie să accepte orice interferență primită, inclusive interferențe care pot cauza operațiune nedorită.

Testele au fost efectuate în perioada 15 mai 2020 – 18 mai 2020 la Shenzhen LCS Compliance Testing Laboratory Ltd.

Data emiterii: 18 mai 2020



BCTC Building & 1-2F, East of B Building, Pengzhou Industrial Park,
Fuyuan 1st Road, Qiaotou, Fuyong Street, Bao'an District, Shenzhen,
Guangdong, China



Certificate of Compliance

Certificate Number: BCTC2006000917C

- Applicant** : Zhengzhou AiQURA Intelligent technology Co., Ltd.
4th floor, building 26, smart phone industrial park, intersection of
xingang avenue and renmin road, zhengzhou airport economic
comprehensive experimental zone
- Manufacturer** : Zhengzhou AiQURA Intelligent technology Co., Ltd.
4th floor, building 26, smart phone industrial park, intersection of
xingang avenue and renmin road, zhengzhou airport economic
comprehensive experimental zone
- Product** : **Finger pulse Oximeter**
- M/N** : **AD805**
- Test Standard** : **EN 60601-1-2:2015**

The EUT described above has been tested according to the listed standards and found in compliance with the council EMC directive 2014/30/EU. The manufacturer should consider this device meets requirement of MDD directive 93/42/EEC. The observations and test results referenced from this Certificate are relevant only to the sample tested. This Certificate is for the exclusive use of BCTC's Client and is provided pursuant to the agreement between BCTC and its Client. This Certificate is part of the full test report(s) and should be read in conjunction with report No. : BCTC2006000917E.



This certificate is for the exclusive use of BCTC's client and is provided pursuant to agreement between BCTC and its client. BCTC's responsibility and liability are limited to the terms and conditions of the agreement. The observation and test results referenced from this certificate are relevant only to the sample tested. This Certificate by itself does not imply that the material, product, or service is or has ever been a BCTC certification program.

Tel: 400-788-9558 or 0755-32936262
[Http://www.bctc-lab.com](http://www.bctc-lab.com)



Certificat de conformitate

Certificat Nr.: BCTC2006000917C

Prin prezentul certificate, Shenzhen Hongcai Testing Technology Co., Ltd. declară că testările au fost efectuate și raportul a fost generat pentru:

Aplicant: Zhengzhou AiQURA Intelligent Technology Co., LTD.

Producător: Zhengzhou AiQURA Intelligent Technology Co., LTD.

Produs: Pulsoximetru

Model: AD805

Standard de testare: EN 60601-1-2:2015

Dispozitivul descris mai sus a fost testat în conformitate cu standardele enumerate și a s-a stabilit că este în conformitate cu directiva 2014/30 / UE a Consiliului EMC. Producătorul ar trebui să ia în considerare faptul că acest dispozitiv îndeplinește cerințele directivei MDD 93/42 / CEE. Observațiile și rezultatele testelor la care se face referire în acest certificat sunt relevante numai pentru eșantionul testat. Acest certificat este pentru utilizarea exclusivă a clientului BCTC și este furnizat în conformitate cu acordul dintre BCTC și clientul său. Acest certificat face parte din raportul (rapoartele) complet (e) de testare și trebuie citit împreună cu raportul nr.: BCTC2006000917E.



Data emiterii: 19 iunie 2020

