

BIOBASE

JINAN BIOBASE BIOTECH CO., LTD

OLABO Intelligent Manufacturing Industrial Park, Ancheng Town, Pingyin County, Jinan
City, Shandong, 250409, China

Tel: +86-531-55697912; Fax: +86-531-81219804

E-mail: sales5@biobase.com Web: www.biobase.cc

Date: November, 13th, 2025

To Whom It May Concern

This letter is to confirm that a BKC-TL5VII centrifuge completed with angle rotor №3 (12*10 ml) and angle rotor № 4 (12*15 ml) fulfills the following requirements:

- laboratory centrifuge is used as an instrument for separating serum, plasma, precipitation of proteins, or detection of urinary sediment;
- has minimum rpm of 100;
- has minimum time range 0-99 min. Increment \leq 1 min;
- has automatic lid lock system: lid is locked and protected against opening while in operation;
- has visual and acoustic warnings (unbalance, start/stop, open lid);
- has unbalance detection function;
- has buttons for controlling on/off and lid opening;
- has digital display;
- has a port in the lid for operation checking;
- both angle rotor № 3 (12*10 ml) and angle rotor № 4 (12*15 ml) are compatible with the tubes of the following dimensions: 17 mm (diameter)*102 mm (length).

Signature

Name Clark Wang

Title General Manager

