

COMPLIANCE SHEET

To whom may concern:

Center for Centralized Public Procurement in Health

We MEDCAPTAIN MEDICAL TECHNOLOGY CO., LTD, 12th Floor, Baiwang Research Building, No.5158 Shahe West Road, Xili, Nanshan, 518055, Shenzhen, Guangdong, PEOPLE'S REPUBLIC OF CHINA

Do confirm, that the offered infusion pumps

HP-30 Pro

Offered in cooperation with Aelo-Grup SRL for the public procurement

**# 21040135 Lot 30
6 Units**



Complies with the following technical specification:

REQUIRED:

Infusion pump (infuser) with syringe (basic features)
Description A syringe infuser is a device that allows the intravenous administration of medical fluids with high accuracy.
Parameter Specification
General features Type With 1 channel
LCD display
the possibility of adjusting the brightness
Built-in handle yes
Display Data displayed Infusion rate yes
The infusion volume yes
Infusion time yes
Time to the end of the infusion yes
Syringe size yes
Occlusion level yes
Rate KVO da
Date, time is
Battery charge level yes
Events yes
Alarm is
Pump properties Flow rate range 0.1 - 999 ml / h
0.1 mL / h increment step
Bolus dosage Increase 0.1 - 1200 ml / h
Rate KVO 0.1 - 3 ml / h
Accuracy ≤2%
Compatible with all required syringe sizes produced by different manufacturers min. 8 producers, with the presentation of the list of producers
Possibility to add new types of syringes in the library yes
Accepted syringes 5, 10, 20, 30, 50, 60
Automatic syringe detection yes
Maximum occlusion pressure level 75 - 900 mmHg

OFFERED:

Infusion pump (infuser) with syringe (basic features)
Description A syringe infuser is a device that allows the intravenous administration of medical fluids with high accuracy.
Parameter Specification
General features Type With 1 channel
LCD display
the possibility of adjusting the brightness
Built-in handle yes
Display Data displayed Infusion rate yes
The infusion volume yes
Infusion time yes
Time to the end of the infusion yes
Syringe size yes
Occlusion level yes
Rate KVO da
Date, time is
Battery charge level yes
Events yes
Alarm is
Pump properties Flow rate range 0.1 - 9999.99 ml / h
0.01 mL / h increment step (Page10)
Bolus dosage Increase 0.1 - 2200 ml / h (Page10)
Rate KVO 0.1 - 5 ml / h (Page10)
Accuracy ≤2% (Page11)
Compatible with all required syringe sizes produced by different manufacturers min. 8 producers, with the presentation of the list of producers
Possibility to add new types of syringes in the library yes
Accepted syringes 2,3,5, 10, 20, 30, 50, 60 (Page8)
Automatic syringe detection yes
Maximum occlusion pressure level 50 - 975 mmHg (Page11)

<p>min. 5 levels</p> <p>Protection against liquid penetration into the device minimum IP24</p> <p>Infusion modes Infusion rate yes</p> <p>Dosage in time yes</p> <p>The ascending and descending mode yes</p> <p>Alarms and indicators Occlusion yes</p> <p>System error yes</p> <p>Empty syringe yes</p> <p>Infusion at last yes</p> <p>Discharged battery yes</p> <p>Audible alarm Adjustable volume Yes</p> <p>Silent mode button yes</p> <p>Logbook Stored events Buttons pressed yes</p> <p>Error codes yes</p> <p>Alarm is</p> <p>Medically injected by</p> <p>The amount infused yes</p> <p>Settings yes</p> <p>Bolus is requested</p> <p>Power supply 220 V, 50 Hz yes</p> <p>Internal Rechargeable Battery yes</p> <p>Operating time ≥ 5 h at 5 mL / h</p> <p>Accessories included separately for each pump purchased</p> <p>Support the system for fixing the pump to the infusion stand yes</p> <p>Syringes 50 ml 100 pcs.</p> <p>Syringe interconnect circuit 100 pcs.</p>	<p>min. 13 levels(Page11)</p> <p>Protection against liquid penetration into the device minimum IP24 Better,IP33(Page9)</p> <p>Infusion modes Infusion rate yes</p> <p>Dosage in time yes</p> <p>The ascending and descending mode yes</p> <p>Alarms and indicators Occlusion yes</p> <p>System error yes</p> <p>Empty syringe yes</p> <p>Infusion at last yes</p> <p>Discharged battery yes</p> <p>Audible alarm Adjustable volume Yes</p> <p>Silent mode button yes</p> <p>Logbook Stored events Buttons pressed yes</p> <p>Error codes yes</p> <p>Alarm is</p> <p>Medically injected by</p> <p>The amount infused yes</p> <p>Settings yes</p> <p>Bolus is requested</p> <p>Power supply 220 V, 50 Hz yes</p> <p>Internal Rechargeable Battery yes</p> <p>Operating time ≥ 10 h at 5 mL / h(Page9)</p> <p>Accessories included separately for each pump purchased</p> <p>Support the system for fixing the pump to the infusion stand yes</p> <p>Syringes 50 ml 100 pcs.</p> <p>Syringe interconnect circuit 100 pcs.</p>	
<p>Manufacturer's representative</p>	<p>stamp</p> <p></p>	<p>signature</p> <p></p>