

LETTER OF DECLARATION

To whom it may concern,

We, Shenzhen Mindray Bio-Medical Electronics Co., Ltd., ("Mindray") manufacturer of the Mindray Hematology Analyzers and Reagents, with registered business office at Mindray Building, Keji 12" Road South, High-tech Industrial Park, Nanshan, Shenzhen 518057, P.R. China, hereby declare that:

the reagent consumption for each model of analyzer listed below are as specified herein

Reagent Consumption for BC-20s / BC-20 / BC-30s / BC-30			
Nr	Reagent name	Consumption per 1 test (ml)	Comment
1	M-30CFL LYSE	0.31	at mode CBC+ DIFF
2	M-30D DILUENT	20	at mode CBC+ DIFF
3	PROBE CLEANSER	0	Consumption 1 ml during shutdown of analyzer
4	BC-3D Control	0	Shipped 4 times per year. Term of validity ~4 months.

Reagent Consumption for BC-3600			
Nr	Reagent name	Consumption per 1 test (ml)	Comment
1	M-30CFL LYSE	0.35	at mode CBC+ DIFF
2	M-30D DILUENT	27.66	at mode CBC+ DIFF
3	M-30R RINSE	7.5	at mode CBC+ DIFF
4	PROBE CLEANSER	0	Consumption 2.4 ml during shutdown of analyzer
5	BC-3D Control	0	Shipped 4 times per year. Term of validity ~4 months.

Reagent Consumption for BC-5000 / BC-5150			
Nr	Reagent name	Consumption per 1 test (ml)	Comment
1	M-52D Diluent	27.48	at mode CBC+ DIFF
2	M-52DIFF Lyse	1	at mode CBC+ DIFF
3	M-52LH Lyse	0.2	at mode CBC+ DIFF
4	PROBE CLEANSER	0	Consumption 2 ml during shutdown of analyzer

**SHENZHEN MINDRAY
BIO-MEDICAL ELECTRONICS CO., LTD.**

Mindray Building, Keji 12th Road South,
High-tech Industrial Park, Nanshan,
Shenzhen 518057, P.R. China

Tel: +86 755 81888998

Fax: +86 755 26582680

Website: www.mindray.com



5	BC-5D Control	0	Shipped 6 times per year. Term of validity ~2 months.
---	---------------	---	---

Reagent Consumption for BC-5300

Nr .	Reagent name	Consumption per 1 test (ml)	Comment
1	M-53D Diluent	42	Used at mode CBC+ DIFF
2	M-53 LEO (I) Lyse	1.7	at mode CBC+ DIFF
3	M-53 LEO (II) Lyse	0.14	at mode CBC+ DIFF
4	M-53 LH Lyse	0.5	at mode CBC+ DIFF
5	M-53 Cleanser	0	Consumption 30 during shutdown of analyzer
6	PROBE CLEANSER	0	Consumption 3.6 ml during shutdown of analyzer
7	BC-5D Control	0	Shipped 6 times per year. Term of validity ~2 months.

Reagent Consumption for BC-700

Nr .	Reagent name	Consumption per 1 test (ml)	Comment
1	DS Diluent	28	at mode CBC+DIFF
2	M-6 LD Lyse	1.62	at mode CBC+DIFF
3	M-6 LH Lyse	0.52	at mode CBC+DIFF
4	M-6 FD Dye	0.04	at mode CBC+DIFF
5	ESR Solution Reagent	2.08	Used only when ESR test is done,
6	PROBE CLEANSER	0	Consumption 4 ml during shutdown of analyzer
7	BC-6D Control	0	Shipped 6 times per year. Term of validity ~2 months.

Reagent Consumption for BC-720

Nr .	Reagent name	Consumption per 1 test (ml)	Comment
1	DS Diluent	28	at mode CBC+DIFF
2	M-6 LD Lyse	1.62	at mode CBC+DIFF
3	M-6 LH Lyse	0.52	at mode CBC+DIFF
4	M-6 FD Dye	0.04	at mode CBC+DIFF
5	ESR Solution Reagent	2.08	Used only when ESR test is done
6	M-6 DR DILUENT	1.62	Used only when RET test is done

**SHENZHEN MINDRAY
BIO-MEDICAL ELECTRONICS CO., LTD.**

Mindray Building, Keji 12th Road South,
High-tech Industrial Park, Nanshan,
Shenzhen 518057, P.R. China

Tel: +86 755 81888998

Fax: +86 755 26582680

Website: www.mindray.com

7	M-6 FR Dye	0.02	Used only when RET test is done
8	PROBE CLEANSER	0	Consumption 4 ml during shutdown of analyzer
9	BC-RET Control	0	Shipped 6 times per year. Term of validity ~2 months.
10	BC-6D Control	0	Shipped 6 times per year. Term of validity ~2 months.

Reagent Consumption for BC-760

Nr	Reagent name	Consumption per 1 test (ml)	Comment
1	DS Diluent	28	at mode CBC+DIFF
2	M-6 LD Lyse	1.62	at mode CBC+DIFF
3	M-6 LH Lyse	0.52	at mode CBC+DIFF
4	M-6 FD Dye	0.04	at mode CBC+DIFF
5	ESR Solution Reagent	2.08	Used only when ESR test is done
6	PROBE CLEANSER	0	Consumption 4ml during shutdown of analyzer
7	BC-6D Control	0	Shipped 6 times per year. Term of validity ~2 months.

Reagent Consumption for BC-780

Nr	Reagent name	Consumption per 1 test (ml)	Comment
1	DS Diluent	28	at mode CBC+DIFF
2	M-6 LD Lyse	1.62	at mode CBC+DIFF
3	M-6 LH Lyse	0.52	at mode CBC+DIFF
4	M-6 FD Dye	0.04	at mode CBC+DIFF
5	ESR Solution Reagent	2.08	Used only when ESR test is done
6	M-6 DR DILUENT	1.62	Used only when RET test is done
7	M-6 FR Dye	0.02	Used only when RET test is done
8	PROBE CLEANSER	0	Consumption 4 ml during shutdown of analyzer
9	BC-RET Control	0	Shipped 6 times per year. Term of validity ~2 months.
10	BC-6D Control	0	Shipped 6 times per year. Term of validity ~2 months.

Reagent Consumption for BC-7600

**SHENZHEN MINDRAY
BIO-MEDICAL ELECTRONICS CO., LTD.**

Mindray Building, Keji 12th Road South,
High-tech Industrial Park, Nanshan,
Shenzhen 518057, P.R. China

Tel: +86 755 81888998

Fax: +86 755 26582680

Website: www.mindray.com

Nr .	Reagent name	Consumption per 1 test (ml)	Comment
1	DS Diluent	40	at mode CBC+DIFF
2	M-6 LD Lyse	1.56	at mode CBC+DIFF
3	M-6 LH Lyse	0.52	at mode CBC+DIFF
4	M-6 FD Dye	0.02	at mode CBC+DIFF
5	ESR Solution Reagent	2.08	Used only when ESR test is done
8	PROBE CLEANSER	0	Consumption 4 ml during shutdown of analyzer
9	BC-RET Control	0	Shipped 6 times per year. Term of validity ~2 months.
10	BC-6D Control	0	Shipped 6 times per year. Term of validity ~2 months.

Reagent Consumption for BC-7800

Nr .	Reagent name	Consumption per 1 test (ml)	Comment
1	DS Diluent	40	at mode CBC+DIFF
2	M-6 LD Lyse	1.56	at mode CBC+DIFF
3	M-6 LH Lyse	0.52	at mode CBC+DIFF
4	M-6 FD Dye	0.02	at mode CBC+DIFF
5	ESR Solution Reagent	2.08	Used only when ESR test is done
6	M-6 DR DILUENT	1.56	Used only when RET test is done
7	M-6 FR Dye	0.02	Used only when RET test is done
8	PROBE CLEANSER	0	Consumption 4ml during shutdown of analyzer
9	BC-RET Control	0	Shipped 6 times per year. Term of validity ~2 months.
10	BC-6D Control	0	Shipped 6 times per year. Term of validity ~2 months.

Reagent Consumption for BC-7900

Nr .	Reagent name	Consumption per 1 test (ml)	Comment
1	DS Diluent	40	at mode CBC+DIFF
2	M-6 LD Lyse	1.56	at mode CBC+DIFF
3	M-6 LH Lyse	0.52	at mode CBC+DIFF
4	M-6 FD Dye	0.02	at mode CBC+DIFF
5	ESR Solution Reagent	2.08	Used only when ESR test is done

**SHENZHEN MINDRAY
BIO-MEDICAL ELECTRONICS CO., LTD.**

Mindray Building, Keji 12th Road South,
High-tech Industrial Park, Nanshan,
Shenzhen 518057, P.R. China

Tel: +86 755 81888998


Fax: +86 755 26582680

Website: www.mindray.com

8	PROBE CLEANSER	0	Consumption 4 ml during shutdown of analyzer
9	BC-RET Control	0	Shipped 6 times per year. Term of validity ~2 months.
10	BC-6D Control	0	Shipped 6 times per year. Term of validity ~2 months.

P.S. Please note that this test quantity applies to continuous testing. Consumption due to powering on, powering off, quality control testing and routine maintenance is not included.

Sincerely,



Gao Xiufu
Regional Marketing & Sales Manager, CA Region
Shenzhen Mindray Bio-Medical Electronics Co., Ltd.



**SHENZHEN MINDRAY
BIO-MEDICAL ELECTRONICS CO., LTD.**

Mindray Building, Keji 12th Road South,
High-tech Industrial Park, Nanshan,
Shenzhen 518057, P.R. China
Tel: +86 755 81888998
Fax: +86 755 26582680
Website: www.mindray.com