

# URIT 优利特

## CE Technical Document

Declaration of Conformity

(Sheath)

(Model:URIT 5S 11)

Document No: IVDR-URIT 5S 11-CE09

Version/Revision No: 1.1

Department Distributed: Department II

Prepared by:

Xiong Ling

Technical Engineer

Verified by:

Huangqinbin


Technical Manager



Approved by:

Su Qinghua

PRRC

**Declaration of Conformity**

Manufacturer:	URIT Medical Electronic Co., Ltd. Address: No. D-07 Information Industry District, High-Tech Zone, Guilin, Guangxi 541004, P. R. China
European Representative:	Shanghai International Holding Corp. GmbH (Europe) Address: Eifflstrasse 80, 20537 Hamburg, Germany
Product Name:	Sheath
Model:	URIT 5S 11
GMDN Code:	61165
EMDN Code:	W0103010105
SRN:	CN-MF-000011840
Basic UDI-DI:	69357404IBHS000006BB
Device Photograph:	
Intended Use:	It is used to dilute blood samples to form sheath flow, which is conducive to cell counting and classification by analytical instruments.
Risk Class:	Class A
Classification Rule:	Rule 5 (a) of Annex VIII of the Regulation (EU) 2017/746
Conformity Assessment Route	Article 48 section 10 of the Regulation (EU) 2017/746
We herewith declare that the above-mentioned product(s) meet the Regulation (EU) 2017/746 of THE EUROPEAN PARLIAMENT AND OF THE COUNCIL. All supporting documentation is retained at the premises of the manufacturer. We, the manufacturer, are exclusively responsible for the DoC.	
Applied Standards:	EN ISO 13485:2016 EN ISO 14971:2019 EN ISO 15223-1:2021

		EN ISO 18113-2: 2022 EN ISO 18113-1:2022 EN ISO 23640: 2015 EN 13612: 2002	
Notified Body:		NA	
Identification Number:		NA	
(EC) Certificate(s):	NA	Valid until	NA
Start of CE-marking:			
Place, Date of Issue:		Guilin, Guangxi 2023-12-8	
Signature:		  Name: Shi Ping Function: CEO	