



Declaration of Conformity



according to the In vitro diagnostic Medical Devices Directive 98/79/EC

Manufacturer: Changsha Sinocare Inc.
No.265 Guyuan Road, Hi-tech Zone, Changsha, Hunan Province,
Address: 410205, People's Republic of China.

We declare under our sole responsibility that

the Medical Device

Product Name : Blood Glucose monitoring system
Type/model, batch/serial number, possibly sources and number of items : Safe AQ Angel Blood Glucose Meter
Safe AQ Angel Blood Glucose Strip
Blood Glucose Control Solution
(Where applicable)

of class

according to directive 98/79/EC : The product classification is ListB.

meets all the provisions of the directive 98/79/EC which apply to it.

Applied harmonised standards, national standards or other normative documents

EN ISO 15197:2015

In vitro diagnostic test systems requirements for blood-Glucose monitoring systems for Self-testing in managing diabetes mellitus(ISO 15197:2013)

EN23640:2013

In vitro diagnostic medical devices-evaluation of stability of in vitro diagnostic reagents (ISO 23640:2011)

IEC 61326-1: 2012

Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

IEC-61326-2-6: 2012

Electrical requirements for measurement, control and laboratory use -EMC requirements- Part 2-6: Particular requirements -In vitro diagnostic (IVD) medical equipment

EN ISO14971:2012

Medical devices - Application of risk management to medical devices (ISO 14971:2007, Corrected version 2007-10-01)

IEC 61010-1: 2010

Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements

IEC 61010-2-101: 2015

Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment

EN ISO 18113-1:2011

In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 1: Terms, definitions and general requirements (ISO 18113-1:2009)

EN ISO 18113-4:2011

In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 4 : In vitro diagnostic reagents for self-testing (ISO18113-4:2009)

IEC 62366:2007(First Edition)+A1:2014

Medical devices-Application of usability engineering to medical devices

EN ISO 18113-5: 2011

In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 5: In vitro diagnostic instruments for self-testing (ISO 18113-5:2009)

IEC 62304 : :2015

Medical device software- Software life cycle processes

Conformity assessment procedure Council directive 98/79 /ECAnnex IV except setion 4、 6

Notified Body (if consulted) TÜV SÜD Product Service GmbH
Ridlerstraße 65 · 80339 München Germany NB ID 0123
Shanghai International Holding Corp.GmbH(Europe)

EC-Representative Eiffestraße 80,20537 Hamburg,Germany

No.265 Guyuan Road, Hi-tech
Zone,Changsha, Hunan Province,
410205, People's Republic of
China

Li Shaobo
General Manager

2/28/2017
(place and date)

Li Shaobo.
(name and signature, function)

