

DECLARATION OF CONFORMITY

We, the company

Meditü Kft. H-6900 Makó, Rákosi út 6.

In our capacity as manufacturer declare in own responsibility that the medical device of class 1 according to Annex IX of Directive 93/42/EEC

surgical needles, non-sterile – article see attachment

is designed and manufactured in agreement with the guideline 93/42/EEC. The essential requirements according to Annex 1 are fulfilled. The technical documentation is kept under the address above. The conformity evaluation was performed according to annex VII of the guideline 93/42/EEC, Applicable harmonized standards were used. This declaration is valid until 25.05.2024,

Makó, 30.11.2023.

MEDITÜ
Sebészeti Varrótű Gyártó Kft.
6900 Makó, Rákosi út 6.
Adószám: 10471773-2-06



Dr. Tóth István Tibor
managing director

Diameters and Lengths of Surgical Needles (mm)

Annex No. 1 of the Declaration of Conformity

Type	Size sign	Diameter	Length
A	000	1,80	100
B	00	1,60	92
G	0	1,60	85
AR	1	1,40	80
BR	2	1,40	75
GR	3	1,30	70
BT	4	1,30	65
GT	5	1,20	60
	6	1,20	55
	7	1,10	50
	8	1,00	45
	9	1,00	40
	10	0,90	36
	11	0,80	32
	12	0,80	28
	13	0,70	25
	14	0,70	22
	15	0,60	20
	16	0,60	18
GA	000	2,00	90
GAR	00	1,80	80
GAT	0	1,80	70
	1	1,60	60
	2	1,60	55
	3	1,50	50
	4	1,40	45
	5	1,30	40
	6	1,20	35
	7	1,10	30
	8	1,00	25
	9	0,90	22
	10	0,80	19
	11	0,80	16
E	000	1,30	70
ET	00	1,20	60
	0	1,20	55
	1	1,00	47
	2	1,00	40
	3	0,80	32
	4	0,80	28
	5	0,70	22
	6	0,60	18
	7	0,50	15
	8	0,50	12
OB	0	0,60	36
OG	1	0,60	30
OA	2	0,50	25
	3	0,50	21
	4	0,50	18
	5	0,50	15
	6	0,50	12
H	00	1,40	75
HR	0	1,20	60
	1	1,10	52
	2	1,00	45
	3	0,90	38
	4	0,80	30
	5	0,70	22

Type	Size sign	Diameter	Length
PB	00	0,80	75
PC	0	0,80	65
PD	1	0,70	55
PH	2	0,70	45
PA	3	0,60	36
	4	0,60	30
	5	0,60	25
	6	0,60	22
	7	0,50	20
	8	0,50	18
	9	0,50	15
	10	0,50	12
AV	1	2,00	100
BV	2	2,00	90
	3	1,80	80
	4	1,80	70
	5	1,60	60
	6	1,60	55
	7	1,50	50
	8	1,40	45
	9	1,30	40
	10	1,20	35
	11	1,10	30
	12	1,00	25
	13	0,90	22
BE	1	1,40	80
GE	2	1,40	75
BER	3	1,30	70
	4	1,30	65
	5	1,20	60
	6	1,20	55
	7	1,10	52
	8	1,10	50
	9	1,00	47
	10	1,00	45
	11	1,00	40
	12	0,90	38
	13	0,90	36
	14	0,80	34
	15	0,80	32
	16	0,80	30
	17	0,80	28
	18	0,70	25
	19	0,70	23
	20	0,70	22
	21	0,60	20
	22	0,60	18

Type	Size sign	Diameter	Length
Keith	0	1,00	80
Keith-R	1	0,90	75
	2	0,90	70
	3	0,90	65
	4	0,80	60
	5	0,80	55
	6	0,80	50
	7	0,70	45
König	1	1,20	55
	2	1,10	50
	3	1,00	45
	4	1,00	40
	5	0,90	36
	6	0,80	34
	7	0,80	32
	8	0,80	30
	9	0,80	26
	10	0,70	23
	11	0,60	20
	12	0,60	18
Neumann	1	0,70	23
	2	0,60	18
Kalt	00	0,35	25
	0	0,35	22
	1	0,35	20
	2	0,35	16
	3	0,35	12
	4	0,35	10
Mayo	1	1,50	50
Mayo T	2	1,40	45
	3	1,40	42
	4	1,30	40
PM/S	1	2,00	145
PM/A	2	2,00	130
	3	2,00	115
	4	2,00	100
	5	1,80	85
	6	1,80	75
Payr G	00	0,35	25
	0	0,35	22
	1	0,35	18
	2	0,35	16
	3	0,35	12
Payr C	1	0,35	30
	2	0,35	25
	3	0,35	22
	4	0,35	20
	5	0,35	16
	6	0,35	12
Ferguson	1	1,40	74
	2	1,30	63
	3	1,10	50
	4	1,00	40
	5	0,80	34
	6	0,70	23