

## CERTIFICATE OF ANALYSIS

Certificado de Análisis / Certificat de contrôle / Chargen-Prüfprotokoll

Pourquier\* Brucellosis Negative Control [Brucellosis Negative Control]

---

Product code :	P00131	Expiration date :	24-AUG-2024
Batch :	442		
Tests per kit :	/		
Insert version :	/	Manufacturing date :	AUG-2019

---

Reconstitute\* with 1ml distilled water

\* After reconstitution, the control serum must be stored at a temperature  $\leq -16^{\circ}\text{C}$  for up to 1 year

He can be frozen and unfrozen up to three times

---

Control Test	Spécifications	Results
EAT :	Negative : No agglutination	Pass
SAW :	Negative( < 30 UI/ml)	Pass
FC :	Negative( < 20 ICFTU/ml)	Pass
ELISA :	Negative	Pass

---


This product was performance tested and has met all quality control specifications required for release.

This information is released by :

Title :

Sélectionner

Signature :



Manufacturer: IDEXX Montpellier SAS - 326, rue de la Galéra – Parc Euromédecine - 34090 Montpellier - France

Tél. : 33(0) 4 99 23 24 25 - Fax : 33(0) 4 67 04 20 25 - [info-montpellier@idexx.com](mailto:info-montpellier@idexx.com)

EN 05-01-17 - Vers : C

Document printed on : 18-Sep-2019