



Awareness Technology, Inc.

Declaration of Conformity

Product identification

Product name : Incubator / Shaker
Model/Type : P2200 Series

Manufacturer

Name : Awareness Technology, Inc.
Address : PO Drawer 1679
Palm City, Florida
Country : USA

Authorized Representative in Europe

Representative:
Name : Emergo Europe
Address : Molenstraat 15
2513 BH The Hague
Country : The Netherlands
tel: +31 70 345 8570
fax: +31 70 346 7299

Means of Conformity

Awareness Technology, Inc. declares that the product listed is in conformity with the Annex III, essential requirements and provisions of Council Directive:

98/79/EC

And is in conformance with the following standards:

EN 61326-1 / EN 55022
EN 61010-1

Signature

Place and Date: Awareness Technology, Inc.

December 01,2003

Signature :





Name :

Steve Andrus
Quality Manager

Chris Mauer
Compliance Engineer



Stat Fax[®] 2200 - Incubator Shaker

Features

Holds two 96 well microplates.

Digitally controlled 8-speed shaker with orbital mixing from 575 to 1500 rpm.

Temperature adjustable from ambient to 40°C with resolution of 0.1°C.

Digital timer controls shaking or can be operated independently, solely as a timer.

Compact and easy to operate.

Outer tinted cover provides further insulation and protection from light.

Options & Accessories

Available with either 120V AC or 230V AC adapter.

37°C temperature calibration plate.

Distributed by:



Awareness Technology, Inc. P.O. Box 1679, Palm City, FL 34991 (772) 283 6540 fax (772) 283 8020
web site: <http://www.awaretech.com>, email: info@awaretech.com

Cost Effective by Design[®]

Stat Fax[®] 2200

Incubator / Shaker specifications

Incubator

Temperature Range: Ambient to 40°C.
Resolution: 0.1°C
Uniformity with Cover: Better than +/- 0.4°C.
Mechanism: Enclosed, warm aluminum bed with air circulation.

Mixer

Speeds: 8 gradual steps from 575 to 1500 RPM.
Options: Continuous or pulse.
Mechanism: Rotating counter weight.

Timer

Modes: Set time, view time remaining, audible signal (stops mixing but maintains temperature control after time out).
Resolution: 1 second.
Range: 1 second to 99 minutes, 59 seconds.

Electronic

Display: Four 7-segment red LEDs.
Keyboard: 16 key, domed membrane switch, enunciating.
Power Requirements: 12V AC, 2 amps. Adapters available for 120V or 230V power with US or Euro plugs.
Microprocessor: Z80A or Z180.
Certifications: NRTL Listed, CE Mark.

Other

Vessel: Holds 2 standard microtiter plates or strip trays.
Enclosure: Fire-retardant ABS plastic enclosure with metal base.
Dimensions: Approx. 11 x 12.5 x 5.75 in, 7.2 lbs.
Approx. 43 x 37 x 18cm, 3.3kg.

Ask about these other products
from Awareness Technology:



Stat Fax[®] 2600
Microplate Washer



Dri-Dye[®]
Check Strips



Lumi Stat[®]
Microwell Strip Luminometer

CERTIFICATE OF REGISTRATION

This is to certify that the management system of:

Awareness Technology, Inc.

Main Site: 1935 SW Martin Highway

Palm City, Florida 34990 USA

Additional site: 2325 SW Martin Highway, Palm City, Florida 34990 USA

has been assessed by Intertek as conforming to the requirements of:

ISO 13485:2016

The quality management system is applicable to:

The design, development, manufacture, distribution, installation and service of IVDD General Laboratory Instruments.

Additional site: Manufacturing, Quality Control, Distribution, Shipping, Installation and Service.

Certificate Number:

9362-8

Initial Certification Date:

March 28, 2012

Date of Certification Decision:

March 24, 2021

Issuing Date:

March 27, 2021

Valid Until:

March 27, 2024



Intertek



A handwritten signature in black ink, appearing to read "Calin Moldovean".

Calin Moldovean
President

Intertek Testing Services NA Ltd.,
1829, 32nd avenue, Lachine, QC, H8T 3J1,
Canada

