

1. PRODUCT

STANDARD SURGICAL GOWN

2. PRODUCT CONFIRMITIES

ISO 9001, ISO 13485, EN 13795, ISO 11135:2014, ISO 14644-3:2019, ISO 14791:2012, ISO 14937: 2009, ISO 10993-7:2008, EN 20811

3. DESCRIPTION

Nonwoven , 3 layers SMS 35 gsm, Resistance to penetration of liquids within 50-80 cm H2O according to EN 20811
Velcro adjustment at the Neck, overlapping closure at the back by ties , Elastic Cuff Knitted
Antiallergenic, AirPermeable, Non-Linting
Water Repellent , latex free, full ultrasonic seems(including armpits)

4. PACKING

The product is individually packaged. On the packaging is the inscription with the name and address of the manufacturer, the product model, description, dimensions, date of manufacture, validity date, lot number, etc.

5. STORAGE

0°C - + 30°C, Avoid Sunlight, Avoid Humidity

6. Validity

Sterile EO. 3 years from the date of production

Size	Length (cm)
S (SMALL)	min.120
M (MEDIUM)	min.125
L (LARGE)	min.135
XL (XLARGE)	min.140
2XL (XXLARGE)	min.150



ZEYNI MEDİKAL
TEKSTİL İNŞAAT SANAYİ TİC. LTD. ŞTİ.
Yatırımcı: Akh. 12. Katman Eki. No.17
Yatırımcı: USA
Yatırım V.D. 2007/11/11 No. Sic.No 102911
Mersis No: 0801001410000001

Ultrasonic Welding Seams



Ultrasonic welding is operated by machine, easy to control quality with high output. Make sure good quality. Avoid the problem of stitches coming loose, more firm and durable; Good ability of leak-proof and permeation-resistance;