

SRI TRANG GLOVES (THAILAND) PUBLIC COMPANY LIMITED

10 Soi 10 Phetkasem Road, Hat Yai, Hat Yai, Songkhla 90110 Thailand
Tel. (66) 074-344-663 Fax. (66) 074-344-677

SCT.PD.SP.05.237-100322 R.05

PRODUCT SPECIFICATION SHEET : LXXOGF-B-EU-M-NS

DAR-SCT-2022-0410

PAGE 1 of 1

GLOVE SPECIFICATION

Product Specification Group Code : LXXOGF-B-EU-M-NS
Product Code : DLXFBOG
Product name : Latex Examination Gloves, Powdered, Non Sterile
Grade : EU Spec, Medical Grade

GENERAL INFORMATION

Type : Single-use disposable examination or protective glove
Marking : Gloves are not marked to designated size.
Shape : Ambidextrous
Material : Natural Rubber
Warning: This product contains natural rubber latex which may cause allergic reactions
Cuff : Rolled cuff
Surface : Textured at Finger Tip

Thickness : (single wall according to ASTM D3578, ASTM D3767)	Finger : Min. 0.08mm	Palm : Min. 0.08mm
---	-----------------------------	---------------------------

Powder amount : (according to EN 455-3, EN ISO 21171)	Max. 10mg/dm ²	Force at break : (according to EN 455-2)	Before Aging : Med. ≥ 6.0N After Aging : Med. ≥ 6.0N
---	---------------------------	--	---

Protein content : (according to EN 455-3)	Max. 150µg/g	Tensile Strength : (according to ASTM D3578, ASTM D412, ASTM D573)	Before Aging : Min. 18Mpa After Aging : Min. 14Mpa
---	--------------	--	---

Length : (according to EN 455-2)	XS-XL : Med. ≥ 240mm	% Elongation : (according to ASTM D3578, ASTM D412, ASTM D573)	Before Aging : Min. 650% After Aging : Min. 500%
--	----------------------	--	---

Available Sizes & Palm width (mm) : XS (5-6) : ≤ 80, S (6-7) : 80+/-10, M (7-8) : 95 +/-10, L (8-9) : 110+/-10, XL (9-10) : ≥ 110
(according to EN 455-2)

Storage: 5 years storage life (date of expiration see on the box). Keep area cool, dry and dust free, avoid ventilation and storage close to photocopy equipment. Copper-ions discolor the glove. Protect gloves against ultraviolet light sources, as sunlight and oxidizing agents. Storage above 30°C will lead to accelerated aging and should be avoided.

STANDARDS, GUIDELINES AND QUALITY CERTIFICATION

(Every mentioned standard is used in the latest edition.)

Certification :	ISO 9001, ISO 13485		
Conformity to guidelines :	Medical Device Regulation (EU)2017/745 : Class I under Rule1 and Rule 5 according to Annex VIII (Subject to EU Declaration of Conformity to brand owner) Personal Protective Equipment Regulation (EU) 2016/425 : Category I or Category III (Subject to EU Type-Examination Certificate)		
Conformity to standards :	<u>Medical</u> EN 455 ASTM D3578	<u>PPE</u> CAT I EN ISO 21420	<u>PPE</u> CAT III EN ISO 21420 EN ISO 374 EN 16523-1
Conformity to food regulation :	1935/2004		

WARNING AND REPORTING SYSTEM

Potentially allergenic ingredients : Dithiocarbamate type.
(For further information, the actual version of the ingredients list is given on request.)

Please report any incidents with the products to STGT_TA@sritranggroup.com and regulatory@sritranggroup.com without undue delay. Our Product Safety Officers will review all information provided and might contact the original informant for further clarification.

SRI TRANG GLOVES (THAILAND) PUBLIC COMPANY LIMITED

10 Soi 10 Phetkasem Road, Hat Yai, Hat Yai, Songkhla 90110 Thailand
Tel. (66) 074-344-663 Fax. (66) 074-344-677

SCT.PD.SP.05.237-100322 R.05

PRODUCT SPECIFICATION SHEET : LXXOGF-B-EU-M-NS

DAR-SCT-2022-0410

Appendix of specification

APPENDIX : 2

Product Specification Code : LXXOGF-B-WHT-EU-MC-NS
Product Code : DLXFBOG
Material : Natural Rubber
Treatment and Feature : Corn Starch Powder (Absorbable Dusting Powder-USP Grade)

SAMPLING, INSPECTION AND FINAL RELEASE INFORMATION

Product Specification Code	Glove color	Characteristics			Test-Method
		Barrier Defects*	Major Visual Defects*	Minor Defects*	
LXXOGF-B-WHT-EU-MC-NS	WHITE	AQL ≤ 1.5	AQL ≤ 1.5	AQL ≤ 4.0	EN 455-1

Barrier Defects :

Inspection level G I (ISO 2859-1)

Holes that results in the malfunction of glove as a protection barrier either physically or not physically affect the appearance.

Major Visual Defects :

Inspection level G I (ISO 2859-1)

Defined defect that disrupt the glove's function or defect that cause an uncomfortable feeling to the user.

Minor Defects :

Inspection level G I (ISO 2859-1)

Visual defects aggregated are non-conformities of lower degree of concern, which do not prevent correct use of gloves. Minor cosmetic defect which does not affect on the glove's performance

Remark :

* Defect Characteristics following In-house practice.

Product description / Details of packaging for Change Control No.:				OTC-STA-0821-1734	
1. Brand Name:		SRITRANG		Customer : IMUNOTEHNOMED(MOLDOVA)	
Brand Name (INT):		STGEUVRH-10X100-LXFBOG-EUC-WHT-NS-BX			
2. Product Type:		LATEX POWDERED WHITE TEXTURED AT FINGER TIPS NON-STERILE			
3. KMAT :		DLXFBOG		510K No.:	-
Specification:		LXXOGF-B-WHT-EU-MC-NS		CA No.:	-
4. Material number (FG):					
Size XS :		14018223			
Size S :		14018224			
Size M :		14018225			
Size L :		14018226			
Size XL :		14018227			
Size 2L :		-			
5. Packaging					
Packaging type :		INNER BOX			
Dimension(mm):		PKGPD03	W: 110	L: 200	H: 60
SIZE	COLOR	MATERIAL NO.(PCK)	VERSION		
FS	BLACK,2577C	20017865	C01		
XS					
S					
M					
L					
XL					
2L					
Packaging type :		OUTER BOX			
Dimension(mm):		PKGPD03	W: 225	L: 310	H: 210
SIZE	COLOR	MATERIAL NO.(PCK)	VERSION		
FS	BLACK,2577U	20017866	C01		
XS					
S					
M					
L					
XL					
2L					
Packaging type :		-			
Dimension(mm):		-	W: -	L: -	H: -
SIZE	COLOR	MATERIAL NO.(PCK)	VERSION		
FS	-	-	-		
XS	-	-	-		
S	-	-	-		
M	-	-	-		
L	-	-	-		
XL	-	-	-		
2L	-	-	-		
Packaging type :		-			
Dimension(mm):		-	W: -	L: -	H: -
SIZE	COLOR	MATERIAL NO.(PCK)	VERSION		
FS	-	-	-		
XS	-	-	-		
S	-	-	-		
M	-	-	-		
L	-	-	-		
XL	-	-	-		
2L	-	-	-		
Packaging type :		-			
Dimension(mm):		-	W: -	L: -	H: -
SIZE	COLOR	MATERIAL NO.(PCK)	VERSION		
FS	-	-	-		
XS	-	-	-		
S	-	-	-		
M	-	-	-		
L	-	-	-		
XL	-	-	-		
2L	-	-	-		
Material:		DUPLEX		FSC: NO	
Flap:		LOCKING		Coating: VARNISHING	
Opening:		OTHER AND NON STD-SINGLE LINE			
Marking Size by:		TICK SIZE BY BLACK			
MFG+Lot No.:		YYYY-MM(MFG) XXXX(INT) CCCC(CASE) DLXFBOG(SPECIAL)			
MFG:		-			
EXP :		YYYY-MM(MFG+5YRS)			
Pre-print:		NO			
Remark:		Shared artwork with STGEUVRH-10X100-LXFBOG-EUC-WHT-NS-BX Inkjet DLXFBOG behind (CASE NO.)			
Material:		B.BROWN		FSC: NO	
Flap:		-		Coating: -	
Opening:		-			
Marking Size by:		STAMP TEXT BY BLACK			
MFG+Lot No.:		YYYY-MM(MFG) XXXX(INT) CCCC(CASE).REF.DLXFBOG(SPECIAL)			
MFG:		-			
EXP :		YYYY-MM(MFG+5YRS)			
Pre-print:		NO			
Remark:		Shared artwork with STGEUVRH-10X100-LXFBOG-EUC-WHT-NS-BX Inkjet. REF.DLXFBOG short side 2			
Material:		-		FSC: -	
Flap:		-		Coating: -	
Opening:		-			
Marking Size by:		-			
MFG+Lot No.:		-			
MFG:		-			
EXP :		-			
Pre-print:		-			
Remark:		-			
Material:		-		FSC: -	
Flap:		-		Coating: -	
Opening:		-			
Marking Size by:		-			
MFG+Lot No.:		-			
MFG:		-			
EXP :		-			
Pre-print:		-			
Remark:		-			

6. Packing/Contain instructions.

Dispenser
 Carton
 Packing Method

☐ Use ☒ Not Use Plastic under the perforation of dispenser boxes
☐ Use ☒ Not Use Plastic bag for dispenser boxes - -
☐ Use ☒ Not Use Poly bag for carton Color: -

7. Loading quantity

Container 20ft : Cartons / Container 40ft : Cartons / Container 40ftHC : Cartons / Truck : Cartons

8. Change description

For Dispenser box :




-Change format inkjet
 -Shared artwork with STGEUVRH-10X100-LXFBG-EUC-WHT-NS-BX

For Transport carton :

-Change format inkjet
 -Shared artwork with STGEUVRH-10X100-LXFBG-EUC-WHT-NS-BX

9. Packing format and Inkjet format

Special format : (Packing Format : Outer / Special Ink-Jet)



INKJET BY STGT INTERNAL DEPT.



St.B.