

ATTESTATION OF CONFORMITY

Issued to: Hangzhou Bigfish Bio-tech Co., Ltd
Room 8310, Building No.2, No.22, Longquan Road, Yuhang District,
Hangzhou P.R. China

For the product: Real-time fluorescent quantitative PCR analyzer

Trade name: BIGFISH

Type/Model: BFQP-16, BFQP-48, BFQP-96

Ratings: Analyzer : 24 Vdc, 180 W
Adaptor: GST220A24
Input: 100-240 Vac, 50/60 Hz, 4,0 A; Output: 24 Vdc, 9,2 A, 221 W MAX.

Manufactured by: Hangzhou Bigfish Bio-tech Co., Ltd
Room 8310, Building No.2, No.22, Longquan Road, Yuhang District,
Hangzhou P.R. China

Requirements: EN 61326-1:2013

This Attestation is granted on account of an examination by DEKRA, the results of which are laid down in a test report no 6086193.50

This Attestation implies that the examined types are in accordance with the standards designated under the Electromagnetic compatibility directive 2014/30/EU.

The examination has been carried out on one single specimen or several specimens of the product, submitted by the manufacturer. The Attestation does not include an assessment of the manufacturer's production. Conformity of his production with the specimen tested by DEKRA is not the responsibility of DEKRA.

The CE marking may be affixed on the product if all relevant and effective EC directives are complied with.

Shanghai, 21 September 2020

Number: 6086193.01AOC V1.1

DEKRA Testing and Certification (Shanghai) Ltd.

Kreny Lin
Certification Manager

© Integral publication of this attestation and adjoining reports is allowed

