

# Test Report

Number: GZHT02508749

<b>Report Ref:</b>	GZHT02508749		
<b>Date received/ Test Started:</b>	Dec 17, 2021	<b>Date Issued:</b>	Dec 24, 2021

<b>Company Name:</b>	NEOMATRIX SRL
<b>Address:</b>	# 121, 31 AUGUST 1989 STR MD-2012 CHISINAU, REPUBLIC OF MOLDOVA
<b>Contact Name:</b>	Sergiu Sirbu

The Following Sample Was Submitted And Identified By/On Behalf Of The Applicant As:	
End Uses	: Non-Sterile Medical Face Mask
Ratings	: -
Sample Name	: Matrix Medical Disposable Face Mask Type II R
No. Of Sample	: One (20 Pieces)
Size	: 175x97mm
Colour	: Blue/White
Manufacturer	: NeoMATRIX SRL
Buyer	: NeoMATRIX SRL
Brand Name	: MATRIX
Agent	: Intertek Deutschland GmbH
Standard	: -
Date received/ Test Started	: Dec 17, 2021
Ref	: Type No.: Type IIR Lot No.: 4840782288991/10-21 Product Registration Number: AMDM Nr.DM000274414

Test was conducted on specific items, at our client's request.

Prepared And Checked By:  
For Intertek Testing Services Shenzhen Ltd.Guangzhou Branch



James Ma  
Textile lab Senior Manager

Ula / hilaryxu



Page 1 Of 3

**Original Sample Photo**



Prepared And Checked By:  
For Intertek Testing Services Shenzhen Ltd.Guangzhou Branch

James Ma  
Textile lab Senior Manager

Ula / hilaryxu



**Test Report**

Number: GZHT02508749

Tests Conducted (As Requested By The Applicant)

1 Tensile Strength As Received (ISO 9073-18:2008, 1-Inch Grab Test, Tensile Testing Machine, CRE):

Length	72.1 N
Width	44.5 N

Remark : N = Newton

---

End of Report

*This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Intertek does not accept a duty of care or any other responsibility to any person other than the Client in respect of this report and only accepts liability to the Client insofar as is expressly contained in the terms and conditions governing Intertek's provision of services to you. Intertek makes no warranties or representations either express or implied with respect to this report save as provided for in those terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, except in the event of our gross negligence or wilful misconduct. No copy of the test report(except for full text copy) shall be made without the written approval by Intertek.*

Ula / hilaryxu

Page 3 Of 3