

# Certificate of Analysis



## OPTICAL ADHESIVE COVERS

Product No. 4311971  
Lot No. 33YJ88  
Date of Manufacture 05AUG2024  
Expiration Date 05AUG2026

| TEST                                   | SPECIFICATION        | RESULT |
|----------------------------------------|----------------------|--------|
| <b>Adhesive Testing</b>                |                      |        |
| ADHESION TO STEEL                      | ≥ 30 oz/in           | Pass   |
| <b>Physical Inspection</b>             |                      |        |
| DIMENSION D PER SOP                    | 0.0937 - 0.1562 inch | Pass   |
| TAPE WIDTH PER SOP                     | 3.03 - 3.09 inch     | Pass   |
| TAPE LENGTH BETWEEN PERF LINES PER SOP | 4.53 - 4.59 inch     | Pass   |
| TAPE LENGTH PER SOP                    | 5.53 - 5.59 inch     | Pass   |

***For Research Use Only. Not for use in diagnostic procedures.***

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system.

Austin, TX, USA

Life Technologies  
5791 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@lifetech.com](mailto:cofarequests@lifetech.com).

Quality Assurance  
Issued 13MAY2025

FORM-00013E  
Doc: 4370328 G