



Surgical Power Tools







www.asco-starlite.com

R

Τ

S

Μ





369D.302

369D.305

CANNULATED DRILL MACHINE

It is used for various types of drilling operations. Also Suitable for K-wire surgery & Reaming Operations*

Standard Configuration:

- ★ One Hand Piece
- ★2 pcs. Battery
- ★1 pc. chuck & key
- ★ 1 pc. Charger ★ 2 pcs. Sterilizing Rings

Speed	1000 rpm
Noise	≪65db
Sterilize	≤135 °C
Torque	≥ 3.5 NM

Autoclavable

Packed in Aluminium Carrying Case



OSCILLATING SAW

Standard Configuration:

★One Hand Piece ★2 pcs. Saw Blades ★ 1 pc. Charger

★ 2 pcs. Sterilizing Rings

★2 pcs. Battery

SPECIFICATIONS

Speed	0-20000 osc/m
Noise	≤70db
Sterilize	≤135 °C

Autoclavable

Packed in Aluminium Carrying Case



International **Blade Design**

THE STARLITE RANGE OF ORTHOPAEDIC POWER TOOLS IS SIMPLY S M



- + Battery Operated





369D.304

Battery Operated

369D.300UA

STERNUM SAW

For application in thoracic surgery

Set Contents: ★One Hand Piece ★2 pcs. Saw Blades ★2 pcs. Battery

* 1 pc. Charger ★ 2 pcs. Sterilizing Rings

SPECIFICATIONS

Speed	0-15,000 osc/m
Noise	≤75db
Steriliz	≤135 °C

Autoclavable

Packed in Aluminium Carrying Case

Series 3 Ultra

MULTIFUNCTION ORTHO DRILL & SAW SET

★ 1 pc. Charger

★ 2 pcs. Battery

★ 2 pcs. Sterilizing Rings

★ 2 pcs. Oscillating Saw Blades

Set Contents:

- * Hand Piece Motor, Battery Operated
- ★ Drill Attachment (Jacob Chuck Type)
- * Cannualted Drill Attachment
- ★ Oscillating Saw attachment
- ★ Reamer Attachment
- *Convert Adapter for Quick Coupling Drill Bits
- * Convert Adapter for AO design Reamers

Autoclavable

Packed in Aluminium Carrying Case

Technical Specifications:

Hand	Working Voltage	14.4V
Hand Piece	Output Power	≥90W
	Unit Rise in Temp.	≤20°C
	Input Voltage	AC 110V-240V 50HZ-60HZ
Charger	Output Power	14.4V 900MA
	Input Power	30W
	Voltage	14.4V
Battery	Charging Time	2Hours or more
	Runnimg Time	Continues 45-50 minutes on full charge



Better Battery life

Faster Speed

High Power

Better Grip

Ergonomical Shape





CRANIAL DRILL & BUR

For application in cranial surgery.

- Standard Configuration:
- ★ 1 pc. Cranial Drill Attachment
 ★ 1 pc. Bur Attachment
- *2 pcs. Battery

SPECIFICATIONS

★1 pc. Charger

* 2 pcs. Sterilizing Rings

SFECIFICATIONS		
Speed - Drill	0-1000 r/m	
Speed - Bur	0-40,000 r/m	
Noise	≪65db	
Sterilize	≤135 °C	

Autoclavable

Packed in Aluminium Carrying Case







Apothecaries Sundries Manufacturing Company is a leading provider of quality medical & hospital equipment to doctors, hospitals, and health service agencies worldwide. The company, simply known as ASCO in the global market, manufactures and markets a wide variety of high quality hospital equipment. The brand has gained recognition worldwide as a high quality premier brand has made its mark into the European marketplace where it continues to be recognized for its exceptional quality. ASCO, has been in the field of supplying quality surgical/hospital appliances and allied products for almost six decades now. It is also credited with setting the benchmark of excellence in quality, R&D and customer service.

if you are looking for power tools that offer reliable performance in the most demanding applications, that come with a variety of options to suit your needs and are available in reasonable prices, **Starlite is the answer**.

Apothecaries' Sundries Manufacturing Co.

Regd. Office:

ASCO House, No.13, Community Center, Mayapuri, Phase 1, New Delhi-110064, Delhi, India Tel : +91 9315680579, +91 8178737594 Email: info@ascoindia.com

Branch Office:

ASCO Medical Ltd. United Kingdom



ISO

13485:2003 CERTIFIED

- Government of India recognised -

"STAR EXPORT HOUSE" 🗐

ISO

9001:2000 CERTIFIED

WHO

GMP

www.asco-starlite.com



CERTIFICATE

This Certificate is issued to

Apothecaries' Sundries Manufacturing Co.

'ASCO' House 229, Udyog Vihar Phase IV, Gurugram-122015, Haryana INDIA

Who have implemented a Quality Management System, which Meets the requirements laid down in ISO 13485:2016, with the following scope:

Manufacture and Supply of Medical, Surgical, Laboratory Device & Equipments.

Certificate No : QL-I-1006-1126 Original Issue : 10 February 2021 Latest Issue : 10 February 2021 Valid Till : 9 February 2024

a L O L O

The Continuing Validity of this Certificate is Subject to timely Conduct of Surveillance audits



for QL Certification Authorised Signatory

ISO 13485:2016

To check the validity of the certificate, email to verify@ql-certification.com Colored reproduction of this certificates is not permitted. Upon request, the Certificates shall be returned to QL Certification



Australia Hong Kong

www.gl-certification.com

QL Certification

China

India



CERTIFICATE

The location-wise activities included in certificate No. QL-I-1006-1126 are:

S.No.	Location	Certified Activities
1.	<i>Office:</i> Apothecaries' Sundries Manufacturing Co 'ASCO' House, 229, Udyog Vihar Phase IV, Gurugram-122015, Haryana India	Top Management, Purchase and Marketing
2.	<i>Office:</i> Apothecaries' Sundries Manufacturing Co 'ASCO' House, 13, Community Center, Mayapuri, Phase I New Delhi 110064 India	
3.	<i>Factory:</i> Apothecaries' Sundries Manufacturing Co. F 1, Sector A2 Tronica City, Loni Ghaziabad Uttar Pradesh India	Production, QA, Training, Stores and Dispatch
4.	<i>Factory:</i> Apothecaries' Sundries Manufacturing Co. F 10, Sector A2 Tronica City, Loni Ghaziabad Uttar Pradesh India	Production, QA, Training, Stores and Dispatch
Valid 1	Fill 9 February 2024	Cher Hos
	IN TERNATION PC CPEDITATION FORM	for QL Certification Authorised Signatory

ISO 13485:2016

To check the validity of the certificate, email to verify@ql-certification.com Colored reproduction of this certificates is not permitted. Upon request, the Certificates shall be returned to QL Certification

Australia

Austria

www.gl-certification.com

QL Certification

Hong Kong

India

China



Certificate of Conformity

This Certificate is issued to

Apothecaries' Sundries Manufacturing Co. 'ASCO' House 13, Community Centre Mayapuri, Phase I New Delhi 110 064 INDIA

who have maintained a technical documentation, as required for **Class I** devices and the requirements laid down in Annex VII, section 3 of the Council Directive 93/42/EEC of 14 June 1993 for following products:

Surgical and Orthopedic Machines

(Further details of products are given on page 2 of 3 of this certificate. Any changes undertaken shall be reported immediately to Vexil BPS so that the same may be reviewed for continuing validity of the certificate)

Certificate No. : MD9100296 Original Issue : 21 September 2015 Latest Issue : 21 September 2018 Valid Till : 20 September 2021

100

for Vexil Business Process Services Private Limited



Any changes in the design or construction of the products or any amendments to the Directive as referred above, may render this certificate invalid. The product liability rests with the manufacturer or his representative in accordance with Council Directive 85/374/EEC

For checking validity of the certificate please call +91 11 2384 5001 or email to info@vexilbps.com. The details of the scope of certification can be obtained from the organization. The continuing validity of certificate is subject to fulfilment of requirements laid down in Conditions of Contract document (D1501) Colour photocoy is not permitted. Upon request the Certificate shall be returned to Vexil Business Process Services Private Limited. UK Contact:: 430 Kingston Road, Kingston-upon-Thames Surrey KT1 3PB (U.K.)

F0826A00

BUSINESS PROCESS SERVICES

Page 1 of 3



The Certificate No. MD9100296 issued to M/s Apothecaries' Sundries Manufacturing Co. covers the following devices:

S. No.	Product / Category	Code No.
1.	Surgical and Orthopedic	367.001, 367.002, 367.003, 367.004,
	Machines (Tourniquets,	367.004.01, 367.004.02, 367.005,
	Shaver Systems, Pulse	368.010CA, 368.010CB, 368.010D,
	Lavage Systems, Electric	369.001, 369.002, 369.003, 369.004,
	Drills and Saws, Manual	369.005, 369.006, 369.007, 369.008,
	Drills and Saws, Battery	369.009, 369.010, 369.011, 369C.001,
	Drills and Saws)	369D.001, 369D.002, 369D.003, 369D.004,
		369D.005, 369D.009, 369D.0091,
		369D.101, 369D.102, 369D.103, 369D.104,
		369D.105, 369D.106, 369D.201, 369D.202,
		369D.203, 369D.204, 369D.205, 369D.206,

369D.210A

Valid Till: 20 September 2021

gain 00

369D.207, 369D.208, 369D.209, 369D.210,

for Vexil Business Process Services Private Limited



For checking validity of the certificate please call +91 11 2384 5001 or email to info@vexilbps.com. The details of the scope of certification can be obtained from the organization. The continuing validity of certificate is subject to fulfillment of requirements laid down in Conditions of Contract document (D1501) Colour photocopy is not permitted. Upon request the Certificate shall be returned to Vexil Business Process Services Private Limited.

(F

F0826A00

'Page 2 of 3



VEXIL BUSINESS PROCESS SERVICES

The S. No.	location-wise activities included in Certificate N Location	No. MD9100296 are: Certified activities
1.	Office: Apothecaries' Sundries Manufacturing Co. 'ASCO' House 13, Community Centre Mayapuri, Phase I New Delhi 110 064 India	Top Management, Purchase and Marketing
2.	<i>Factory:</i> Apothecaries' Sundries Manufacturing Co. F 1, Sector A2 Tronica City, Loni Ghaziabad Uttar Pradesh India	Production, QA, Training, Stores and Dispatch
3.	Factory: Apothecaries' Sundries Manufacturing Co. F 10, Sector A2 Tronica City, Loni Ghaziabad Uttar Pradesh India	Production, QA, Training, Stores and Dispatch
4.	<i>Factory:</i> Apothecaries' Sundries Manufacturing Co. 20, Udyog Vihar, Sankhol, Bahadurgarh, Haryana India	Production, QA, Training, Stores and Dispatch
Valic	Till: 20 September 2021	for Vexil Business Process Services Private Limited
	cking validity of the certificate please call +91 11 2384 5001 or email to info@vexilbo ation. The continuing validity of certificate is subject to fulfillment of requirements laid d shotocopy is not permitted. Upon request the Certificate shall be returned to Vexil Busi 00	