



DECLARATION OF CONFORMITY



DLAB Scientific Co., Ltd
Add: Yu An Road 31, Airport
Economic core zone, Shunyi District,
Beijing, 101318 P.R.China

Product: **Clinical Centrifuges**

Model (s): DM0408 DM0424

The designated products are conform to the guideline:

2014/35/EU Low Voltage Directive
2014/30/EU Electromagnetic Compatibility Directive
2015/863/EU RoHS Directive

The conformity of the products with the requisites as set forth in the guideline is evidenced by the full compliance with the following standards:

EN 61010-1	Safety Requirements for Electrical Equipment for Measurement Control, and Laboratory Use – Part 1: General Requirements
EN 61010-2-020	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-020: Particular requirements for laboratory centrifuges
EN 61326-1	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
EN 61000-3-2	Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions
EN 61000-3-3	Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker

Beijing

2020-01-01


DLAB Scientific Co., Ltd
Quality Manager
