

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Urinalysis Reagent Strips

PURCHASE NO.: 37740066C

BATCH NO.: URS0030251

EXPIRATION DATE: 2022-03-28

QUANTITY: 100x100T

ANALYSIS		SPECIFICATIONS (Color card)	TEST RESULT
Glucose	Neg. urine	Negative	Complies
	250mg/dL	>100mg/dL <500mg/dL	Complies
Bilirubin	Neg. urine	Negative	Complies
	Yihua Urine controls	>++	Complies
Ketone	Neg. urine	Negative	Complies
	Yihua Urine controls	>40mg/dL <160mg/dL	Complies
S.G	1.000	≥1.000 <1.005	Complies
	1.025	>1.020 <1.030	Complies
Blood	Neg. urine	Negative	Complies
	Yihua Urine controls	≥++ ≤+++	Complies
pH	5.0	≥5.0 <6.0	Complies
	9.0	>8.0 ≤9.0	Complies
Protein	Neg. urine	Negative	Complies
	Yihua Urine controls	>+, ≤+++	Complies
Urobilinogen	Neg. urine	Negative	Complies
	4~12 mg/dL	>4mg/dl ≤12mg/dl	Complies
Nitrite	Neg. urine	Negative	Complies
	Yihua Urine controls	+	Complies
Leukocyte	Neg. urine	Negative	Complies
	Yihua Urine controls	>+, <+++	Complies



Quality Department: Arbog. Me

Date: July. 02, 2022