








SPARE PARTS LIST


Remark: All the mentioned spare parts are compatible with the **Plusteam Series, TBT** autoclave.

PRODUCT CODE	DESCRIPTION OF THE GOODS	PICTURE/INFO
5M1SS2	HEPA	Retains at least 99.99% of particles
4C0002	CONTACTOR-1	
S89649	STEAM TRAP	Discharges condensate from the steam line or equipment
1S0031	Pressure sensor chamber	
1S0032	Pressure sensor jaket	
1S0032	Pressure sensor generator	
5M1SS4	WATER LEVEL SET	a sensor setting that controls the filling or stopping of water supply;
5M1SS76	Solenoid Coil pneumatic valve	
5M1SS998	Gasket 67	
5M1SS998	Gasket 39	
1S0018	VACUUM PUMP-2	
5M1SS999	Gasket lubrikant	Silicone-based lubricant intended for autoclave gaskets with a temperature resistance of up to +250°C.

5MS9991	Pneumatic valve (jaket)	
7M1SS45	Safety valve chamber	3 BAR; To ensure safe operation by releasing excess pressure when it exceeds the system's design limit
7M1SS46	Safety valve jaket	3 BAR; To ensure safe operation by releasing excess pressure when it exceeds the system's design limit
7M1SS47	Safety valve generator	4 BAR; To ensure safe operation by releasing excess pressure when it exceeds the system's design limit

Bidder
Name/Stamp/Sign

Ergi EKLER


STERILMED MEDICAL
ELEKTRİK ELEKTRONİK OTOMASYON İNŞAAT
GIDA SANAYİ VE TİCARET LİMİTED SİRKETİ
Suyaniye Çarşı No: 15/10 Blok Kat: 18 C/40 No:43
06539 Maltepe/İstanbul / TÜRKİYE
Tel: +90 312 375 91 00 Fax: +90 312 375 92 92
Sınkay No: 7818644577 Mersis No: 0781044457700019
Ticaret Sicil No: 284029 info@sterilmed.com.tr

**STERILMED MEDICAL ELEKTRİK
ELEKTRONİK OTOMASYON İNŞAAT GIDA
SANAYİ VE TİCARET LİMİTED SİRKETİ**

