

Certificate of Analysis



Prep-n-Go™ Buffer

Product No. **4471406**
Lot No. **2412054**
Date of Manufacture **13DEC2024**
Expiration Date **09MAY2026**

KIT CONFIGURATIONS:

PART NO.	COMPONENT DESCRIPTION	LOT NO.
4471434	(1) Bottle Prep-n-Go™ Buffer	2411051

SPECIFICATIONS:

RESULT

QC Method for Prep-n-Go™ Buffer: The analytical QC results are obtained from PN 4467081 BULK, Prep-n-Go™ Buffer. The human DNA contamination results are obtained from P/N 4471434 BOTTLE, Prep-n-Go™ Buffer and are analyzed on an Applied Biosystems Genetic Analyzer using the AmpF ℓ STR® Identifiler® Direct PCR Amplification Kit.

No human STR profile detected using AmpF ℓ STR® Identifiler® Direct PCR Amplification Kit at 28 cycles	PASS
pH (8.15 - 8.30)	PASS
Conductivity (1.45 - 1.88 mS/cm)	PASS
Refractive Index (1.3330 - 1.3340)	PASS

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Warrington, UK

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 18DEC2024

Doc: 4471046 Rev.E