

Test Report

– Eppendorf PCR clean & sterile –

ep Dualfilter T.I.P.S.[®] 0.1 - 10 µL S



10x 96 Eppendorf Tips
PCR clean / sterile

LOT M212567L

2028-06-28

STERILE R

Order no.: 0030 078.500

Eppendorf SE
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 0030078500

Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

Human DNA

Method: real-time PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

DNase

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable
(LOD: 1.0 x 10⁻⁶ Kunitz units)

RNase

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable
(LOD: 1.0 x 10⁻⁹ Kunitz units)

PCR Inhibition

Method: real-time PCR acc. to SOP 100559/V1
Limit: less than 10 targets amplifiable

Sterility

Method: Sterilization with SAL 10⁻⁶ by
R: Irradiation according DIN EN ISO 11137:2015* or
EO: Ethylene oxide*

*Please see attached label for the standard method of sterilization for this product

Test method: USP, Ph. Eur. 8.0 / 2.6.12

Endotoxin

Method: Kinetic-chromogenic LAL test (Ph. Eur. 8.0 / 2.6.14)
Limit: < 0.001 EU/mL

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 12.07.23

Testing of the above described lot showed conformity within the limits of detection.

18.07.23

Date of release

Epp 2007-3531

Registration number

i.v. P. Weigel

Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.