

# CE DECLARATION OF CONFORMITY ECLIPSE/COPERNIC Temporary occlusion catheters

DCE-31 rev.2

Page 1 of 2

Manufacturer

**BALT Extrusion SAS** 

Address

10 rue de la Croix Vigneron

95160 Montmorency

France

Product

**ECLIPSE/COPERNIC – Temporary occlusion catheters** 

(See detailed references on page 2)

GMDN code 32584: Temporary occlusion catheters

Classification (MDD, Annex IX): Rule 6 Class III

We herewith declare in sole responsibility that as of the date of this declaration, the above mentioned product meets the provisions of the following EC Council Directives and Standards. All supporting documentation is retained under the premises of the manufacturer and the notified body.

### **Directive**

Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 concerning medical devices (MDD 93/42/EEC), Rule 6 per Annex IX of the MDD

### **Standards:**

Harmonized Standards (published in the Official Journal of the European Communities) applicable to these products can be referenced in TF-02 Section 3-4.

## Notified Body:

# **DQS Medizinprodukte GmbH**

August-Schanz-Straße 21 DE-60433 Frankfurt am Main Germany

Reg. #: 0297

Annex II section 3:

CE Certificate n° 513975 MR2

Valid Until:

26 May 2024

Annex II section 4:

CE Certificate n° 536988 MRA

Valid Until:

13 November 2022

Place, Date

Montmorency, France, 22 April 2021

Signed

Sonia TAMAZIRT, Senior Manager, Regulatory Affairs