

BOEN HEALTHCARE CO., LTD.

CERTIFICATE OF ANALYSIS

Report No.: GM03-BN23

Product Name: Inoculating Loops			Specification: Flexible, 10μL				
Lot. No.: 20230501			MFG. Date: 2023-05				
EXP. Date: 2028-04			Raw Material: ABS				
Standard: Enterprise Standard							
Package: 20pcs/bag, 10000pcs/carton							
Quantity: 500000pcs			Report Date: 2023.06.10				
Item	Technical Requirement	Inspection Level	Sample	AQL		Defect Qty	Result
Dimension	10μL	5pcs/lot	5pcs	0		0	Pass
				Ac	Re		
				0	1		
Appearance	No burr, black dot, dust, scratch, flash≤0.5mm, no oil spot, breakage. The color should be without obvious color difference.	GB/T 2828.1-2012 Normal test Level II	800pcs	1.5		0	Pass
				Ac	Re		
				21	22		
Function	The loops are no flash edge, capacity is 10μL	10pcs/lot	10pcs	0		0	Pass
				Ac	Re		
				0	1		
Package	The printing should be clean and clear.	1 outer box/lot	1 outer box	0		0	Pass
				Ac	Re		
				0	1		
Conclusion	The product meets all the requirement by out test.						

Inspector: Tom Li

Date: 2023.06.10

Checker: Leo Sun

Date: 2023.06.10

