



SwiftNINJA® Steerable Coronary Microcatheter











Catheter Specifications

Part Number	Inner Diameter	Outer Diameter Proximal Portion	Outer Diameter Distal Portion	Usable Length	Hydrophilic Coating Length	Max Guide wire	Maximium Injection Pressure	Distance From Tip To 1 st Markerband	Distance Between 1 st And 2 nd Markerbands	Cath Vol.
SNCM125	0.021" (0.54 mm)	2.9F (0.97 mm)	2.4F (0.80 mm)	125 cm	80 cm	≤ 0.018" (0.46 mm)	6,900 kPa (1,000 psi)	0.5 mm	13.5 mm	0.49 mL

Catheter Preparation for Use

To activate the hydrophilic coating: Flush the catheter with heparinized saline via the standard luer BEFORE removal from the protective catheter sleeve.

To Unlock Steering Dial (white wheel): Hold the lower housing body and gently pull the steering dial toward the luer connector until a clicking sound is heard.





Prior to attempting catheter tip movement or deflection, you must engage the steering dial

Ordering Information:

Part Number	Description					
SNCM125	SwiftNINJA® Steerable Coronary Microcatheter 2.9F-2.4F 125 cm					



Before using, refer to Instructions for Use for indications, contraindications, warnings, precautions and directions for use.



Understand. Innovate. Deliver."

Merit Medical Systems, Inc. 1600 West Merit Parkway South Jordan, Utah 84095 1 801 208 4300 1.800.35.MERIT

merit.com

Merit Medical Europe, Middle East, & Africa (EMEA) Amerikalaan 42, 6199 AE Maastricht-Airport

The Netherlands

+31 43 358 82 22

Merit Medical Ireland Ltd.

Parkmore Business Park West Galway, Ireland +353 (0) 91 703 733

Australia +61 (0) 3 8373 4943

Austria

0800 295 374

Belgium 0800 72 906 (Dutch) 0800 73 172 (Français)

+55 11 4561 - 0062

Canada

China +86-10 8561 0788

1.800.364.4370

Hong Kong +852-2207 6888

Denmark

Finland

France

Germany

80 88 00 24

0800 770 586

0800 91 60 30

0800 182 0871

+91-80-41223376

Ireland (Republic) 1800 553 163

800 897 005

+8210 2205 0707

Luxembourg 8002 25 22

Netherlands

0800 022 81 84

800 11629

Portugal 308 801 034

Russia +7 495 221 89 02 +34 911238406

Sweden

020 792 445

Switzerland +41 22 518 02 30

(Français) +41 22 518 02 52

(Italiano) +41 22 518 00 35

0800 973 115