

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: HBcAb EIA Test Kit

PURCHASE NO.: 37740068A

BATCH NO.: 2008762

EXPIRATION DATE: 2021/11/08

QUANTITY: 11kits

SPECIFICATION	STANDARD	CONCLUSION
Appearance	No visible deposit or floccules in liquid component	Complies
Negative Control	OD > 1.000	Complies
Positive Control	OD < 0.080	Complies
QCO reference	$0.500 \leq \text{QCO/Cutoff value} \leq 1.500$	Complies
Compliance rate of negative reference	15/15	Complies
Compliance rate of positive references	15/15	Complies
Precision criteria	CV < 20%	Complies

Quality Department: Hedy zhang

Date: Sep. 23, 2020

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Free T4 EIA Test Kit

PURCHASE NO.: 37740068A

BATCH NO.: 2008764

EXPIRATION DATE: 2021/06/24

QUANTITY: 2kits

SPECIFICATION	STANDARD	CONCLUSION
Physics Check	No sediment or floccule	Complies
Calibrator1#	$OD \geq 1.800$	Complies
Calibrator 6#	$OD \leq 0.400$	Complies
QL	$0.44 \leq R \leq 1.33$	Complies
QM	$0.66 \leq R \leq 1.34$	Complies
QH	$0.66 \leq R \leq 1.34$	Complies
QN1-QN8	Concentration $\leq 2\text{ng/dL}$ (8/8)	Complies
QH1,QH2	Concentration $> 2\text{ng/dL}$ (2/2)	Complies
Precision	$CV \leq 15\%$	Complies

Quality Department: Hedy zhang

Date: Sep. 23, 2020