


Model S-240



Model S-242



Features

- Designed to be stored in small places
 - Folding / Unfolding Chair
 - Telescoping front handles with two positions
 - Two rear handles (S-240) / Four rear handles (S-242)
 - Two rear wheels dia 125 mm (S-240)
 - Two wheels dia 80 mm with locks / Two wheels dia 125 mm (S-242)
 - Vinyl-coated nylon cover is simple to clean and resists stains, blood and bodily fluids
 - Three quick-release patient restraints and leg restraints hold the patient securely
- 60-0110-001 CHAIR S-240
 - 60-0110-002 CHAIR S-242


The Models S-240, S-242 Chair were designed, manufactured and tested to meet the specification of EN 1789 and EN 1865 standards.

Specifications



Length	Width	Height	Weight/without restraints/S-240	Weight/without lower rear handles/S-242	Load limit
74 cm	52.5 cm	92 cm	8,65 kg	10 kg	170 kg
Length in folded position		Height in folded position		Width in folded position	
92 cm		17 cm		53 cm	

EU Declaration of Conformity

Manufacturers Name:	Ferno Slovakia s.r.o.
Manufacturers Address:	Bošáca 893 / 913 07 / Slovakia
SRN (Single Registration Number):	SK-MF-000001474
Authorized Representative (UKCA REPRESENTATIVE)	Ferno House Stubs Beck Lane, Cleckheaton, West Yorkshire, BD19 4TZ United Kingdom +44 (0) 1274 851999 www.ferno.co.uk
	
Basic UDI-DI:	8588008129600110ChairsWU
Name of the Device (s):	CHAIR S-242
Product code:	60-0110-002
Classification:	Class I
Notified Body name:	N/A
Notified Body Address:	N/A
Notified Body Identification number:	N/A
Conformity assessment route:	Ferno Slovakia s.r.o. uses the following procedures for the CE-labeling of their products according the Regulation MDR 2017/745: EN 1865; EN 1789 <u>Class I:</u> EU conformity declaration according to Annex IV + Annex VIII of MDR 2017/745

This DoC is issued based on the ISO 13485 Certification.

This declaration of conformity is issued by Ferno Slovakia s.r.o.

We hereby declare that the medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices.

All supporting documentation is retained at the premises of the manufacturer.

Signature:




Silvia Vančová
Managing Director

Place and date of issue:

Bošáca, 24.6.2024

Attachment nr.1 to declaration of conformity Chair S-242**Chair S-242**

Part Number	Product description	GTIN (UDI-DI)
60-0110-002	CHAIR S-242	8588008129685



Slovakia

CERTIFICATE

TÜV SÜD Slovakia s.r.o.
Certification Body for Management Systems

Accredited by SNAS
Certificate on accreditation No. Q-011

certifies that



Ferno Slovakia s.r.o.
893
SK – 913 07 Bošáca
IČO: 35 809 400

has established and applies
a Quality Management System for

**Development, production, repair and modification of transport
and fixation equipment, medical devices.**

An audit was performed, Report No. **0401/30/23/Q/AS/R5**
Proof has been furnished that the requirements
according to

STN EN ISO 9001:2016

are fulfilled. The certificate is valid from **2023-05-07** until **2026-05-06**

Certificate Registration No. **Q 0401-6**



Bratislava, 2023-05-04

TÜV SÜD Slovakia s.r.o.
Certification Body for Management Systems
Member of Group TÜV SÜD
Jašíkova 6, 821 03 Bratislava



MAKE SURE LOCK BAR IS FULLY ENGAGED
LOCK BAR LEFT TO REMAIN ENGAGED AT ALL TIMES

FERNO
www.ferno.com

CE