

AB-0583-T

21003760-
İNG

02-21

TEST RESULT

MICROBIAL CLEANLINESS (Bioburden)

Test Metod: Ref: EN ISO 11737-1:2018

The sample is put in extraciton liquid after shaking well, inoculated on the agar.
After incubation at 30 ± 1 ° C for 72 hours, growth microorganisms are counted on the agar.

	<u>RESULTS</u>	<u>REQUIREMENT</u>
Microbial cleanliness (cfu/ 100 cm ²)	120 cfu/100 cm ²	≤300 cfu/100 cm ²

*cfu= Colony forming unit.