

To: Centrul pentru Achizitii Centralizate in Sanatate

Subject: 21120873 dated 22.12.2023 for procurement "Achiziționarea Dispozitivelor medicale (Tomograf Computerizat Ortopedic) conform necesităților IMSP Spitalul Clinic de Traumatologie și Ortopedie".

Declaration

We, Planmed Oy, Sorvaajankatu 7 00880 Helsinki Finland, manufacturer of Planmed Verity, hereby declare that the Planmed Verity has and comply with the following technical parameters:

- Scanning time in 3D: 24 seconds for standard protocol
- Exposure time: 8 seconds for standard protocol
- Fast GPU accelerated reconstruction
- Scan angle: 210°
- Bore diameter: 40 cm
- Service mode with intergrated QA tools

Helsinki, 2023-12-18,


Juha Huttunen
Vice President, Operations

