

Faraş Sedye (ES-300) Scoop Stretcher



Uzunluk Length	Genişlik Width	Ağırlık Weight
1650 mm.	430 mm.	9.6 Kg.

- Sertleştirilmiş özel alüminyumdan imal edilmiştir.
 - Maksimum 160 Kg. taşıma kapasitesine sahiptir.
 - Yerde yatan hasta ve yaralının sarsmadan yerden kaldırılabilmesini sağlar.
 - Her iki taraftan ikiye ayrılabilir, ayarlanabilir ve sabitlenebilir.
 - 3 adet hasta sabitleyici emniyet kemeri vardır.
-
- *Made of hardened aluminum.*
 - *160 Kg. maximum carrying capacity.*
 - *Designed for lifting patient or injured lying on the floor without shaking.*
 - *Both sides can be separated, adjusted and fixed.*
 - *3 fixing safety belts for patient.*

Kayar Tip Hasta Taşıma Sandalyesi (ES-220) Stair Chair



- Bu ürün hastaların merdivenlerden konfor içinde indirilmesi için kullanılır.
 - Sandalyenin arka çerçevesine yerleştirilmiş olan kayma sistemi ile hastanın uzun merdiven basamaklarında taşınması hastaya zarar vermeden ve kullanıcılar için de ilave bir insan gücüne ihtiyaç kalmayacak şekilde kolaylaştırmaktadır. Artık koltuğu hasta ile birlikte kaldırmanıza gerek yoktur, çünkü kaydırma sistemi merdivenden aşağı istikamete doğru güvenli ve basit bir kayma işlemini gerçekleştirir.
 - Kaydırma paletleri dayanıklı ve plastik esaslı bir maddeden ile imal edilmiş olup, kullanılmadığı durumlarda yukarı doğru katlanabilir şekilde bırakılır.
 - Zemin ve paletler arasında yeterli mesafe ve açı mevcuttur. Bunun amacı sandalyenin her türlü merdivende kaydırılabilir olmasıdır. Paletler merdiven haricinde normal kullanımı etkilemezler.
-
- *This product is used for transporting the patient by sliding down the stairs with comfort.*
 - *Sliding mechanism placed on the rear part of the chair helps to move patient on long stairs and provides harmless operation and removes additional human power need, for users. There is no need for you to carry patient with chair anymore, since sliding mechanism enables safe and simple motion to downstairs.*
 - *Sliding tracks made of durable and plastic based material and can be folded up in stationary conditions.*
 - *Sufficient distance and angle exist between tracks and ground to enable sliding on all kind of stairs. Tracks do not affect normal usage of the stair chair on regular surfaces.*

DECLARATION OF CONFORMITY



EMS MOBİL SİSTEMLER A.Ş..

1. Organize Sanayi Bölgesi Kırım Hanlığı Caddesi No: 9

Sincan-Ankara /TURKEY

Tel : +90 312 267 18 67 - Faks : +90 312 267 18 66

Product : Scoop Stretcher

Product Code : ES 300

Classification : Class I- Rule I

We herewith declare with this document that compliance evaluation was applied upon the above mentioned products as per the Council Directive 93/42/EEC Medical Devices Regulation Annex 7 for Medical Device Directive

General Applicable Directives

Medical Device Directive council directive 93/42/EEC of June 1993 concerning medical devices.
(MDD 93/42/EEC) Compliance Declaration evaluation was applied as per Annex 7.

Standarts

- TS EN 1865-1+A1:2015 Patient handling equipment used in road ambulances - Part 1: General stretcher systems and patient handling equipment
- TS EN 980:2011 Symbols for use in the labeling of medical devices
- TS EN 1041+A1:2015 Information supplied by the manufacturer of medical devices
- TS EN ISO 14971:2013 Medical devices - Application of risk management to medical devices

Signature

EMS MOBİL SİSTEMLER A.Ş.
1. Organize Sanayi Bölgesi
Kırımhanlığı Caddesi No: 9 Sincan/ANKARA
Tel: 0312 267 18 67 - Fax: 0312 267 18 66
Mersis No: 0475005500500012
Sincan Vergi Dairesi: 475 005 5005

Name : Mevlüt Şahin
Position : General Manager

Date : 15.05.2016